

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042023\
 Data File : PP057071.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Apr 2023 17:42
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042023AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 20 18:11:29 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 20 16:59:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.404	3.631	107.0E6	81305362	48.376	48.129
2) SA Decachlor...	10.217	8.686	75464849	88508791	47.025	49.475
Target Compounds						
26) L6 AR-1254-1	6.439	5.537	36285587	48415243	480.468	480.437
27) L6 AR-1254-2	6.659	5.685	51733978	42907029	476.995	480.177
28) L6 AR-1254-3	7.027	6.092	50011255	67635333	483.135	482.754
29) L6 AR-1254-4	7.314	6.321	29963257	36413511	482.291	483.313
30) L6 AR-1254-5	7.736	6.742	34785560	58487759	482.921	475.131



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042023\
Data File : PP057071.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Apr 2023 17:42
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042023AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 20 18:11:29 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 20 16:59:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.404 min
Delta R.T.: 0.000 min
Response: 107007192
Conc: 48.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042023AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.631 min
Delta R.T.: 0.000 min
Response: 81305362
Conc: 48.13 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.217 min
Delta R.T.: 0.000 min
Response: 75464849
Conc: 47.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.686 min
Delta R.T.: 0.000 min
Response: 88508791
Conc: 49.48 ng/ml



#26 AR-1254-1


R.T.: 6.439 min
Delta R.T.: 0.000 min
Response: 36285587
Conc: 480.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042023AR1254




#26 AR-1254-1

R.T.: 5.537 min
Delta R.T.: 0.000 min
Response: 48415243
Conc: 480.44 ng/ml




#27 AR-1254-2

R.T.: 6.659 min
Delta R.T.: 0.000 min
Response: 51733978
Conc: 477.00 ng/ml



#27 AR-1254-2


R.T.: 5.685 min
Delta R.T.: 0.000 min
Response: 42907029
Conc: 480.18 ng/ml



#28 AR-1254-3


R.T.: 7.027 min
Delta R.T.: 0.000 min
Response: 50011255
Conc: 483.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042023AR1254




#28 AR-1254-3

R.T.: 6.092 min
Delta R.T.: 0.000 min
Response: 67635333
Conc: 482.75 ng/ml




#29 AR-1254-4

R.T.: 7.314 min
Delta R.T.: 0.000 min
Response: 29963257
Conc: 482.29 ng/ml



#29 AR-1254-4


R.T.: 6.321 min
Delta R.T.: 0.000 min
Response: 36413511
Conc: 483.31 ng/ml



#30 AR-1254-5

R.T.: 7.736 min
Delta R.T.: 0.000 min
Response: 34785560
Conc: 482.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042023AR1254



#30 AR-1254-5

R.T.: 6.742 min
Delta R.T.: 0.001 min
Response: 58487759
Conc: 475.13 ng/ml