

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042023\
 Data File : PP057074.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Apr 2023 18:32
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 01:11:44 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 20 16:59:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.404	3.631	112.8E6	85752505	51.008	50.761
2)	SA Decachlor...	10.217	8.687	77362535	90795714	48.207	50.754
Target Compounds							
16)	L4 AR-1242-1	5.579	4.728	24933623	23214057	487.133	496.096
17)	L4 AR-1242-2	5.601	4.747	37296351	32329500	497.944	497.746
18)	L4 AR-1242-3	5.664	4.925	23625428	17888739	501.559	503.650
19)	L4 AR-1242-4	5.762	5.009	18740578	19942346	502.381	501.547
20)	L4 AR-1242-5	6.503	5.536	16884349	22378136	501.271	500.219



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042023\
Data File : PP057074.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Apr 2023 18:32
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 01:11:44 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 20 16:59:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.404 min
Delta R.T.: 0.000 min
Response: 112829138
Conc: 51.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.631 min
Delta R.T.: 0.000 min
Response: 85752505
Conc: 50.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.217 min
Delta R.T.: 0.000 min
Response: 77362535
Conc: 48.21 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.687 min
Delta R.T.: 0.001 min
Response: 90795714
Conc: 50.75 ng/ml



#16 AR-1242-1


R.T.: 5.579 min
Delta R.T.: 0.001 min
Response: 24933623
Conc: 487.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.728 min
Delta R.T.: 0.000 min
Response: 23214057
Conc: 496.10 ng/ml




#17 AR-1242-2

R.T.: 5.601 min
Delta R.T.: 0.000 min
Response: 37296351
Conc: 497.94 ng/ml



#17 AR-1242-2


R.T.: 4.747 min
Delta R.T.: 0.000 min
Response: 32329500
Conc: 497.75 ng/ml



#18 AR-1242-3


R.T.: 5.664 min
Delta R.T.: 0.000 min
Response: 23625428
Conc: 501.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.925 min
Delta R.T.: 0.000 min
Response: 17888739
Conc: 503.65 ng/ml




#19 AR-1242-4

R.T.: 5.762 min
Delta R.T.: 0.000 min
Response: 18740578
Conc: 502.38 ng/ml



#19 AR-1242-4


R.T.: 5.009 min
Delta R.T.: 0.000 min
Response: 19942346
Conc: 501.55 ng/ml



#20 AR-1242-5

R.T.: 6.503 min
Delta R.T.: 0.000 min
Response: 16884349
Conc: 501.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.536 min
Delta R.T.: 0.000 min
Response: 22378136
Conc: 500.22 ng/ml