

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042023\
 Data File : PP057104.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Apr 2023 03:55
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 05:13:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 20 16:59:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.403	3.631	110.8E6	88393836	50.088	52.325
2) SA Decachlor...	10.216	8.685	80570640	83608309	50.206	46.736
Target Compounds						
16) L4 AR-1242-1	5.578	4.728	24382941	23768951	476.375	507.954
17) L4 AR-1242-2	5.601	4.747	34953250	33156210	466.661	510.474
18) L4 AR-1242-3	5.664	4.925	22266926	18162348	472.718	511.353
19) L4 AR-1242-4	5.762	5.009	18137847	20079160	486.224	504.988
20) L4 AR-1242-5	6.503	5.536	16922424	22767022	502.401	508.912



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042023\
Data File : PP057104.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Apr 2023 03:55
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 05:13:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 20 16:59:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.403 min
Delta R.T.: 0.000 min
Response: 110794148
Conc: 50.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.631 min
Delta R.T.: 0.000 min
Response: 88393836
Conc: 52.33 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.216 min
Delta R.T.: -0.001 min
Response: 80570640
Conc: 50.21 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.685 min
Delta R.T.: 0.000 min
Response: 83608309
Conc: 46.74 ng/ml



#16 AR-1242-1


R.T.: 5.578 min
Delta R.T.: 0.000 min
Response: 24382941
Conc: 476.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.728 min
Delta R.T.: 0.000 min
Response: 23768951
Conc: 507.95 ng/ml




#17 AR-1242-2

R.T.: 5.601 min
Delta R.T.: 0.000 min
Response: 34953250
Conc: 466.66 ng/ml



#17 AR-1242-2


R.T.: 4.747 min
Delta R.T.: 0.000 min
Response: 33156210
Conc: 510.47 ng/ml



#18 AR-1242-3


R.T.: 5.664 min
Delta R.T.: 0.000 min
Response: 22266926
Conc: 472.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.925 min
Delta R.T.: 0.000 min
Response: 18162348
Conc: 511.35 ng/ml




#19 AR-1242-4

R.T.: 5.762 min
Delta R.T.: 0.000 min
Response: 18137847
Conc: 486.22 ng/ml



#19 AR-1242-4


R.T.: 5.009 min
Delta R.T.: 0.000 min
Response: 20079160
Conc: 504.99 ng/ml



#20 AR-1242-5

R.T.: 6.503 min
Delta R.T.: 0.000 min
Response: 16922424
Conc: 502.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.536 min
Delta R.T.: 0.000 min
Response: 22767022
Conc: 508.91 ng/ml