

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042023\
 Data File : PP057116.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Apr 2023 07:45
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 08:23:37 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 20 16:59:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.404	3.631	114.2E6	84259193	51.626	49.877
2)	SA Decachlor...	10.219	8.687	84955745	83208155	52.939	46.512
Target Compounds							
3)	L1 AR-1016-1	5.579	4.728	30503169	29182545	486.235	513.727
4)	L1 AR-1016-2	5.602	4.747	44779862	41469529	490.728	517.524
5)	L1 AR-1016-3	5.664	4.925	28387538	22413888	492.208	516.000
6)	L1 AR-1016-4	5.763	4.967	22497632	19629113	495.082	505.843
7)	L1 AR-1016-5	6.060	5.182	24168708	27236554	501.600	543.220
31)	L7 AR-1260-1	7.194	6.223	45752452	50847478	525.317	501.782
32)	L7 AR-1260-2	7.452	6.411	47999383	58092586	522.685	499.797
33)	L7 AR-1260-3	7.813	6.567	38655004	54248088	532.681	490.368
34)	L7 AR-1260-4	8.040	7.041	43931471	44398764	522.803	484.707
35)	L7 AR-1260-5	8.363	7.283	78087714	85560341	544.123	474.035



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042023\
Data File : PP057116.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Apr 2023 07:45
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 08:23:37 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 20 16:59:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.404 min
Delta R.T.: 0.000 min
Response: 114196645
Conc: 51.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.631 min
Delta R.T.: 0.000 min
Response: 84259193
Conc: 49.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.219 min
Delta R.T.: 0.002 min
Response: 84955745
Conc: 52.94 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.687 min
Delta R.T.: 0.002 min
Response: 83208155
Conc: 46.51 ng/ml



#3 AR-1016-1


R.T.: 5.579 min
Delta R.T.: 0.000 min
Response: 30503169
Conc: 486.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.728 min
Delta R.T.: 0.000 min
Response: 29182545
Conc: 513.73 ng/ml




#4 AR-1016-2

R.T.: 5.602 min
Delta R.T.: 0.000 min
Response: 44779862
Conc: 490.73 ng/ml



#4 AR-1016-2


R.T.: 4.747 min
Delta R.T.: 0.000 min
Response: 41469529
Conc: 517.52 ng/ml



#5 AR-1016-3


R.T.: 5.664 min
Delta R.T.: 0.000 min
Response: 28387538
Conc: 492.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.925 min
Delta R.T.: 0.000 min
Response: 22413888
Conc: 516.00 ng/ml




#6 AR-1016-4

R.T.: 5.763 min
Delta R.T.: 0.000 min
Response: 22497632
Conc: 495.08 ng/ml



#6 AR-1016-4


R.T.: 4.967 min
Delta R.T.: 0.000 min
Response: 19629113
Conc: 505.84 ng/ml



#7 AR-1016-5


R.T.: 6.060 min
Delta R.T.: 0.000 min
Response: 24168708
Conc: 501.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.182 min
Delta R.T.: 0.000 min
Response: 27236554
Conc: 543.22 ng/ml




#31 AR-1260-1

R.T.: 7.194 min
Delta R.T.: 0.001 min
Response: 45752452
Conc: 525.32 ng/ml



#31 AR-1260-1


R.T.: 6.223 min
Delta R.T.: 0.000 min
Response: 50847478
Conc: 501.78 ng/ml



#32 AR-1260-2


R.T.: 7.452 min
Delta R.T.: 0.001 min
Response: 47999383
Conc: 522.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.411 min
Delta R.T.: 0.001 min
Response: 58092586
Conc: 499.80 ng/ml




#33 AR-1260-3

R.T.: 7.813 min
Delta R.T.: 0.000 min
Response: 38655004
Conc: 532.68 ng/ml



#33 AR-1260-3


R.T.: 6.567 min
Delta R.T.: 0.000 min
Response: 54248088
Conc: 490.37 ng/ml



#34 AR-1260-4


R.T.: 8.040 min
Delta R.T.: 0.001 min
Response: 43931471
Conc: 522.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.041 min
Delta R.T.: 0.000 min
Response: 44398764
Conc: 484.71 ng/ml



#35 AR-1260-5

R.T.: 8.363 min
Delta R.T.: 0.001 min
Response: 78087714
Conc: 544.12 ng/ml



#35 AR-1260-5

R.T.: 7.283 min
Delta R.T.: 0.002 min
Response: 85560341
Conc: 474.03 ng/ml