

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042121\
 Data File : PP035090.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Apr 2021 15:17
 Operator : DD\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 17:20:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 21 17:19:31 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.836	4.251	7776100	4753249	97.169	95.983
2)	SA Decachlor...	10.749	9.544	5979574	2110753	94.977	94.892
Target Compounds							
26)	L6 AR-1254-1	7.053	6.369	2364829	1543963	936.629	926.342
27)	L6 AR-1254-2	7.282	6.525	3501085	1354874	943.308	926.401
28)	L6 AR-1254-3	7.662	6.945	3579477	1987538	968.890	949.019
29)	L6 AR-1254-4	7.958	7.183	2057656	870412	1005.858	997.421
30)	L6 AR-1254-5	8.385	7.607	2877397	1715632	999.137	991.719



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042121\
Data File : PP035090.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Apr 2021 15:17
Operator : DD\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 17:20:13 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042121.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 21 17:19:31 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.836 min
Delta R.T.: 0.000 min
Response: 7776100
Conc: 97.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#1 Tetrachloro-m-xylene

R.T.: 4.251 min
Delta R.T.: -0.001 min
Response: 4753249
Conc: 95.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.749 min
Delta R.T.: -0.001 min
Response: 5979574
Conc: 94.98 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.544 min
Delta R.T.: 0.000 min
Response: 2110753
Conc: 94.89 ng/ml



#26 AR-1254-1


R.T.: 7.053 min
Delta R.T.: 0.000 min
Response: 2364829
Conc: 936.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#26 AR-1254-1

R.T.: 6.369 min
Delta R.T.: 0.000 min
Response: 1543963
Conc: 926.34 ng/ml




#27 AR-1254-2

R.T.: 7.282 min
Delta R.T.: -0.001 min
Response: 3501085
Conc: 943.31 ng/ml



#27 AR-1254-2


R.T.: 6.525 min
Delta R.T.: 0.000 min
Response: 1354874
Conc: 926.40 ng/ml



#28 AR-1254-3


R.T.: 7.662 min
Delta R.T.: -0.002 min
Response: 3579477
Conc: 968.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#28 AR-1254-3

R.T.: 6.945 min
Delta R.T.: 0.000 min
Response: 1987538
Conc: 949.02 ng/ml




#29 AR-1254-4

R.T.: 7.958 min
Delta R.T.: -0.001 min
Response: 2057656
Conc: 1005.86 ng/ml



#29 AR-1254-4


R.T.: 7.183 min
Delta R.T.: 0.000 min
Response: 870412
Conc: 997.42 ng/ml



#30 AR-1254-5

R.T.: 8.385 min
Delta R.T.: 0.000 min
Response: 2877397
Conc: 999.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000



#30 AR-1254-5

R.T.: 7.607 min
Delta R.T.: 0.000 min
Response: 1715632
Conc: 991.72 ng/ml