

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042121\
 Data File : PP035094.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Apr 2021 16:25
 Operator : DD\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 17:21:06 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 21 17:19:31 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.836	4.252	409896	256980	5.122	5.189
2)	SA Decachlor...	10.748	9.545	319989	118289	5.083	5.318
Target Compounds							
26)	L6 AR-1254-1	7.052	6.369	145081	92720	57.462	55.630
27)	L6 AR-1254-2	7.282	6.526	211862	83451	57.083	57.060
28)	L6 AR-1254-3	7.663	6.946	197249	115774	53.391	55.280
29)	L6 AR-1254-4	7.960	7.184	107290	45038	52.447	51.609
30)	L6 AR-1254-5	8.385	7.608	145340	91863	50.467	53.101



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042121\
Data File : PP035094.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Apr 2021 16:25
Operator : DD\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 17:21:06 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042121.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 21 17:19:31 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.836 min
Delta R.T.: 0.000 min
Response: 409896
Conc: 5.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050




#1 Tetrachloro-m-xylene

R.T.: 4.252 min
Delta R.T.: 0.000 min
Response: 256980
Conc: 5.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.748 min
Delta R.T.: -0.002 min
Response: 319989
Conc: 5.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.545 min
Delta R.T.: 0.000 min
Response: 118289
Conc: 5.32 ng/ml



#26 AR-1254-1


R.T.: 7.052 min
Delta R.T.: -0.002 min
Response: 145081
Conc: 57.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050




#26 AR-1254-1

R.T.: 6.369 min
Delta R.T.: 0.000 min
Response: 92720
Conc: 55.63 ng/ml




#27 AR-1254-2

R.T.: 7.282 min
Delta R.T.: 0.000 min
Response: 211862
Conc: 57.08 ng/ml



#27 AR-1254-2


R.T.: 6.526 min
Delta R.T.: 0.000 min
Response: 83451
Conc: 57.06 ng/ml



#28 AR-1254-3


R.T.: 7.663 min
Delta R.T.: -0.001 min
Response: 197249
Conc: 53.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050




#28 AR-1254-3

R.T.: 6.946 min
Delta R.T.: 0.000 min
Response: 115774
Conc: 55.28 ng/ml




#29 AR-1254-4

R.T.: 7.960 min
Delta R.T.: 0.000 min
Response: 107290
Conc: 52.45 ng/ml



#29 AR-1254-4


R.T.: 7.184 min
Delta R.T.: 0.000 min
Response: 45038
Conc: 51.61 ng/ml



#30 AR-1254-5

R.T.: 8.385 min
Delta R.T.: -0.002 min
Response: 145340
Conc: 50.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050



#30 AR-1254-5

R.T.: 7.608 min
Delta R.T.: 0.000 min
Response: 91863
Conc: 53.10 ng/ml