

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042122\
 Data File : PP046426.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Apr 2022 20:45
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 01:08:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 22 01:08:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.866	3.129	65168558	43183773	25.048	25.525
2) SA Decachlor...	9.952	8.455	39023387	40852968	25.489	26.361
Target Compounds						
26) L6 AR-1254-1	6.034	5.111	19886539	22501466	256.460	269.919
27) L6 AR-1254-2	6.272	5.271	30176744	19451995	257.593	270.915
28) L6 AR-1254-3	6.672	5.700	29802407	29971352	255.936	263.959
29) L6 AR-1254-4	6.985	5.949	22067200	20200283	256.998	266.716
30) L6 AR-1254-5	7.445	6.398	22021831	27919546	253.944	263.142

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042122\
Data File : PP046426.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Apr 2022 20:45
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 01:08:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042122.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 22 01:08:11 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

