

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042123\
 Data File : PP057122.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Apr 2023 09:26
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 21:32:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 20 16:59:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.403	3.631	112.3E6	87102521	50.761	51.561
2) SA Decachlor...	10.220	8.688	83927647	85513623	52.298	47.801
Target Compounds						
3) L1 AR-1016-1	5.579	4.728	31947291	29001245	509.255	510.536
4) L1 AR-1016-2	5.601	4.747	46350909	40922009	507.945	510.691
5) L1 AR-1016-3	5.664	4.926	29583858	22549536	512.951	519.123
6) L1 AR-1016-4	5.762	4.967	23391376	19958939	514.749	514.343
7) L1 AR-1016-5	6.060	5.182	24969344	27295233	518.216	544.390
31) L7 AR-1260-1	7.194	6.223	46928700	52208233	538.823	515.211
32) L7 AR-1260-2	7.451	6.412	48231002	59022040	525.207	507.793
33) L7 AR-1260-3	7.814	6.568	39467239	56465566	543.874	510.412
34) L7 AR-1260-4	8.041	7.042	44799945	46869767	533.138	511.683
35) L7 AR-1260-5	8.364	7.284	76445852	87514876	532.683	484.864



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042123\
Data File : PP057122.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Apr 2023 09:26
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 21:32:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 20 16:59:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.403 min
Delta R.T.: 0.000 min
Response: 112284409
Conc: 50.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.631 min
Delta R.T.: 0.000 min
Response: 87102521
Conc: 51.56 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.220 min
Delta R.T.: 0.003 min
Response: 83927647
Conc: 52.30 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.688 min
Delta R.T.: 0.003 min
Response: 85513623
Conc: 47.80 ng/ml



#3 AR-1016-1


R.T.: 5.579 min
Delta R.T.: 0.000 min
Response: 31947291
Conc: 509.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.728 min
Delta R.T.: 0.000 min
Response: 29001245
Conc: 510.54 ng/ml




#4 AR-1016-2

R.T.: 5.601 min
Delta R.T.: 0.000 min
Response: 46350909
Conc: 507.94 ng/ml



#4 AR-1016-2


R.T.: 4.747 min
Delta R.T.: 0.000 min
Response: 40922009
Conc: 510.69 ng/ml



#5 AR-1016-3


R.T.: 5.664 min
Delta R.T.: 0.000 min
Response: 29583858
Conc: 512.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.926 min
Delta R.T.: 0.000 min
Response: 22549536
Conc: 519.12 ng/ml




#6 AR-1016-4

R.T.: 5.762 min
Delta R.T.: 0.000 min
Response: 23391376
Conc: 514.75 ng/ml



#6 AR-1016-4


R.T.: 4.967 min
Delta R.T.: 0.000 min
Response: 19958939
Conc: 514.34 ng/ml



#7 AR-1016-5


R.T.: 6.060 min
Delta R.T.: 0.000 min
Response: 24969344
Conc: 518.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.182 min
Delta R.T.: 0.000 min
Response: 27295233
Conc: 544.39 ng/ml




#31 AR-1260-1

R.T.: 7.194 min
Delta R.T.: 0.000 min
Response: 46928700
Conc: 538.82 ng/ml



#31 AR-1260-1


R.T.: 6.223 min
Delta R.T.: 0.000 min
Response: 52208233
Conc: 515.21 ng/ml



#32 AR-1260-2


R.T.: 7.451 min
Delta R.T.: 0.000 min
Response: 48231002
Conc: 525.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.412 min
Delta R.T.: 0.002 min
Response: 59022040
Conc: 507.79 ng/ml




#33 AR-1260-3

R.T.: 7.814 min
Delta R.T.: 0.002 min
Response: 39467239
Conc: 543.87 ng/ml



#33 AR-1260-3


R.T.: 6.568 min
Delta R.T.: 0.002 min
Response: 56465566
Conc: 510.41 ng/ml



#34 AR-1260-4


R.T.: 8.041 min
Delta R.T.: 0.002 min
Response: 44799945
Conc: 533.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.042 min
Delta R.T.: 0.002 min
Response: 46869767
Conc: 511.68 ng/ml



#35 AR-1260-5

R.T.: 8.364 min
Delta R.T.: 0.002 min
Response: 76445852
Conc: 532.68 ng/ml



#35 AR-1260-5

R.T.: 7.284 min
Delta R.T.: 0.002 min
Response: 87514876
Conc: 484.86 ng/ml