

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042123\
 Data File : PP057148.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Apr 2023 18:14
 Operator : YP\AJ
 Sample : 02448-04MSD
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 20230324-COMP-2MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 22:09:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 20 16:59:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.405	3.632	46233852	33074988	20.901	19.579
2) SA Decachlor...	10.219	8.686	29619673	31514266	18.457	17.616
Target Compounds						
3) L1 AR-1016-1	5.579	4.729	31792030	28926770	506.780	509.224
4) L1 AR-1016-2	5.602	4.748	46652040	40728479	511.245	508.276
5) L1 AR-1016-3	5.665	4.927	28853049	22371659	500.280	515.028
6) L1 AR-1016-4	5.763	4.968	22841156	20216417	502.641	520.978
7) L1 AR-1016-5	6.060	5.183	25000412	25742619	518.861	513.424
31) L7 AR-1260-1	7.193	6.224	45341616	50832422	520.600	501.634
32) L7 AR-1260-2	7.451	6.412	46225506	56987838	503.368	490.292
33) L7 AR-1260-3	7.813	6.567	32462746	54900035	447.350	496.261
34) L7 AR-1260-4	8.040	7.042	39162557	39972123	466.051	436.381
35) L7 AR-1260-5	8.364	7.283	63293685	76160693	441.037	421.957



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042123\
Data File : PP057148.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Apr 2023 18:14
Operator : YP\AJ
Sample : 02448-04MSD
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
20230324-COMP-2MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 22:09:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 20 16:59:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.405 min
Delta R.T.: 0.000 min
Response: 46233852
Conc: 20.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
20230324-COMP-2MSD




#1 Tetrachloro-m-xylene

R.T.: 3.632 min
Delta R.T.: 0.001 min
Response: 33074988
Conc: 19.58 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.219 min
Delta R.T.: 0.002 min
Response: 29619673
Conc: 18.46 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.686 min
Delta R.T.: 0.000 min
Response: 31514266
Conc: 17.62 ng/ml



#3 AR-1016-1


R.T.: 5.579 min
Delta R.T.: 0.000 min
Response: 31792030
Conc: 506.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
20230324-COMP-2MSD




#3 AR-1016-1

R.T.: 4.729 min
Delta R.T.: 0.001 min
Response: 28926770
Conc: 509.22 ng/ml




#4 AR-1016-2

R.T.: 5.602 min
Delta R.T.: 0.000 min
Response: 46652040
Conc: 511.24 ng/ml



#4 AR-1016-2


R.T.: 4.748 min
Delta R.T.: 0.000 min
Response: 40728479
Conc: 508.28 ng/ml



#5 AR-1016-3


R.T.: 5.665 min
Delta R.T.: 0.000 min
Response: 28853049
Conc: 500.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
20230324-COMP-2MSD




#5 AR-1016-3

R.T.: 4.927 min
Delta R.T.: 0.001 min
Response: 22371659
Conc: 515.03 ng/ml




#6 AR-1016-4

R.T.: 5.763 min
Delta R.T.: 0.000 min
Response: 22841156
Conc: 502.64 ng/ml



#6 AR-1016-4


R.T.: 4.968 min
Delta R.T.: 0.001 min
Response: 20216417
Conc: 520.98 ng/ml



#7 AR-1016-5


R.T.: 6.060 min
Delta R.T.: 0.000 min
Response: 25000412
Conc: 518.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
20230324-COMP-2MSD




#7 AR-1016-5

R.T.: 5.183 min
Delta R.T.: 0.001 min
Response: 25742619
Conc: 513.42 ng/ml




#31 AR-1260-1

R.T.: 7.193 min
Delta R.T.: 0.000 min
Response: 45341616
Conc: 520.60 ng/ml



#31 AR-1260-1


R.T.: 6.224 min
Delta R.T.: 0.001 min
Response: 50832422
Conc: 501.63 ng/ml



#32 AR-1260-2


R.T.: 7.451 min
Delta R.T.: 0.000 min
Response: 46225506
Conc: 503.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
20230324-COMP-2MSD




#32 AR-1260-2

R.T.: 6.412 min
Delta R.T.: 0.002 min
Response: 56987838
Conc: 490.29 ng/ml




#33 AR-1260-3

R.T.: 7.813 min
Delta R.T.: 0.000 min
Response: 32462746
Conc: 447.35 ng/ml



#33 AR-1260-3


R.T.: 6.567 min
Delta R.T.: 0.001 min
Response: 54900035
Conc: 496.26 ng/ml



#34 AR-1260-4


R.T.: 8.040 min
Delta R.T.: 0.001 min
Response: 39162557
Conc: 466.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
20230324-COMP-2MSD




#34 AR-1260-4

R.T.: 7.042 min
Delta R.T.: 0.001 min
Response: 39972123
Conc: 436.38 ng/ml



#35 AR-1260-5

R.T.: 8.364 min
Delta R.T.: 0.001 min
Response: 63293685
Conc: 441.04 ng/ml



#35 AR-1260-5

R.T.: 7.283 min
Delta R.T.: 0.002 min
Response: 76160693
Conc: 421.96 ng/ml