

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042222\
 Data File : PP046459.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2022 17:33
 Operator : YP\AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1262ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/25/2022
 Supervised By :mohammad ahmed 04/25/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 18:12:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 22 18:11:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.866	3.129	126.5E6	180.2E6	50.000	50.000
2) SA Decachlor...	9.954	8.455	78305320	129.3E6	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	7.528	6.439	60434294	123.5E6	500.000	500.000
37) L8 AR-1262-2	8.116	6.718	96154416	96644208	500.000	500.000
38) L8 AR-1262-3	8.441	7.289	67130684	81181919	500.000	500.000
39) L8 AR-1262-4	8.536	7.360	31490837	143.2E6	491.223m	500.000
40) L8 AR-1262-5	9.206	7.903	34497924	65603356	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042222\
Data File : PP046459.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2022 17:33
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/25/2022
Supervised By :mohammad ahmed 04/25/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 18:12:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 22 18:11:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.866 min
Delta R.T.: 0.000 min
Response: 126485897
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/25/2022
Supervised By :mohammad ahmed 04/25/2022




#1 Tetrachloro-m-xylene

R.T.: 3.129 min
Delta R.T.: 0.000 min
Response: 180243739
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.954 min
Delta R.T.: 0.000 min
Response: 78305320
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.455 min
Delta R.T.: 0.000 min
Response: 129334800
Conc: 50.00 ng/ml




#36 AR-1262-1

R.T.: 7.528 min
Delta R.T.: 0.000 min
Response: 60434294
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/25/2022
Supervised By :mohammad ahmed 04/25/2022




#36 AR-1262-1

R.T.: 6.439 min
Delta R.T.: 0.000 min
Response: 123516187
Conc: 500.00 ng/ml




#37 AR-1262-2

R.T.: 8.116 min
Delta R.T.: 0.000 min
Response: 96154416
Conc: 500.00 ng/ml



#37 AR-1262-2

R.T.: 6.718 min
Delta R.T.: 0.000 min
Response: 96644208
Conc: 500.00 ng/ml




#38 AR-1262-3

R.T.: 8.441 min
Delta R.T.: 0.000 min
Response: 67130684
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/25/2022
Supervised By :mohammad ahmed 04/25/2022




#38 AR-1262-3

R.T.: 7.289 min
Delta R.T.: 0.000 min
Response: 81181919
Conc: 500.00 ng/ml




#39 AR-1262-4

R.T.: 8.536 min
Delta R.T.: 0.000 min
Response: 31490837
Conc: 491.22 ng/ml m



#39 AR-1262-4

R.T.: 7.360 min
Delta R.T.: 0.000 min
Response: 143199437
Conc: 500.00 ng/ml




#40 AR-1262-5

R.T.: 9.206 min
Delta R.T.: 0.000 min
Response: 34497924
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/25/2022
Supervised By :mohammad ahmed 04/25/2022



#40 AR-1262-5

R.T.: 7.903 min
Delta R.T.: 0.000 min
Response: 65603356
Conc: 500.00 ng/ml