

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042321\
 Data File : PP035213.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Apr 2021 15:24
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 16:50:51 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 21 17:37:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.839	4.246	3650606	2298204	44.801	44.628
2)	SA Decachlor...	10.747	9.533	2902502	1069176	45.931	44.999
Target Compounds							
3)	L1 AR-1016-1	6.151	5.497	944841	512710	459.722	485.003
4)	L1 AR-1016-2	6.174	5.517	1467309	815555	445.687	441.859
5)	L1 AR-1016-3	6.240	5.709	923695	433165	423.594	482.513
6)	L1 AR-1016-4	6.346	5.756	730291	347123	421.633	414.827
7)	L1 AR-1016-5	6.658	5.987	725758	452749	409.965	446.354
31)	L7 AR-1260-1	7.828	7.072	1503230	744228	446.363	425.381
32)	L7 AR-1260-2	8.093	7.266	1743157	862244	512.027	483.695
33)	L7 AR-1260-3	8.458	7.422	1080866	776143	430.622	430.445
34)	L7 AR-1260-4	8.688	7.900	1307053	555685	445.295	436.553
35)	L7 AR-1260-5	9.005	8.144	2552246	1237269	518.787	508.898



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042321\
Data File : PP035213.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Apr 2021 15:24
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 16:50:51 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042121.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 21 17:37:28 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.839 min
Delta R.T.: 0.003 min
Response: 3650606
Conc: 44.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.246 min
Delta R.T.: -0.006 min
Response: 2298204
Conc: 44.63 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.747 min
Delta R.T.: -0.002 min
Response: 2902502
Conc: 45.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.533 min
Delta R.T.: -0.011 min
Response: 1069176
Conc: 45.00 ng/ml



#3 AR-1016-1


R.T.: 6.151 min
Delta R.T.: 0.001 min
Response: 944841
Conc: 459.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.497 min
Delta R.T.: -0.007 min
Response: 512710
Conc: 485.00 ng/ml




#4 AR-1016-2

R.T.: 6.174 min
Delta R.T.: 0.001 min
Response: 1467309
Conc: 445.69 ng/ml



#4 AR-1016-2


R.T.: 5.517 min
Delta R.T.: -0.008 min
Response: 815555
Conc: 441.86 ng/ml



#5 AR-1016-3


R.T.: 6.240 min
Delta R.T.: 0.002 min
Response: 923695
Conc: 423.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.709 min
Delta R.T.: -0.007 min
Response: 433165
Conc: 482.51 ng/ml




#6 AR-1016-4

R.T.: 6.346 min
Delta R.T.: 0.000 min
Response: 730291
Conc: 421.63 ng/ml



#6 AR-1016-4


R.T.: 5.756 min
Delta R.T.: -0.006 min
Response: 347123
Conc: 414.83 ng/ml



#7 AR-1016-5


R.T.: 6.658 min
Delta R.T.: 0.000 min
Response: 725758
Conc: 409.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.987 min
Delta R.T.: -0.007 min
Response: 452749
Conc: 446.35 ng/ml




#31 AR-1260-1

R.T.: 7.828 min
Delta R.T.: 0.000 min
Response: 1503230
Conc: 446.36 ng/ml



#31 AR-1260-1


R.T.: 7.072 min
Delta R.T.: -0.007 min
Response: 744228
Conc: 425.38 ng/ml



#32 AR-1260-2


R.T.: 8.093 min
Delta R.T.: -0.001 min
Response: 1743157
Conc: 512.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 7.266 min
Delta R.T.: -0.007 min
Response: 862244
Conc: 483.69 ng/ml




#33 AR-1260-3

R.T.: 8.458 min
Delta R.T.: 0.000 min
Response: 1080866
Conc: 430.62 ng/ml



#33 AR-1260-3


R.T.: 7.422 min
Delta R.T.: -0.007 min
Response: 776143
Conc: 430.44 ng/ml



#34 AR-1260-4


R.T.: 8.688 min
Delta R.T.: 0.000 min
Response: 1307053
Conc: 445.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.900 min
Delta R.T.: -0.007 min
Response: 555685
Conc: 436.55 ng/ml



#35 AR-1260-5

R.T.: 9.005 min
Delta R.T.: -0.001 min
Response: 2552246
Conc: 518.79 ng/ml



#35 AR-1260-5

R.T.: 8.144 min
Delta R.T.: -0.008 min
Response: 1237269
Conc: 508.90 ng/ml