

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042423\
 Data File : PP057172.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Apr 2023 13:11
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 24 22:37:29 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 20 16:59:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.404	3.632	105.3E6	79112707	47.613	46.831
2) SA Decachlor...	10.222	8.690	80709426	78964354	50.293	44.140
Target Compounds						
3) L1 AR-1016-1	5.580	4.729	29425728	27613327	469.060	486.103
4) L1 AR-1016-2	5.603	4.748	43060098	38931394	471.882	485.849
5) L1 AR-1016-3	5.665	4.927	27831971	21236719	482.575	488.900
6) L1 AR-1016-4	5.764	4.968	21406075	18784039	471.061	484.066
7) L1 AR-1016-5	6.061	5.183	23293487	24676942	483.435	492.170
31) L7 AR-1260-1	7.195	6.224	43676722	49153337	501.484	485.064
32) L7 AR-1260-2	7.454	6.413	46369512	56202043	504.936	483.531
33) L7 AR-1260-3	7.815	6.569	37965544	52717014	523.180	476.528
34) L7 AR-1260-4	8.041	7.043	42949769	43473682	511.120	474.608
35) L7 AR-1260-5	8.365	7.285	75073488	83058188	523.120	460.172



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042423\
Data File : PP057172.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Apr 2023 13:11
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 24 22:37:29 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 20 16:59:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.404 min
Delta R.T.: 0.000 min
Response: 105320832
Conc: 47.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.632 min
Delta R.T.: 0.001 min
Response: 79112707
Conc: 46.83 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.222 min
Delta R.T.: 0.005 min
Response: 80709426
Conc: 50.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.690 min
Delta R.T.: 0.004 min
Response: 78964354
Conc: 44.14 ng/ml



#3 AR-1016-1


R.T.: 5.580 min
Delta R.T.: 0.001 min
Response: 29425728
Conc: 469.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.729 min
Delta R.T.: 0.001 min
Response: 27613327
Conc: 486.10 ng/ml




#4 AR-1016-2

R.T.: 5.603 min
Delta R.T.: 0.001 min
Response: 43060098
Conc: 471.88 ng/ml



#4 AR-1016-2


R.T.: 4.748 min
Delta R.T.: 0.001 min
Response: 38931394
Conc: 485.85 ng/ml



#5 AR-1016-3


R.T.: 5.665 min
Delta R.T.: 0.000 min
Response: 27831971
Conc: 482.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.927 min
Delta R.T.: 0.001 min
Response: 21236719
Conc: 488.90 ng/ml




#6 AR-1016-4

R.T.: 5.764 min
Delta R.T.: 0.001 min
Response: 21406075
Conc: 471.06 ng/ml



#6 AR-1016-4


R.T.: 4.968 min
Delta R.T.: 0.001 min
Response: 18784039
Conc: 484.07 ng/ml



#7 AR-1016-5


R.T.: 6.061 min
Delta R.T.: 0.001 min
Response: 23293487
Conc: 483.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.183 min
Delta R.T.: 0.001 min
Response: 24676942
Conc: 492.17 ng/ml




#31 AR-1260-1

R.T.: 7.195 min
Delta R.T.: 0.002 min
Response: 43676722
Conc: 501.48 ng/ml



#31 AR-1260-1


R.T.: 6.224 min
Delta R.T.: 0.002 min
Response: 49153337
Conc: 485.06 ng/ml



#32 AR-1260-2


R.T.: 7.454 min
Delta R.T.: 0.003 min
Response: 46369512
Conc: 504.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.413 min
Delta R.T.: 0.003 min
Response: 56202043
Conc: 483.53 ng/ml




#33 AR-1260-3

R.T.: 7.815 min
Delta R.T.: 0.003 min
Response: 37965544
Conc: 523.18 ng/ml



#33 AR-1260-3


R.T.: 6.569 min
Delta R.T.: 0.003 min
Response: 52717014
Conc: 476.53 ng/ml



#34 AR-1260-4


R.T.: 8.041 min
Delta R.T.: 0.002 min
Response: 42949769
Conc: 511.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.043 min
Delta R.T.: 0.003 min
Response: 43473682
Conc: 474.61 ng/ml



#35 AR-1260-5

R.T.: 8.365 min
Delta R.T.: 0.003 min
Response: 75073488
Conc: 523.12 ng/ml



#35 AR-1260-5

R.T.: 7.285 min
Delta R.T.: 0.004 min
Response: 83058188
Conc: 460.17 ng/ml