

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042425\
 Data File : PP071439.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Apr 2025 17:35
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 25 00:26:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.812	100.1E6	71056265	50.467	50.443
2) SA Decachlor...	10.230	8.846	75200506	45217011	51.934	51.556
Target Compounds						
16) L4 AR-1242-1	5.667	4.897	29151449	25492118	532.239	551.257
17) L4 AR-1242-2	5.688	4.915	45097016	36030300	530.675	545.726
18) L4 AR-1242-3	5.750	5.093	26526857	20178338	518.880	553.616
19) L4 AR-1242-4	5.847	5.176	22167708	19472942	525.195	545.299
20) L4 AR-1242-5	6.579	5.700	23647385	23354263	510.847	536.780



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042425\
Data File : PP071439.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Apr 2025 17:35
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 25 00:26:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: -0.004 min
Response: 100057300
Conc: 50.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 71056265
Conc: 50.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.230 min
Delta R.T.: -0.008 min
Response: 75200506
Conc: 51.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.846 min
Delta R.T.: -0.003 min
Response: 45217011
Conc: 51.56 ng/ml



#16 AR-1242-1


R.T.: 5.667 min
Delta R.T.: -0.002 min
Response: 29151449
Conc: 532.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 25492118
Conc: 551.26 ng/ml




#17 AR-1242-2

R.T.: 5.688 min
Delta R.T.: -0.001 min
Response: 45097016
Conc: 530.67 ng/ml



#17 AR-1242-2


R.T.: 4.915 min
Delta R.T.: 0.000 min
Response: 36030300
Conc: 545.73 ng/ml



#18 AR-1242-3


R.T.: 5.750 min
Delta R.T.: -0.002 min
Response: 26526857
Conc: 518.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.093 min
Delta R.T.: 0.000 min
Response: 20178338
Conc: 553.62 ng/ml




#19 AR-1242-4

R.T.: 5.847 min
Delta R.T.: -0.002 min
Response: 22167708
Conc: 525.19 ng/ml



#19 AR-1242-4


R.T.: 5.176 min
Delta R.T.: -0.001 min
Response: 19472942
Conc: 545.30 ng/ml



#20 AR-1242-5

R.T.: 6.579 min
Delta R.T.: -0.002 min
Response: 23647385
Conc: 510.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.700 min
Delta R.T.: -0.001 min
Response: 23354263
Conc: 536.78 ng/ml