

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042519\
 Data File : PP029996.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Apr 2019 15:33
 Operator : EI/SJ
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 26 10:01:26 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042519.M
 Quant Title : 8080.M
 QLast Update : Fri Apr 26 09:59:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.149	12.219	421.1E6	386.7E6	1009.456	1028.005
Target Compounds							
1) T	Dalapon	3.231	2.980	723.1E6	495.2E6	953.418	952.104
2) T	3,5-DICHL...	10.537	10.308	683.2E6	469.1E6	954.511	951.346
3) T	4-Nitroph...	11.745	11.376	349.1E6	192.0E6	948.614	933.328
5) T	DICAMBA	12.473	12.526	1853.1E6	1425.4E6	962.064	966.640
6) T	MCPD	12.808	12.720	103.5E6	73703557	96.815	97.045
7) T	MCPA	13.066	13.114	149.1E6	116.3E6	94.886	96.993
8) T	DICHLORPROP	13.677	13.649	421.0E6	340.3E6	949.314	965.068
9) T	2,4-D	14.042	14.142	455.7E6	388.4E6	966.025	978.594
10) T	Pentachlo...	14.456	14.760	7235.7E6	5356.5E6	956.293	952.476
11) T	2,4,5-TP ...	15.318	15.326	2460.4E6	1970.3E6	969.362	979.581
12) T	2,4,5-T	15.723	15.973	2303.5E6	1953.0E6	985.068	988.154
13) T	2,4-DB	16.483	16.571	285.3E6	225.6E6	989.600	1011.371
14) T	DINOSEB	18.003	16.995	1519.6E6	1224.6E6	995.575	1010.668
15) T	Picloram	17.783	18.360	3541.3E6	2745.0E6	1042.322	1055.887
16) T	DCPA	18.355	18.240	3357.7E6	2581.5E6	993.291	1000.960



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042519\
Data File : PP029996.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Apr 2019 15:33
Operator : EI/SJ
Sample : HSTDICC1000
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HSTDICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 26 10:01:26 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042519.M
Quant Title : 8080.M
QLast Update : Fri Apr 26 09:59:52 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.231 min
Delta R.T.: 0.000 min
Response: 723098473
Conc: 953.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDICC1000




#1 Dalapon

R.T.: 2.980 min
Delta R.T.: 0.001 min
Response: 495170493
Conc: 952.10 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 10.537 min
Delta R.T.: -0.001 min
Response: 683225699
Conc: 954.51 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 10.308 min
Delta R.T.: 0.000 min
Response: 469070875
Conc: 951.35 ng/ml



#3 4-Nitrophenol


R.T.: 11.745 min
Delta R.T.: -0.002 min
Response: 349136678
Conc: 948.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDICC1000




#3 4-Nitrophenol

R.T.: 11.376 min
Delta R.T.: 0.000 min
Response: 191989938
Conc: 933.33 ng/ml




#4 2,4-DCAA

R.T.: 12.149 min
Delta R.T.: 0.000 min
Response: 421110215
Conc: 1009.46 ng/ml



#4 2,4-DCAA


R.T.: 12.219 min
Delta R.T.: 0.000 min
Response: 386745902
Conc: 1028.01 ng/ml



#5 DICAMBA


R.T.: 12.473 min
Delta R.T.: 0.000 min
Response: 1853105784
Conc: 962.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDICC1000




#5 DICAMBA

R.T.: 12.526 min
Delta R.T.: 0.001 min
Response: 1425356900
Conc: 966.64 ng/ml




#6 MCP

R.T.: 12.808 min
Delta R.T.: 0.002 min
Response: 103531164
Conc: 96.81 ug/ml



#6 MCP


R.T.: 12.720 min
Delta R.T.: 0.003 min
Response: 73703557
Conc: 97.05 ug/ml



#7 MCPA


R.T.: 13.066 min
Delta R.T.: 0.002 min
Response: 149097461
Conc: 94.89 ug/ml

Instrument :
ECD_P
ClientSampleId :
HSTDICC1000




#7 MCPA

R.T.: 13.114 min
Delta R.T.: 0.003 min
Response: 116304543
Conc: 96.99 ug/ml




#8 DICHLORPROP

R.T.: 13.677 min
Delta R.T.: 0.000 min
Response: 420992324
Conc: 949.31 ng/ml



#8 DICHLORPROP


R.T.: 13.649 min
Delta R.T.: 0.000 min
Response: 340344014
Conc: 965.07 ng/ml



#9 2,4-D


R.T.: 14.042 min
Delta R.T.: -0.001 min
Response: 455664479
Conc: 966.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDICC1000




#9 2,4-D

R.T.: 14.142 min
Delta R.T.: 0.000 min
Response: 388352714
Conc: 978.59 ng/ml




#10 Pentachlorophenol

R.T.: 14.456 min
Delta R.T.: -0.001 min
Response: 7235715755
Conc: 956.29 ng/ml



#10 Pentachlorophenol


R.T.: 14.760 min
Delta R.T.: 0.000 min
Response: 5356533582
Conc: 952.48 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 15.318 min
Delta R.T.: -0.001 min
Response: 2460441704
Conc: 969.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDICC1000




#11 2,4,5-TP (SILVEX)

R.T.: 15.326 min
Delta R.T.: 0.000 min
Response: 1970337314
Conc: 979.58 ng/ml




#12 2,4,5-T

R.T.: 15.723 min
Delta R.T.: -0.002 min
Response: 2303543519
Conc: 985.07 ng/ml



#12 2,4,5-T


R.T.: 15.973 min
Delta R.T.: 0.000 min
Response: 1953009786
Conc: 988.15 ng/ml



#13 2,4-DB


R.T.: 16.483 min
Delta R.T.: 0.000 min
Response: 285310908
Conc: 989.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDICC1000




#13 2,4-DB

R.T.: 16.571 min
Delta R.T.: 0.000 min
Response: 225556767
Conc: 1011.37 ng/ml




#14 DINOSEB

R.T.: 18.003 min
Delta R.T.: 0.000 min
Response: 1519583785
Conc: 995.58 ng/ml



#14 DINOSEB


R.T.: 16.995 min
Delta R.T.: 0.001 min
Response: 1224584756
Conc: 1010.67 ng/ml



#15 Picloram


R.T.: 17.783 min
Delta R.T.: 0.000 min
Response: 3541295491
Conc: 1042.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDICC1000




#15 Picloram

R.T.: 18.360 min
Delta R.T.: 0.000 min
Response: 2744985079
Conc: 1055.89 ng/ml



#16 DCPA

R.T.: 18.355 min
Delta R.T.: -0.001 min
Response: 3357673341
Conc: 993.29 ng/ml



#16 DCPA

R.T.: 18.240 min
Delta R.T.: 0.000 min
Response: 2581506913
Conc: 1000.96 ng/ml