

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042522\
 Data File : PP046615.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Apr 2022 19:58
 Operator : YP\AJ
 Sample : N2532-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 26 02:08:37 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 22 19:00:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.863	3.127	57459730	82486265	23.803	25.145
2) SA Decachlor...	9.946	8.448	36715158	63967544	24.951	26.817
Target Compounds						
3) L1 AR-1016-1	5.111	4.250	380638	10126	5.055	0.088 #
4) L1 AR-1016-2	5.142	4.280	584013	188334	5.129	1.110 #
5) L1 AR-1016-3	5.203	4.470	732465	81789	10.246	0.934 #
6) L1 AR-1016-4	5.298	4.516	682634	66059	11.354	0.949 #
7) L1 AR-1016-5	5.619	4.740	636655	147692	12.024	1.736 #
8) L2 AR-1221-1	4.084	3.356	11091926	431218	402.196	9.766 #
9) L2 AR-1221-2	4.185	3.460f	723967	564385	35.658	16.710 #
10) L2 AR-1221-3	4.265	3.520	135302	85061	2.076	0.856 #
11) L3 AR-1232-1	4.265	3.520	135302	85061	2.763	1.121 #
12) L3 AR-1232-2	4.822	4.280	274318	188334	11.301	2.612 #
13) L3 AR-1232-3	5.142	4.470	584013	81789	12.365	2.246 #
14) L3 AR-1232-4	5.298	4.558	682634	65256	28.055	2.048 #
15) L3 AR-1232-5	5.411	4.740	831092	147692	49.950	5.127 #
16) L4 AR-1242-1	5.111	4.250	380638	10126	6.003	0.103 #
17) L4 AR-1242-2	5.142	4.280	584013	188334	6.130	1.300 #
18) L4 AR-1242-3	5.203	4.470	732465	81789	12.235	1.101 #
19) L4 AR-1242-4	5.317	4.558	524248	65256	10.365	0.914 #
20) L4 AR-1242-5	6.113	5.104	152908	363844	3.034	3.771
21) L5 AR-1248-1	5.111	4.250	380638	10126	9.107	0.147 #
22) L5 AR-1248-2	5.411	4.516	831092	66059	14.062	0.707 #
23) L5 AR-1248-3	5.619	4.558	636655	65256	9.114	0.668 #
24) L5 AR-1248-4	6.067	4.734	25862	300345	0.337	2.551 #
25) L5 AR-1248-5	6.113	5.154	152908	179031	1.966	1.469 #
26) L6 AR-1254-1	6.037	5.104	67583	363844	1.164	1.886 #
27) L6 AR-1254-2	6.278	5.266	341827	230815	2.774	1.386 #
28) L6 AR-1254-3	6.674	5.696	363780	567088	2.749	2.274
29) L6 AR-1254-4	6.979	5.952	263232	436985	2.648	2.547
30) L6 AR-1254-5	7.447	6.391	132984	595142	1.512	2.727 #
31) L7 AR-1260-1	6.852	5.831	82716	476997	0.980	2.998 #
32) L7 AR-1260-2	7.143	6.030f	45852	871950	0.455	4.287 #
33) L7 AR-1260-3	7.533	6.219	72941	1290375	0.900	7.554 #
34) L7 AR-1260-4	7.772	6.710	228854	311126	2.873	2.141 #
35) L7 AR-1260-5	8.103	6.973	87249	323993	0.549	0.923 #
36) L8 AR-1262-1	7.533	6.430	72941	287686	0.603	1.165 #
37) L8 AR-1262-2	8.103	6.710	87249	311126	0.454	1.610 #
38) L8 AR-1262-3	8.440	7.289	128341	697174	0.956	4.294 #
39) L8 AR-1262-4	8.537	7.355	134503	54774	2.136	0.191 #
40) L8 AR-1262-5	9.208	7.904	56867	40803	0.824	0.311 #
41) L9 AR-1268-1	8.440	7.289	128341	697174	0.534	1.455 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042522\
 Data File : PP046615.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Apr 2022 19:58
 Operator : YP\AJ
 Sample : N2532-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 26 02:08:37 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 22 19:00:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.537	7.355	134503	54774	0.607	0.129 #
43) L9	AR-1268-3	8.773	7.586	105875	361148	0.551	1.021 #
44) L9	AR-1268-4	9.208	7.904	56867	40803	0.737	0.277 #
45) L9	AR-1268-5	9.623	8.197	177757	202196	0.317	0.198 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042522\
Data File : PP046615.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Apr 2022 19:58
Operator : YP\AJ
Sample : N2532-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 26 02:08:37 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 22 19:00:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.863 min
Delta R.T.: -0.004 min
Response: 57459730
Conc: 23.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.127 min
Delta R.T.: -0.002 min
Response: 82486265
Conc: 25.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.946 min
Delta R.T.: -0.009 min
Response: 36715158
Conc: 24.95 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.448 min
Delta R.T.: -0.008 min
Response: 63967544
Conc: 26.82 ng/ml



#3 AR-1016-1


R.T.: 5.111 min
Delta R.T.: -0.001 min
Response: 380638
Conc: 5.05 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.250 min
Delta R.T.: -0.014 min
Response: 10126
Conc: 0.09 ng/ml




#4 AR-1016-2

R.T.: 5.142 min
Delta R.T.: 0.005 min
Response: 584013
Conc: 5.13 ng/ml



#4 AR-1016-2


R.T.: 4.280 min
Delta R.T.: -0.002 min
Response: 188334
Conc: 1.11 ng/ml



#5 AR-1016-3


R.T.: 5.203 min
Delta R.T.: 0.002 min
Response: 732465
Conc: 10.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.470 min
Delta R.T.: 0.004 min
Response: 81789
Conc: 0.93 ng/ml




#6 AR-1016-4

R.T.: 5.298 min
Delta R.T.: -0.008 min
Response: 682634
Conc: 11.35 ng/ml



#6 AR-1016-4


R.T.: 4.516 min
Delta R.T.: 0.000 min
Response: 66059
Conc: 0.95 ng/ml



#7 AR-1016-5


R.T.: 5.619 min
Delta R.T.: -0.006 min
Response: 636655
Conc: 12.02 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.740 min
Delta R.T.: 0.001 min
Response: 147692
Conc: 1.74 ng/ml




#8 AR-1221-1

R.T.: 4.084 min
Delta R.T.: -0.007 min
Response: 11091926
Conc: 402.20 ng/ml



#8 AR-1221-1


R.T.: 3.356 min
Delta R.T.: 0.002 min
Response: 431218
Conc: 9.77 ng/ml



#9 AR-1221-2


R.T.: 4.185 min
Delta R.T.: 0.002 min
Response: 723967
Conc: 35.66 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.460 min
Delta R.T.: 0.017 min
Response: 564385
Conc: 16.71 ng/ml




#10 AR-1221-3

R.T.: 4.265 min
Delta R.T.: 0.003 min
Response: 135302
Conc: 2.08 ng/ml



#10 AR-1221-3


R.T.: 3.520 min
Delta R.T.: -0.001 min
Response: 85061
Conc: 0.86 ng/ml



#11 AR-1232-1


R.T.: 4.265 min
Delta R.T.: 0.003 min
Response: 135302
Conc: 2.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.520 min
Delta R.T.: -0.002 min
Response: 85061
Conc: 1.12 ng/ml




#12 AR-1232-2

R.T.: 4.822 min
Delta R.T.: 0.000 min
Response: 274318
Conc: 11.30 ng/ml



#12 AR-1232-2


R.T.: 4.280 min
Delta R.T.: -0.002 min
Response: 188334
Conc: 2.61 ng/ml



#13 AR-1232-3


R.T.: 5.142 min
Delta R.T.: 0.007 min
Response: 584013
Conc: 12.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.470 min
Delta R.T.: 0.004 min
Response: 81789
Conc: 2.25 ng/ml




#14 AR-1232-4

R.T.: 5.298 min
Delta R.T.: -0.009 min
Response: 682634
Conc: 28.05 ng/ml



#14 AR-1232-4


R.T.: 4.558 min
Delta R.T.: 0.000 min
Response: 65256
Conc: 2.05 ng/ml



#15 AR-1232-5


R.T.: 5.411 min
Delta R.T.: 0.003 min
Response: 831092
Conc: 49.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.740 min
Delta R.T.: 0.002 min
Response: 147692
Conc: 5.13 ng/ml




#16 AR-1242-1

R.T.: 5.111 min
Delta R.T.: -0.003 min
Response: 380638
Conc: 6.00 ng/ml



#16 AR-1242-1


R.T.: 4.250 min
Delta R.T.: -0.014 min
Response: 10126
Conc: 0.10 ng/ml



#17 AR-1242-2


R.T.: 5.142 min
Delta R.T.: 0.004 min
Response: 584013
Conc: 6.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.280 min
Delta R.T.: -0.002 min
Response: 188334
Conc: 1.30 ng/ml




#18 AR-1242-3

R.T.: 5.203 min
Delta R.T.: 0.000 min
Response: 732465
Conc: 12.24 ng/ml



#18 AR-1242-3


R.T.: 4.470 min
Delta R.T.: 0.004 min
Response: 81789
Conc: 1.10 ng/ml



#19 AR-1242-4


R.T.: 5.317 min
Delta R.T.: 0.008 min
Response: 524248
Conc: 10.37 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.558 min
Delta R.T.: 0.000 min
Response: 65256
Conc: 0.91 ng/ml




#20 AR-1242-5

R.T.: 6.113 min
Delta R.T.: 0.007 min
Response: 152908
Conc: 3.03 ng/ml



#20 AR-1242-5


R.T.: 5.104 min
Delta R.T.: -0.009 min
Response: 363844
Conc: 3.77 ng/ml



#21 AR-1248-1


R.T.: 5.111 min
Delta R.T.: 0.000 min
Response: 380638
Conc: 9.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.250 min
Delta R.T.: -0.013 min
Response: 10126
Conc: 0.15 ng/ml




#22 AR-1248-2

R.T.: 5.411 min
Delta R.T.: 0.003 min
Response: 831092
Conc: 14.06 ng/ml



#22 AR-1248-2


R.T.: 4.516 min
Delta R.T.: 0.002 min
Response: 66059
Conc: 0.71 ng/ml



#23 AR-1248-3


R.T.: 5.619 min
Delta R.T.: -0.005 min
Response: 636655
Conc: 9.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.558 min
Delta R.T.: 0.000 min
Response: 65256
Conc: 0.67 ng/ml




#24 AR-1248-4

R.T.: 6.067 min
Delta R.T.: 0.002 min
Response: 25862
Conc: 0.34 ng/ml



#24 AR-1248-4


R.T.: 4.734 min
Delta R.T.: -0.002 min
Response: 300345
Conc: 2.55 ng/ml



#25 AR-1248-5


R.T.: 6.113 min
Delta R.T.: 0.007 min
Response: 152908
Conc: 1.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.154 min
Delta R.T.: -0.001 min
Response: 179031
Conc: 1.47 ng/ml




#26 AR-1254-1

R.T.: 6.037 min
Delta R.T.: 0.000 min
Response: 67583
Conc: 1.16 ng/ml



#26 AR-1254-1


R.T.: 5.104 min
Delta R.T.: -0.007 min
Response: 363844
Conc: 1.89 ng/ml



#27 AR-1254-2


R.T.: 6.278 min
Delta R.T.: 0.006 min
Response: 341827
Conc: 2.77 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.266 min
Delta R.T.: -0.005 min
Response: 230815
Conc: 1.39 ng/ml




#28 AR-1254-3

R.T.: 6.674 min
Delta R.T.: 0.001 min
Response: 363780
Conc: 2.75 ng/ml



#28 AR-1254-3


R.T.: 5.696 min
Delta R.T.: -0.005 min
Response: 567088
Conc: 2.27 ng/ml



#29 AR-1254-4


R.T.: 6.979 min
Delta R.T.: -0.007 min
Response: 263232
Conc: 2.65 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 5.952 min
Delta R.T.: 0.002 min
Response: 436985
Conc: 2.55 ng/ml




#30 AR-1254-5

R.T.: 7.447 min
Delta R.T.: 0.001 min
Response: 132984
Conc: 1.51 ng/ml



#30 AR-1254-5


R.T.: 6.391 min
Delta R.T.: -0.008 min
Response: 595142
Conc: 2.73 ng/ml



#31 AR-1260-1


R.T.: 6.852 min
Delta R.T.: 0.000 min
Response: 82716
Conc: 0.98 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.831 min
Delta R.T.: -0.009 min
Response: 476997
Conc: 3.00 ng/ml




#32 AR-1260-2

R.T.: 7.143 min
Delta R.T.: 0.008 min
Response: 45852
Conc: 0.46 ng/ml



#32 AR-1260-2


R.T.: 6.030 min
Delta R.T.: -0.017 min
Response: 871950
Conc: 4.29 ng/ml



#33 AR-1260-3


R.T.: 7.533 min
Delta R.T.: 0.005 min
Response: 72941
Conc: 0.90 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.219 min
Delta R.T.: 0.011 min
Response: 1290375
Conc: 7.55 ng/ml




#34 AR-1260-4

R.T.: 7.772 min
Delta R.T.: -0.005 min
Response: 228854
Conc: 2.87 ng/ml



#34 AR-1260-4


R.T.: 6.710 min
Delta R.T.: -0.007 min
Response: 311126
Conc: 2.14 ng/ml



#35 AR-1260-5


R.T.: 8.103 min
Delta R.T.: -0.012 min
Response: 87249
Conc: 0.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 6.973 min
Delta R.T.: -0.013 min
Response: 323993
Conc: 0.92 ng/ml




#36 AR-1262-1

R.T.: 7.533 min
Delta R.T.: 0.005 min
Response: 72941
Conc: 0.60 ng/ml



#36 AR-1262-1


R.T.: 6.430 min
Delta R.T.: -0.009 min
Response: 287686
Conc: 1.16 ng/ml



#37 AR-1262-2


R.T.: 8.103 min
Delta R.T.: -0.014 min
Response: 87249
Conc: 0.45 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.710 min
Delta R.T.: -0.007 min
Response: 311126
Conc: 1.61 ng/ml




#38 AR-1262-3

R.T.: 8.440 min
Delta R.T.: 0.000 min
Response: 128341
Conc: 0.96 ng/ml



#38 AR-1262-3


R.T.: 7.289 min
Delta R.T.: 0.000 min
Response: 697174
Conc: 4.29 ng/ml



#39 AR-1262-4


R.T.: 8.537 min
Delta R.T.: 0.001 min
Response: 134503
Conc: 2.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.355 min
Delta R.T.: -0.005 min
Response: 54774
Conc: 0.19 ng/ml




#40 AR-1262-5

R.T.: 9.208 min
Delta R.T.: 0.002 min
Response: 56867
Conc: 0.82 ng/ml



#40 AR-1262-5


R.T.: 7.904 min
Delta R.T.: 0.001 min
Response: 40803
Conc: 0.31 ng/ml



#41 AR-1268-1


R.T.: 8.440 min
Delta R.T.: 0.002 min
Response: 128341
Conc: 0.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.289 min
Delta R.T.: -0.001 min
Response: 697174
Conc: 1.45 ng/ml




#42 AR-1268-2

R.T.: 8.537 min
Delta R.T.: 0.000 min
Response: 134503
Conc: 0.61 ng/ml



#42 AR-1268-2


R.T.: 7.355 min
Delta R.T.: -0.007 min
Response: 54774
Conc: 0.13 ng/ml



#43 AR-1268-3


R.T.: 8.773 min
Delta R.T.: 0.005 min
Response: 105875
Conc: 0.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.586 min
Delta R.T.: 0.009 min
Response: 361148
Conc: 1.02 ng/ml




#44 AR-1268-4

R.T.: 9.208 min
Delta R.T.: 0.002 min
Response: 56867
Conc: 0.74 ng/ml



#44 AR-1268-4


R.T.: 7.904 min
Delta R.T.: 0.001 min
Response: 40803
Conc: 0.28 ng/ml



#45 AR-1268-5

R.T.: 9.623 min
Delta R.T.: 0.003 min
Response: 177757
Conc: 0.32 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.197 min
Delta R.T.: -0.005 min
Response: 202196
Conc: 0.20 ng/ml