

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042522\
 Data File : PP046647.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Apr 2022 08:32
 Operator : YP\AJ
 Sample : PB144357BSD
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB144357BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 26 11:38:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 22 19:00:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.862	3.126	55147119	79375056	22.845	24.197
2) SA Decachlor...	9.943	8.447	35414652	62860012	24.067	26.353
Target Compounds						
3) L1 AR-1016-1	5.107	4.259	3898443	6105898	51.771	53.148
4) L1 AR-1016-2	5.131	4.276	5866736	8258939	51.521	48.678
5) L1 AR-1016-3	5.196	4.460	4620343	4423356	64.634	50.526
6) L1 AR-1016-4	5.302	4.508	3586822	3621204	59.659	52.014
7) L1 AR-1016-5	5.621	4.731	3493174	4739185	65.975	55.721
31) L7 AR-1260-1	6.845	5.833	4992524	9488161	59.161	59.627
32) L7 AR-1260-2	7.128	6.040	5679159	12898888	56.402	63.412
33) L7 AR-1260-3	7.521	6.200	3639279	10729087	44.923	62.811 #
34) L7 AR-1260-4	7.769	6.710	3718968	8573912	46.691	58.991 #
35) L7 AR-1260-5	8.109	6.978	6927346	19977494	43.611	56.896 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042522\
Data File : PP046647.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Apr 2022 08:32
Operator : YP\AJ
Sample : PB144357BSD
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB144357BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 26 11:38:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 22 19:00:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.862 min
Delta R.T.: -0.005 min
Response: 55147119
Conc: 22.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144357BSD




#1 Tetrachloro-m-xylene

R.T.: 3.126 min
Delta R.T.: -0.003 min
Response: 79375056
Conc: 24.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.943 min
Delta R.T.: -0.012 min
Response: 35414652
Conc: 24.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.447 min
Delta R.T.: -0.009 min
Response: 62860012
Conc: 26.35 ng/ml



#3 AR-1016-1


R.T.: 5.107 min
Delta R.T.: -0.005 min
Response: 3898443
Conc: 51.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144357BSD




#3 AR-1016-1

R.T.: 4.259 min
Delta R.T.: -0.005 min
Response: 6105898
Conc: 53.15 ng/ml




#4 AR-1016-2

R.T.: 5.131 min
Delta R.T.: -0.006 min
Response: 5866736
Conc: 51.52 ng/ml



#4 AR-1016-2


R.T.: 4.276 min
Delta R.T.: -0.006 min
Response: 8258939
Conc: 48.68 ng/ml



#5 AR-1016-3


R.T.: 5.196 min
Delta R.T.: -0.005 min
Response: 4620343
Conc: 64.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144357BSD




#5 AR-1016-3

R.T.: 4.460 min
Delta R.T.: -0.006 min
Response: 4423356
Conc: 50.53 ng/ml




#6 AR-1016-4

R.T.: 5.302 min
Delta R.T.: -0.003 min
Response: 3586822
Conc: 59.66 ng/ml



#6 AR-1016-4


R.T.: 4.508 min
Delta R.T.: -0.007 min
Response: 3621204
Conc: 52.01 ng/ml



#7 AR-1016-5


R.T.: 5.621 min
Delta R.T.: -0.004 min
Response: 3493174
Conc: 65.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144357BSD




#7 AR-1016-5

R.T.: 4.731 min
Delta R.T.: -0.008 min
Response: 4739185
Conc: 55.72 ng/ml




#31 AR-1260-1

R.T.: 6.845 min
Delta R.T.: -0.007 min
Response: 4992524
Conc: 59.16 ng/ml



#31 AR-1260-1


R.T.: 5.833 min
Delta R.T.: -0.007 min
Response: 9488161
Conc: 59.63 ng/ml



#32 AR-1260-2


R.T.: 7.128 min
Delta R.T.: -0.007 min
Response: 5679159
Conc: 56.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144357BSD




#32 AR-1260-2

R.T.: 6.040 min
Delta R.T.: -0.007 min
Response: 12898888
Conc: 63.41 ng/ml




#33 AR-1260-3

R.T.: 7.521 min
Delta R.T.: -0.007 min
Response: 3639279
Conc: 44.92 ng/ml



#33 AR-1260-3


R.T.: 6.200 min
Delta R.T.: -0.008 min
Response: 10729087
Conc: 62.81 ng/ml



#34 AR-1260-4


R.T.: 7.769 min
Delta R.T.: -0.008 min
Response: 3718968
Conc: 46.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144357BSD




#34 AR-1260-4

R.T.: 6.710 min
Delta R.T.: -0.008 min
Response: 8573912
Conc: 58.99 ng/ml



#35 AR-1260-5

R.T.: 8.109 min
Delta R.T.: -0.006 min
Response: 6927346
Conc: 43.61 ng/ml



#35 AR-1260-5

R.T.: 6.978 min
Delta R.T.: -0.008 min
Response: 19977494
Conc: 56.90 ng/ml