

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042522\
 Data File : PP046652.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Apr 2022 10:46
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 26 13:57:02 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 22 19:00:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.862	3.127	132.6E6	192.7E6	54.947	58.733
2) SA Decachlor...	9.943	8.446	78902887	143.1E6	53.620	60.008
Target Compounds						
21) L5 AR-1248-1	5.107	4.258	23104716	39243834	552.809	571.103
22) L5 AR-1248-2	5.403	4.509	31774160	53003638	537.600	566.931
23) L5 AR-1248-3	5.620	4.551	37203185	55396391	532.605	567.456
24) L5 AR-1248-4	6.059	4.732	39567006	63933161	515.993	543.106
25) L5 AR-1248-5	6.099	5.148	38235586	67536324	491.595	554.209



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042522\
Data File : PP046652.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Apr 2022 10:46
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 26 13:57:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 22 19:00:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.862 min
Delta R.T.: -0.005 min
Response: 132638005
Conc: 54.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.127 min
Delta R.T.: -0.003 min
Response: 192667142
Conc: 58.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.943 min
Delta R.T.: -0.012 min
Response: 78902887
Conc: 53.62 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.446 min
Delta R.T.: -0.010 min
Response: 143137911
Conc: 60.01 ng/ml



#21 AR-1248-1


R.T.: 5.107 min
Delta R.T.: -0.005 min
Response: 23104716
Conc: 552.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.258 min
Delta R.T.: -0.005 min
Response: 39243834
Conc: 571.10 ng/ml




#22 AR-1248-2

R.T.: 5.403 min
Delta R.T.: -0.005 min
Response: 31774160
Conc: 537.60 ng/ml



#22 AR-1248-2


R.T.: 4.509 min
Delta R.T.: -0.005 min
Response: 53003638
Conc: 566.93 ng/ml



#23 AR-1248-3


R.T.: 5.620 min
Delta R.T.: -0.004 min
Response: 37203185
Conc: 532.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.551 min
Delta R.T.: -0.006 min
Response: 55396391
Conc: 567.46 ng/ml




#24 AR-1248-4

R.T.: 6.059 min
Delta R.T.: -0.006 min
Response: 39567006
Conc: 515.99 ng/ml



#24 AR-1248-4


R.T.: 4.732 min
Delta R.T.: -0.005 min
Response: 63933161
Conc: 543.11 ng/ml



#25 AR-1248-5

R.T.: 6.099 min
Delta R.T.: -0.006 min
Response: 38235586
Conc: 491.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.148 min
Delta R.T.: -0.007 min
Response: 67536324
Conc: 554.21 ng/ml