

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042621\
 Data File : PP035304.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Apr 2021 1:59
 Operator : DD\AJ
 Sample : PP042621ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 07:47:18 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 27 07:44:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	81.003	78.157 E3	3.5	98	0.00
2 SA Decachlorobiphenyl	71.428	70.358 E3	1.5	99	0.00
3 L1 AR-1016-1	2.271	2.143 E3	5.6	98	0.00
4 L1 AR-1016-2	3.866	3.608 E3	6.7	96	0.00
5 L1 AR-1016-3	2.559	2.402 E3	6.1	96	0.00
6 L1 AR-1016-4	1.954	1.889 E3	3.3	97	0.00
7 L1 AR-1016-5	1.907	1.849 E3	3.0	98	0.00
31 L7 AR-1260-1	3.609	3.517 E3	2.5	98	0.00
32 L7 AR-1260-2	4.005	3.888 E3	2.9	98	0.00
33 L7 AR-1260-3	2.793	2.786 E3	0.3	99	0.00
34 L7 AR-1260-4	3.339	3.304 E3	1.0	98	0.00
35 L7 AR-1260-5	6.265	6.043 E3	3.5	96	0.00

Signal #2

1 SA Tetrachloro-m-xylene	52.216	49.906 E3	4.4	97	0.00
2 SA Decachlorobiphenyl	27.610	26.114 E3	5.4	95	0.00
3 L1 AR-1016-1	1.099	1.049 E3	4.5	96	0.00
4 L1 AR-1016-2	2.222	2.005 E3	9.8	91	0.00
5 L1 AR-1016-3	1.005	0.931 E3	7.4	90	0.00
6 L1 AR-1016-4	938.373	869.676	7.3	92	0.00
7 L1 AR-1016-5	1.083	1.044 E3	3.6	95	0.00
31 L7 AR-1260-1	1.947	1.851 E3	4.9	96	0.00
32 L7 AR-1260-2	2.148	2.030 E3	5.5	94	0.00
33 L7 AR-1260-3	2.112	1.974 E3	6.5	95	0.00
34 L7 AR-1260-4	1.477	1.398 E3	5.3	95	0.00
35 L7 AR-1260-5	3.020	2.872 E3	4.9	94	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	885.563	0.000	100.0#	0#	-5.09#
9 L2 AR-1221-2	660.755	0.000	100.0#	0#	-5.19#
10 L2 AR-1221-3	1.986	0.000 E3	100.0#	0#	-5.27#
11 L3 AR-1232-1	1.494	0.000 E3	100.0#	0#	-5.27#
12 L3 AR-1232-2	859.051	0.000	100.0#	0#	-5.86#
13 L3 AR-1232-3	1.494	0.000 E3	100.0#	0#	-6.18#
14 L3 AR-1232-4	761.885	0.000	100.0#	0#	-6.35#
15 L3 AR-1232-5	527.521	0.000	100.0#	0#	-6.44#
16 L4 AR-1242-1	1.671	0.000 E3	100.0#	0#	-6.15#
17 L4 AR-1242-2	2.846	0.000 E3	100.0#	0#	-6.17#
18 L4 AR-1242-3	1.868	0.000 E3	100.0#	0#	-6.24#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042621\
 Data File : PP035304.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Apr 2021 1:59
 Operator : DD\AJ
 Sample : PP042621ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 07:47:18 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 27 07:44:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
19 L4 AR-1242-4	1.432	0.000 E3	100.0#	0#	-6.35#
20 L4 AR-1242-5	1.381	0.000 E3	100.0#	0#	-7.12#
21 L5 AR-1248-1	1.304	0.000 E3	100.0#	0#	-6.15#
22 L5 AR-1248-2	2.040	0.000 E3	100.0#	0#	-6.44#
23 L5 AR-1248-3	2.421	0.000 E3	100.0#	0#	-6.66#
24 L5 AR-1248-4	2.258	0.000 E3	100.0#	0#	-7.08#
25 L5 AR-1248-5	2.362	0.000 E3	100.0#	0#	-7.12#
26 L6 AR-1254-1	2.715	0.000 E3	100.0#	0#	-7.05#
27 L6 AR-1254-2	4.136	0.000 E3	100.0#	0#	-7.28#
28 L6 AR-1254-3	4.210	0.000 E3	100.0#	0#	-7.66#
29 L6 AR-1254-4	2.716	0.000 E3	100.0#	0#	-7.96#
30 L6 AR-1254-5	3.378	0.000 E3	100.0#	0#	-8.38#
36 L8 AR-1262-1	4.247	0.000 E3	100.0#	0#	-8.45#
37 L8 AR-1262-2	7.164	0.000 E3	100.0#	0#	-9.00#
38 L8 AR-1262-3	5.148	0.000 E3	100.0#	0#	-9.30#
39 L8 AR-1262-4	4.384	0.000 E3	100.0#	0#	-9.39#
40 L8 AR-1262-5	2.633	0.000 E3	100.0#	0#	-10.03#
41 L9 AR-1268-1	9.549	0.000 E3	100.0#	0#	-9.30#
42 L9 AR-1268-2	9.372	0.000 E3	100.0#	0#	-9.39#
43 L9 AR-1268-3	8.423	0.000 E3	100.0#	0#	-9.60#
44 L9 AR-1268-4	2.889	0.000 E3	100.0#	0#	-10.03#
45 L9 AR-1268-5	24.893	0.000 E3	100.0#	0#	-10.42#

Signal #2

8 L2 AR-1221-1	533.184	0.000	100.0#	0#	-4.50#
9 L2 AR-1221-2	391.379	0.000	100.0#	0#	-4.60#
10 L2 AR-1221-3	1.204	0.000 E3	100.0#	0#	-4.69#
11 L3 AR-1232-1	926.048	0.000	100.0#	0#	-4.69#
12 L3 AR-1232-2	844.123	0.000	100.0#	0#	-5.51#
13 L3 AR-1232-3	393.526	0.000	100.0#	0#	-5.71#
14 L3 AR-1232-4	385.170	0.000	100.0#	0#	-5.80#
15 L3 AR-1232-5	390.170	0.000	100.0#	0#	-5.98#
16 L4 AR-1242-1	806.555	0.000	100.0#	0#	-5.49#
17 L4 AR-1242-2	1.636	0.000 E3	100.0#	0#	-5.51#
18 L4 AR-1242-3	744.479	0.000	100.0#	0#	-5.70#
19 L4 AR-1242-4	840.679	0.000	100.0#	0#	-5.79#
20 L4 AR-1242-5	836.519	0.000	100.0#	0#	-6.36#
21 L5 AR-1248-1	656.749	0.000	100.0#	0#	-5.49#
22 L5 AR-1248-2	1.145	0.000 E3	100.0#	0#	-5.75#
23 L5 AR-1248-3	1.179	0.000 E3	100.0#	0#	-5.79#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042621\
 Data File : PP035304.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Apr 2021 1:59
 Operator : DD\AJ
 Sample : PP042621ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 07:47:18 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 27 07:44:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	1.346	0.000 E3	100.0#	0#	-5.98#
25 L5 AR-1248-5	1.217	0.000 E3	100.0#	0#	-6.40#
26 L6 AR-1254-1	1.856	0.000 E3	100.0#	0#	-6.35#
27 L6 AR-1254-2	1.659	0.000 E3	100.0#	0#	-6.51#
28 L6 AR-1254-3	2.321	0.000 E3	100.0#	0#	-6.93#
29 L6 AR-1254-4	1.053	0.000 E3	100.0#	0#	-7.17#
30 L6 AR-1254-5	2.068	0.000 E3	100.0#	0#	-7.59#
36 L8 AR-1262-1	960.538	0.000	100.0#	0#	-7.59#
37 L8 AR-1262-2	3.273	0.000 E3	100.0#	0#	-8.13#
38 L8 AR-1262-3	1.485	0.000 E3	100.0#	0#	-8.42#
39 L8 AR-1262-4	2.701	0.000 E3	100.0#	0#	-8.48#
40 L8 AR-1262-5	1.047	0.000 E3	100.0#	0#	-8.98#
41 L9 AR-1268-1	4.252	0.000 E3	100.0#	0#	-8.42#
42 L9 AR-1268-2	4.103	0.000 E3	100.0#	0#	-8.48#
43 L9 AR-1268-3	3.584	0.000 E3	100.0#	0#	-8.68#
44 L9 AR-1268-4	1.175	0.000 E3	100.0#	0#	-8.98#
45 L9 AR-1268-5	9.602	0.000 E3	100.0#	0#	-9.27#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0