

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042621\
 Data File : PP035307.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Apr 2021 2:49
 Operator : DD\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 07:48:46 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 27 07:44:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	81.003	77.393 E3	4.5	97	0.00
2 SA Decachlorobiphenyl	71.428	67.986 E3	4.8	95	0.00
26 L6 AR-1254-1	2.715	2.668 E3	1.7	100	0.00
27 L6 AR-1254-2	4.136	4.029 E3	2.6	99	0.00
28 L6 AR-1254-3	4.210	4.186 E3	0.6	98	0.00
29 L6 AR-1254-4	2.716	2.625 E3	3.4	99	0.00
30 L6 AR-1254-5	3.378	3.389 E3	-0.3	98	0.00

Signal #2

1 SA Tetrachloro-m-xylene	52.216	49.432 E3	5.3	96	0.00
2 SA Decachlorobiphenyl	27.610	25.231 E3	8.6	92	0.00
26 L6 AR-1254-1	1.856	1.828 E3	1.5	99	0.00
27 L6 AR-1254-2	1.659	1.630 E3	1.7	99	0.00
28 L6 AR-1254-3	2.321	2.346 E3	-1.1	98	0.00
29 L6 AR-1254-4	1.053	1.039 E3	1.3	95	0.00
30 L6 AR-1254-5	2.068	2.025 E3	2.1	96	0.00

Evaluate Continuing Calibration Report - Not Found

3 L1 AR-1016-1	2.271	0.000 E3	100.0#	0#	-6.15#
4 L1 AR-1016-2	3.866	0.000 E3	100.0#	0#	-6.17#
5 L1 AR-1016-3	2.559	0.000 E3	100.0#	0#	-6.24#
6 L1 AR-1016-4	1.954	0.000 E3	100.0#	0#	-6.35#
7 L1 AR-1016-5	1.907	0.000 E3	100.0#	0#	-6.66#
8 L2 AR-1221-1	885.563	0.000	100.0#	0#	-5.09#
9 L2 AR-1221-2	660.755	0.000	100.0#	0#	-5.19#
10 L2 AR-1221-3	1.986	0.000 E3	100.0#	0#	-5.27#
11 L3 AR-1232-1	1.494	0.000 E3	100.0#	0#	-5.27#
12 L3 AR-1232-2	859.051	0.000	100.0#	0#	-5.86#
13 L3 AR-1232-3	1.494	0.000 E3	100.0#	0#	-6.18#
14 L3 AR-1232-4	761.885	0.000	100.0#	0#	-6.35#
15 L3 AR-1232-5	527.521	0.000	100.0#	0#	-6.44#
16 L4 AR-1242-1	1.671	0.000 E3	100.0#	0#	-6.15#
17 L4 AR-1242-2	2.846	0.000 E3	100.0#	0#	-6.17#
18 L4 AR-1242-3	1.868	0.000 E3	100.0#	0#	-6.24#
19 L4 AR-1242-4	1.432	0.000 E3	100.0#	0#	-6.35#
20 L4 AR-1242-5	1.381	0.000 E3	100.0#	0#	-7.12#
21 L5 AR-1248-1	1.304	0.000 E3	100.0#	0#	-6.15#
22 L5 AR-1248-2	2.040	0.000 E3	100.0#	0#	-6.44#
23 L5 AR-1248-3	2.421	0.000 E3	100.0#	0#	-6.66#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042621\
 Data File : PP035307.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Apr 2021 2:49
 Operator : DD\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 07:48:46 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 27 07:44:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	2.258	0.000 E3	100.0#	0#	-7.08#
25 L5 AR-1248-5	2.362	0.000 E3	100.0#	0#	-7.12#
31 L7 AR-1260-1	3.609	0.000 E3	100.0#	0#	-7.83#
32 L7 AR-1260-2	4.005	0.000 E3	100.0#	0#	-8.09#
33 L7 AR-1260-3	2.793	0.000 E3	100.0#	0#	-8.46#
34 L7 AR-1260-4	3.339	0.000 E3	100.0#	0#	-8.69#
35 L7 AR-1260-5	6.265	0.000 E3	100.0#	0#	-9.00#
36 L8 AR-1262-1	4.247	0.000 E3	100.0#	0#	-8.45#
37 L8 AR-1262-2	7.164	0.000 E3	100.0#	0#	-9.00#
38 L8 AR-1262-3	5.148	0.000 E3	100.0#	0#	-9.30#
39 L8 AR-1262-4	4.384	0.000 E3	100.0#	0#	-9.39#
40 L8 AR-1262-5	2.633	0.000 E3	100.0#	0#	-10.03#
41 L9 AR-1268-1	9.549	0.000 E3	100.0#	0#	-9.30#
42 L9 AR-1268-2	9.372	0.000 E3	100.0#	0#	-9.39#
43 L9 AR-1268-3	8.423	0.000 E3	100.0#	0#	-9.60#
44 L9 AR-1268-4	2.889	0.000 E3	100.0#	0#	-10.03#
45 L9 AR-1268-5	24.893	0.000 E3	100.0#	0#	-10.42#

Signal #2

3 L1 AR-1016-1	1.099	0.000 E3	100.0#	0#	-5.49#
4 L1 AR-1016-2	2.222	0.000 E3	100.0#	0#	-5.51#
5 L1 AR-1016-3	1.005	0.000 E3	100.0#	0#	-5.70#
6 L1 AR-1016-4	938.373	0.000	100.0#	0#	-5.75#
7 L1 AR-1016-5	1.083	0.000 E3	100.0#	0#	-5.98#
8 L2 AR-1221-1	533.184	0.000	100.0#	0#	-4.50#
9 L2 AR-1221-2	391.379	0.000	100.0#	0#	-4.60#
10 L2 AR-1221-3	1.204	0.000 E3	100.0#	0#	-4.69#
11 L3 AR-1232-1	926.048	0.000	100.0#	0#	-4.69#
12 L3 AR-1232-2	844.123	0.000	100.0#	0#	-5.51#
13 L3 AR-1232-3	393.526	0.000	100.0#	0#	-5.71#
14 L3 AR-1232-4	385.170	0.000	100.0#	0#	-5.80#
15 L3 AR-1232-5	390.170	0.000	100.0#	0#	-5.98#
16 L4 AR-1242-1	806.555	0.000	100.0#	0#	-5.49#
17 L4 AR-1242-2	1.636	0.000 E3	100.0#	0#	-5.51#
18 L4 AR-1242-3	744.479	0.000	100.0#	0#	-5.70#
19 L4 AR-1242-4	840.679	0.000	100.0#	0#	-5.79#
20 L4 AR-1242-5	836.519	0.000	100.0#	0#	-6.36#
21 L5 AR-1248-1	656.749	0.000	100.0#	0#	-5.49#
22 L5 AR-1248-2	1.145	0.000 E3	100.0#	0#	-5.75#
23 L5 AR-1248-3	1.179	0.000 E3	100.0#	0#	-5.79#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042621\
 Data File : PP035307.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Apr 2021 2:49
 Operator : DD\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 07:48:46 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 27 07:44:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	1.346	0.000 E3	100.0#	0#	-5.98#
25 L5 AR-1248-5	1.217	0.000 E3	100.0#	0#	-6.40#
31 L7 AR-1260-1	1.947	0.000 E3	100.0#	0#	-7.06#
32 L7 AR-1260-2	2.148	0.000 E3	100.0#	0#	-7.26#
33 L7 AR-1260-3	2.112	0.000 E3	100.0#	0#	-7.41#
34 L7 AR-1260-4	1.477	0.000 E3	100.0#	0#	-7.89#
35 L7 AR-1260-5	3.020	0.000 E3	100.0#	0#	-8.14#
36 L8 AR-1262-1	960.538	0.000	100.0#	0#	-7.59#
37 L8 AR-1262-2	3.273	0.000 E3	100.0#	0#	-8.13#
38 L8 AR-1262-3	1.485	0.000 E3	100.0#	0#	-8.42#
39 L8 AR-1262-4	2.701	0.000 E3	100.0#	0#	-8.48#
40 L8 AR-1262-5	1.047	0.000 E3	100.0#	0#	-8.98#
41 L9 AR-1268-1	4.252	0.000 E3	100.0#	0#	-8.42#
42 L9 AR-1268-2	4.103	0.000 E3	100.0#	0#	-8.48#
43 L9 AR-1268-3	3.584	0.000 E3	100.0#	0#	-8.68#
44 L9 AR-1268-4	1.175	0.000 E3	100.0#	0#	-8.98#
45 L9 AR-1268-5	9.602	0.000 E3	100.0#	0#	-9.27#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0