

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042621\
 Data File : PP035290.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Apr 2021 22:03
 Operator : DD\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 06:11:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 27 06:10:21 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.838	4.241	5719666	3624883	73.879	72.989
2) SA Decachlor...	10.738	9.522	4994148	1857677	72.351	71.537
Target Compounds						
26) L6 AR-1254-1	7.051	6.354	1927670	1328025	720.096	718.730
27) L6 AR-1254-2	7.280	6.511	2951011	1180483	725.837	716.884
28) L6 AR-1254-3	7.660	6.931	3104622	1722604	728.976	721.082
29) L6 AR-1254-4	7.956	7.169	1916706	762572	724.960	701.184
30) L6 AR-1254-5	8.381	7.591	2501043	1530949	726.787	725.249



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042621\
Data File : PP035290.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Apr 2021 22:03
Operator : DD\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 27 06:11:04 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042621.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 27 06:10:21 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.838 min
Delta R.T.: 0.000 min
Response: 5719666
Conc: 73.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#1 Tetrachloro-m-xylene

R.T.: 4.241 min
Delta R.T.: 0.000 min
Response: 3624883
Conc: 72.99 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.738 min
Delta R.T.: 0.000 min
Response: 4994148
Conc: 72.35 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.522 min
Delta R.T.: 0.000 min
Response: 1857677
Conc: 71.54 ng/ml



#26 AR-1254-1


R.T.: 7.051 min
Delta R.T.: 0.000 min
Response: 1927670
Conc: 720.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#26 AR-1254-1

R.T.: 6.354 min
Delta R.T.: 0.000 min
Response: 1328025
Conc: 718.73 ng/ml




#27 AR-1254-2

R.T.: 7.280 min
Delta R.T.: 0.000 min
Response: 2951011
Conc: 725.84 ng/ml



#27 AR-1254-2


R.T.: 6.511 min
Delta R.T.: 0.000 min
Response: 1180483
Conc: 716.88 ng/ml



#28 AR-1254-3


R.T.: 7.660 min
Delta R.T.: 0.000 min
Response: 3104622
Conc: 728.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#28 AR-1254-3

R.T.: 6.931 min
Delta R.T.: 0.000 min
Response: 1722604
Conc: 721.08 ng/ml




#29 AR-1254-4

R.T.: 7.956 min
Delta R.T.: 0.000 min
Response: 1916706
Conc: 724.96 ng/ml



#29 AR-1254-4


R.T.: 7.169 min
Delta R.T.: 0.000 min
Response: 762572
Conc: 701.18 ng/ml



#30 AR-1254-5

R.T.: 8.381 min
Delta R.T.: 0.000 min
Response: 2501043
Conc: 726.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750



#30 AR-1254-5

R.T.: 7.591 min
Delta R.T.: 0.000 min
Response: 1530949
Conc: 725.25 ng/ml