

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042622\
 Data File : PP046708.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Apr 2022 12:51
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 13:03:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 22 19:00:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.860	3.125	58795739	84793405	24.357	25.849
2) SA Decachlor...	9.941	8.443	36893670	67074896	25.072	28.120
Target Compounds						
3) L1 AR-1016-1	5.109	4.276	477113	122222	6.336	1.064 #
4) L1 AR-1016-2	5.134	4.282	154099	34045	1.353	0.201 #
5) L1 AR-1016-3	5.209	4.457	402014	229584	5.624	2.622 #
6) L1 AR-1016-4	5.308	4.515	110316	59425	1.835	0.854 #
7) L1 AR-1016-5	5.622	4.748	821640	4645636	15.518	54.621 #
8) L2 AR-1221-1	4.087	3.355	103942	42358	3.769	0.959 #
9) L2 AR-1221-2	4.184	3.436	652459	1130139	32.136	33.461
10) L2 AR-1221-3	4.260	3.525	70421	130928	1.080	1.318
11) L3 AR-1232-1	4.260	3.525	70421	130928	1.438	1.726
12) L3 AR-1232-2	4.824	4.282	177755	34045	7.323	0.472 #
13) L3 AR-1232-3	5.134	4.457	154099	229584	3.263	6.306 #
14) L3 AR-1232-4	5.308	4.555	110316	42278	4.534	1.327 #
15) L3 AR-1232-5	5.394	4.748	342490	4645636	20.584	161.272 #
16) L4 AR-1242-1	5.109	4.276	477113	122222	7.525	1.240 #
17) L4 AR-1242-2	5.134	4.282	154099	34045	1.617	0.235 #
18) L4 AR-1242-3	5.209	4.457	402014	229584	6.715	3.089 #
19) L4 AR-1242-4	5.308	4.561	110316	37080	2.181	0.519 #
20) L4 AR-1242-5	6.103	5.126	888417	3869765	17.631	40.105 #
21) L5 AR-1248-1	5.109	4.276	477113	122222	11.416	1.779 #
22) L5 AR-1248-2	5.394	4.515	342490	59425	5.795	0.636 #
23) L5 AR-1248-3	5.622	4.555	821640	42278	11.763	0.433 #
24) L5 AR-1248-4	6.061	4.748	467411	4645636	6.095	39.464 #
25) L5 AR-1248-5	6.103	5.150	888417	151800	11.422	1.246 #
27) L6 AR-1254-2	6.253f	5.268	1112211	94290	9.024	0.566 #
28) L6 AR-1254-3	6.672	5.699	1192621	125191	9.013	0.502 #
29) L6 AR-1254-4	6.987	5.947	364784	58523	3.670	0.341 #
30) L6 AR-1254-5	7.440	6.383f	84527	970987	0.961	4.450 #
31) L7 AR-1260-1	6.855	5.841	195292	48539	2.314	0.305 #
32) L7 AR-1260-2	7.145	6.053	125949	261384	1.251	1.285
33) L7 AR-1260-3	7.534	6.191f	42829	743750	0.529	4.354 #
34) L7 AR-1260-4	7.771	6.699f	22675	574030	0.285	3.949 #
35) L7 AR-1260-5	8.117	6.988	78611	143050	0.495	0.407
36) L8 AR-1262-1	7.534	6.433	42829	647002	0.354	2.619 #
37) L8 AR-1262-2	8.117	6.699f	78611	574030	0.409	2.970 #
38) L8 AR-1262-3	8.440	7.309f	67698	168837	0.504	1.040 #
39) L8 AR-1262-4	8.535	7.364	72158	355583	1.146	1.242
40) L8 AR-1262-5	9.208	7.927f	55437	623321	0.803	4.751 #
41) L9 AR-1268-1	8.440	7.309f	67698	168837	0.282	0.352 #
42) L9 AR-1268-2	8.535	7.364	72158	355583	0.326	0.840 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042622\
 Data File : PP046708.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Apr 2022 12:51
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 13:03:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 22 19:00:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43) L9	AR-1268-3	8.770	7.566	73039	466479	0.380	1.319 #
44) L9	AR-1268-4	9.208	7.927f	55437	623321	0.719	4.239 #
45) L9	AR-1268-5	9.616	8.191	529166	381030	0.945	0.373 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042622\
Data File : PP046708.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Apr 2022 12:51
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 27 13:03:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 22 19:00:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.860 min
Delta R.T.: -0.007 min
Response: 58795739
Conc: 24.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.125 min
Delta R.T.: -0.004 min
Response: 84793405
Conc: 25.85 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.941 min
Delta R.T.: -0.014 min
Response: 36893670
Conc: 25.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.443 min
Delta R.T.: -0.012 min
Response: 67074896
Conc: 28.12 ng/ml



#3 AR-1016-1


R.T.: 5.109 min
Delta R.T.: -0.003 min
Response: 477113
Conc: 6.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.276 min
Delta R.T.: 0.012 min
Response: 122222
Conc: 1.06 ng/ml




#4 AR-1016-2

R.T.: 5.134 min
Delta R.T.: -0.003 min
Response: 154099
Conc: 1.35 ng/ml



#4 AR-1016-2


R.T.: 4.282 min
Delta R.T.: 0.000 min
Response: 34045
Conc: 0.20 ng/ml



#5 AR-1016-3


R.T.: 5.209 min
Delta R.T.: 0.008 min
Response: 402014
Conc: 5.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.457 min
Delta R.T.: -0.009 min
Response: 229584
Conc: 2.62 ng/ml




#6 AR-1016-4

R.T.: 5.308 min
Delta R.T.: 0.003 min
Response: 110316
Conc: 1.83 ng/ml



#6 AR-1016-4


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 59425
Conc: 0.85 ng/ml



#7 AR-1016-5


R.T.: 5.622 min
Delta R.T.: -0.002 min
Response: 821640
Conc: 15.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.748 min
Delta R.T.: 0.010 min
Response: 4645636
Conc: 54.62 ng/ml




#8 AR-1221-1

R.T.: 4.087 min
Delta R.T.: -0.004 min
Response: 103942
Conc: 3.77 ng/ml



#8 AR-1221-1


R.T.: 3.355 min
Delta R.T.: 0.000 min
Response: 42358
Conc: 0.96 ng/ml



#9 AR-1221-2


R.T.: 4.184 min
Delta R.T.: 0.000 min
Response: 652459
Conc: 32.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.436 min
Delta R.T.: -0.008 min
Response: 1130139
Conc: 33.46 ng/ml




#10 AR-1221-3

R.T.: 4.260 min
Delta R.T.: -0.002 min
Response: 70421
Conc: 1.08 ng/ml



#10 AR-1221-3


R.T.: 3.525 min
Delta R.T.: 0.004 min
Response: 130928
Conc: 1.32 ng/ml



#11 AR-1232-1


R.T.: 4.260 min
Delta R.T.: -0.003 min
Response: 70421
Conc: 1.44 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.525 min
Delta R.T.: 0.004 min
Response: 130928
Conc: 1.73 ng/ml




#12 AR-1232-2

R.T.: 4.824 min
Delta R.T.: 0.000 min
Response: 177755
Conc: 7.32 ng/ml



#12 AR-1232-2


R.T.: 4.282 min
Delta R.T.: 0.000 min
Response: 34045
Conc: 0.47 ng/ml



#13 AR-1232-3


R.T.: 5.134 min
Delta R.T.: -0.001 min
Response: 154099
Conc: 3.26 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.457 min
Delta R.T.: -0.009 min
Response: 229584
Conc: 6.31 ng/ml




#14 AR-1232-4

R.T.: 5.308 min
Delta R.T.: 0.001 min
Response: 110316
Conc: 4.53 ng/ml



#14 AR-1232-4


R.T.: 4.555 min
Delta R.T.: -0.002 min
Response: 42278
Conc: 1.33 ng/ml



#15 AR-1232-5


R.T.: 5.394 min
Delta R.T.: -0.015 min
Response: 342490
Conc: 20.58 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.748 min
Delta R.T.: 0.010 min
Response: 4645636
Conc: 161.27 ng/ml




#16 AR-1242-1

R.T.: 5.109 min
Delta R.T.: -0.004 min
Response: 477113
Conc: 7.52 ng/ml



#16 AR-1242-1


R.T.: 4.276 min
Delta R.T.: 0.011 min
Response: 122222
Conc: 1.24 ng/ml



#17 AR-1242-2


R.T.: 5.134 min
Delta R.T.: -0.003 min
Response: 154099
Conc: 1.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.282 min
Delta R.T.: 0.000 min
Response: 34045
Conc: 0.24 ng/ml




#18 AR-1242-3

R.T.: 5.209 min
Delta R.T.: 0.007 min
Response: 402014
Conc: 6.72 ng/ml



#18 AR-1242-3


R.T.: 4.457 min
Delta R.T.: -0.010 min
Response: 229584
Conc: 3.09 ng/ml



#19 AR-1242-4


R.T.: 5.308 min
Delta R.T.: 0.000 min
Response: 110316
Conc: 2.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.561 min
Delta R.T.: 0.002 min
Response: 37080
Conc: 0.52 ng/ml




#20 AR-1242-5

R.T.: 6.103 min
Delta R.T.: -0.003 min
Response: 888417
Conc: 17.63 ng/ml



#20 AR-1242-5


R.T.: 5.126 min
Delta R.T.: 0.012 min
Response: 3869765
Conc: 40.11 ng/ml



#21 AR-1248-1


R.T.: 5.109 min
Delta R.T.: -0.002 min
Response: 477113
Conc: 11.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.276 min
Delta R.T.: 0.012 min
Response: 122222
Conc: 1.78 ng/ml




#22 AR-1248-2

R.T.: 5.394 min
Delta R.T.: -0.015 min
Response: 342490
Conc: 5.79 ng/ml



#22 AR-1248-2


R.T.: 4.515 min
Delta R.T.: 0.001 min
Response: 59425
Conc: 0.64 ng/ml



#23 AR-1248-3


R.T.: 5.622 min
Delta R.T.: -0.002 min
Response: 821640
Conc: 11.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.555 min
Delta R.T.: -0.002 min
Response: 42278
Conc: 0.43 ng/ml




#24 AR-1248-4

R.T.: 6.061 min
Delta R.T.: -0.004 min
Response: 467411
Conc: 6.10 ng/ml



#24 AR-1248-4


R.T.: 4.748 min
Delta R.T.: 0.012 min
Response: 4645636
Conc: 39.46 ng/ml



#25 AR-1248-5


R.T.: 6.103 min
Delta R.T.: -0.002 min
Response: 888417
Conc: 11.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.150 min
Delta R.T.: -0.005 min
Response: 151800
Conc: 1.25 ng/ml




#26 AR-1254-1

R.T.: 6.050 min
Delta R.T.: 0.014 min
Response: -990309
Conc: N.D.



#26 AR-1254-1


R.T.: 5.098 min
Delta R.T.: -0.013 min
Response: 20162
Conc: 0.10 ng/ml



#27 AR-1254-2


R.T.: 6.253 min
Delta R.T.: -0.019 min
Response: 1112211
Conc: 9.02 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.268 min
Delta R.T.: -0.004 min
Response: 94290
Conc: 0.57 ng/ml




#28 AR-1254-3

R.T.: 6.672 min
Delta R.T.: 0.000 min
Response: 1192621
Conc: 9.01 ng/ml



#28 AR-1254-3


R.T.: 5.699 min
Delta R.T.: -0.001 min
Response: 125191
Conc: 0.50 ng/ml



#29 AR-1254-4


R.T.: 6.987 min
Delta R.T.: 0.001 min
Response: 364784
Conc: 3.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 5.947 min
Delta R.T.: -0.003 min
Response: 58523
Conc: 0.34 ng/ml




#30 AR-1254-5

R.T.: 7.440 min
Delta R.T.: -0.005 min
Response: 84527
Conc: 0.96 ng/ml



#30 AR-1254-5


R.T.: 6.383 min
Delta R.T.: -0.015 min
Response: 970987
Conc: 4.45 ng/ml



#31 AR-1260-1


R.T.: 6.855 min
Delta R.T.: 0.004 min
Response: 195292
Conc: 2.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.841 min
Delta R.T.: 0.002 min
Response: 48539
Conc: 0.31 ng/ml




#32 AR-1260-2

R.T.: 7.145 min
Delta R.T.: 0.010 min
Response: 125949
Conc: 1.25 ng/ml



#32 AR-1260-2


R.T.: 6.053 min
Delta R.T.: 0.006 min
Response: 261384
Conc: 1.28 ng/ml



#33 AR-1260-3


R.T.: 7.534 min
Delta R.T.: 0.006 min
Response: 42829
Conc: 0.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.191 min
Delta R.T.: -0.016 min
Response: 743750
Conc: 4.35 ng/ml




#34 AR-1260-4

R.T.: 7.771 min
Delta R.T.: -0.006 min
Response: 22675
Conc: 0.28 ng/ml



#34 AR-1260-4


R.T.: 6.699 min
Delta R.T.: -0.019 min
Response: 574030
Conc: 3.95 ng/ml



#35 AR-1260-5


R.T.: 8.117 min
Delta R.T.: 0.002 min
Response: 78611
Conc: 0.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 6.988 min
Delta R.T.: 0.002 min
Response: 143050
Conc: 0.41 ng/ml




#36 AR-1262-1

R.T.: 7.534 min
Delta R.T.: 0.005 min
Response: 42829
Conc: 0.35 ng/ml



#36 AR-1262-1


R.T.: 6.433 min
Delta R.T.: -0.006 min
Response: 647002
Conc: 2.62 ng/ml



#37 AR-1262-2


R.T.: 8.117 min
Delta R.T.: 0.000 min
Response: 78611
Conc: 0.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.699 min
Delta R.T.: -0.018 min
Response: 574030
Conc: 2.97 ng/ml




#38 AR-1262-3

R.T.: 8.440 min
Delta R.T.: -0.001 min
Response: 67698
Conc: 0.50 ng/ml



#38 AR-1262-3


R.T.: 7.309 min
Delta R.T.: 0.020 min
Response: 168837
Conc: 1.04 ng/ml



#39 AR-1262-4


R.T.: 8.535 min
Delta R.T.: -0.001 min
Response: 72158
Conc: 1.15 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.364 min
Delta R.T.: 0.004 min
Response: 355583
Conc: 1.24 ng/ml




#40 AR-1262-5

R.T.: 9.208 min
Delta R.T.: 0.001 min
Response: 55437
Conc: 0.80 ng/ml



#40 AR-1262-5


R.T.: 7.927 min
Delta R.T.: 0.025 min
Response: 623321
Conc: 4.75 ng/ml



#41 AR-1268-1


R.T.: 8.440 min
Delta R.T.: 0.001 min
Response: 67698
Conc: 0.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.309 min
Delta R.T.: 0.020 min
Response: 168837
Conc: 0.35 ng/ml




#42 AR-1268-2

R.T.: 8.535 min
Delta R.T.: -0.002 min
Response: 72158
Conc: 0.33 ng/ml



#42 AR-1268-2


R.T.: 7.364 min
Delta R.T.: 0.002 min
Response: 355583
Conc: 0.84 ng/ml



#43 AR-1268-3


R.T.: 8.770 min
Delta R.T.: 0.002 min
Response: 73039
Conc: 0.38 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.566 min
Delta R.T.: -0.012 min
Response: 466479
Conc: 1.32 ng/ml




#44 AR-1268-4

R.T.: 9.208 min
Delta R.T.: 0.002 min
Response: 55437
Conc: 0.72 ng/ml



#44 AR-1268-4


R.T.: 7.927 min
Delta R.T.: 0.025 min
Response: 623321
Conc: 4.24 ng/ml



#45 AR-1268-5

R.T.: 9.616 min
Delta R.T.: -0.004 min
Response: 529166
Conc: 0.95 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.191 min
Delta R.T.: -0.010 min
Response: 381030
Conc: 0.37 ng/ml