

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042622\
 Data File : PP046664.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Apr 2022 18:19
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 01:45:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 22 19:00:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.862	3.125	125.7E6	179.3E6	52.057	54.673
2)	SA Decachlor...	9.942	8.445	74385485	133.0E6	50.551	55.777
Target Compounds							
26)	L6 AR-1254-1	6.028	5.104	35278443	96440717	607.352	499.787
27)	L6 AR-1254-2	6.267	5.264	57562838	85493665	467.064	513.404
28)	L6 AR-1254-3	6.666	5.693	56695899	128.7E6	428.479	516.192
29)	L6 AR-1254-4	6.979	5.942	41391727	90337721	416.453	526.523 #
30)	L6 AR-1254-5	7.438	6.390	42165580	118.1E6	479.412	541.069



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042622\
Data File : PP046664.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Apr 2022 18:19
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 27 01:45:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 22 19:00:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.862 min
Delta R.T.: -0.005 min
Response: 125662720
Conc: 52.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.125 min
Delta R.T.: -0.004 min
Response: 179349155
Conc: 54.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.942 min
Delta R.T.: -0.013 min
Response: 74385485
Conc: 50.55 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.445 min
Delta R.T.: -0.011 min
Response: 133045684
Conc: 55.78 ng/ml



#26 AR-1254-1


R.T.: 6.028 min
Delta R.T.: -0.008 min
Response: 35278443
Conc: 607.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.104 min
Delta R.T.: -0.007 min
Response: 96440717
Conc: 499.79 ng/ml




#27 AR-1254-2

R.T.: 6.267 min
Delta R.T.: -0.005 min
Response: 57562838
Conc: 467.06 ng/ml



#27 AR-1254-2


R.T.: 5.264 min
Delta R.T.: -0.007 min
Response: 85493665
Conc: 513.40 ng/ml



#28 AR-1254-3


R.T.: 6.666 min
Delta R.T.: -0.007 min
Response: 56695899
Conc: 428.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.693 min
Delta R.T.: -0.007 min
Response: 128711657
Conc: 516.19 ng/ml




#29 AR-1254-4

R.T.: 6.979 min
Delta R.T.: -0.007 min
Response: 41391727
Conc: 416.45 ng/ml



#29 AR-1254-4


R.T.: 5.942 min
Delta R.T.: -0.007 min
Response: 90337721
Conc: 526.52 ng/ml



#30 AR-1254-5

R.T.: 7.438 min
Delta R.T.: -0.007 min
Response: 42165580
Conc: 479.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.390 min
Delta R.T.: -0.008 min
Response: 118062749
Conc: 541.07 ng/ml