

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042623\
 Data File : PP057225.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Apr 2023 17:24
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 01:57:12 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 27 01:57:04 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.410	3.638	172.8E6	126.9E6	75.732	75.739
2) SA Decachlor...	10.244	8.704	95911052	125.1E6	75.140	74.891
Target Compounds						
16) L4 AR-1242-1	5.588	4.737	36298924	33023507	758.543	757.609
17) L4 AR-1242-2	5.611	4.756	52862133	46124488	757.422	753.903
18) L4 AR-1242-3	5.674	4.935	33277028	25720597	757.845	754.789
19) L4 AR-1242-4	5.772	5.019	26646537	28463741	754.908	752.069
20) L4 AR-1242-5	6.514	5.547	23310401	32326100	759.435	757.234

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042623\
Data File : PP057225.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Apr 2023 17:24
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 27 01:57:12 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 27 01:57:04 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

