

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042623\
 Data File : PP057241.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Apr 2023 21:46
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 03:58:49 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 27 03:58:43 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.411	3.638	169.6E6	123.4E6	75.312	75.318
2) SA Decachlor...	10.245	8.706	176.3E6	227.4E6	76.795	76.040
Target Compounds						
41) L9 AR-1268-1	8.696	7.583	137.8E6	187.0E6	768.951	758.647
42) L9 AR-1268-2	8.793	7.648	112.8E6	174.0E6	772.696	761.889
43) L9 AR-1268-3	9.027	7.856	122.1E6	159.2E6	764.222	752.293
44) L9 AR-1268-4	9.463	8.145	37059281	57140601	784.022	768.399
45) L9 AR-1268-5	9.893	8.443	316.1E6	424.1E6	781.552	765.166

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042623\
Data File : PP057241.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Apr 2023 21:46
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 27 03:58:49 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 27 03:58:43 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

