

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042623\
 Data File : PP057238.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Apr 2023 20:57
 Operator : YP\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 03:32:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 27 03:31:58 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.411	3.638	10148457	7727160	4.580	4.714
2) SA Decachlor...	10.244	8.705	6077543	9206291	4.769	5.258
Target Compounds						
26) L6 AR-1254-1	6.452	5.549	3911759	5173828	52.021	51.651
27) L6 AR-1254-2	6.673	5.697	5530142	4596616	52.516	52.229
28) L6 AR-1254-3	7.041	6.105	4950825	6841285	50.726	49.916
29) L6 AR-1254-4	7.328	6.334	3022268	3736692	51.514	49.578
30) L6 AR-1254-5	7.750	6.756	3043263	5800927	48.129	50.139



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042623\
Data File : PP057238.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Apr 2023 20:57
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 27 03:32:05 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 27 03:31:58 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.411 min
Delta R.T.: 0.000 min
Response: 10148457
Conc: 4.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254IC050




#1 Tetrachloro-m-xylene

R.T.: 3.638 min
Delta R.T.: 0.000 min
Response: 7727160
Conc: 4.71 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.244 min
Delta R.T.: 0.000 min
Response: 6077543
Conc: 4.77 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.705 min
Delta R.T.: 0.000 min
Response: 9206291
Conc: 5.26 ng/ml



#26 AR-1254-1


R.T.: 6.452 min
Delta R.T.: 0.000 min
Response: 3911759
Conc: 52.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050




#26 AR-1254-1

R.T.: 5.549 min
Delta R.T.: 0.000 min
Response: 5173828
Conc: 51.65 ng/ml




#27 AR-1254-2

R.T.: 6.673 min
Delta R.T.: 0.000 min
Response: 5530142
Conc: 52.52 ng/ml



#27 AR-1254-2


R.T.: 5.697 min
Delta R.T.: 0.000 min
Response: 4596616
Conc: 52.23 ng/ml



#28 AR-1254-3


R.T.: 7.041 min
Delta R.T.: 0.000 min
Response: 4950825
Conc: 50.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050




#28 AR-1254-3

R.T.: 6.105 min
Delta R.T.: 0.000 min
Response: 6841285
Conc: 49.92 ng/ml




#29 AR-1254-4

R.T.: 7.328 min
Delta R.T.: 0.000 min
Response: 3022268
Conc: 51.51 ng/ml



#29 AR-1254-4


R.T.: 6.334 min
Delta R.T.: 0.000 min
Response: 3736692
Conc: 49.58 ng/ml



#30 AR-1254-5

R.T.: 7.750 min
Delta R.T.: 0.000 min
Response: 3043263
Conc: 48.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254\ICC050



#30 AR-1254-5

R.T.: 6.756 min
Delta R.T.: 0.000 min
Response: 5800927
Conc: 50.14 ng/ml