

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042623\
 Data File : PP057246.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Apr 2023 23:08
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042623AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 02:04:24 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 27 02:03:31 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.412	3.639	118.4E6	87102898	52.536	52.208
2) SA Decachlor...	10.246	8.704	65179793	89362382	50.304	51.599
Target Compounds						
16) L4 AR-1242-1	5.590	4.738	25590244	23184889	520.367	515.100
17) L4 AR-1242-2	5.612	4.757	37095528	32346087	520.374	521.129
18) L4 AR-1242-3	5.675	4.936	23576582	18151810	526.109	525.163
19) L4 AR-1242-4	5.774	5.020	19013995	20379227	528.313	519.195
20) L4 AR-1242-5	6.516	5.548	16685764	22688260	527.691	519.572



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042623\
Data File : PP057246.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Apr 2023 23:08
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042623AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 27 02:04:24 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 27 02:03:31 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.412 min
Delta R.T.: 0.000 min
Response: 118367289
Conc: 52.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042623AR1242




#1 Tetrachloro-m-xylene

R.T.: 3.639 min
Delta R.T.: 0.000 min
Response: 87102898
Conc: 52.21 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.246 min
Delta R.T.: 0.001 min
Response: 65179793
Conc: 50.30 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.704 min
Delta R.T.: 0.000 min
Response: 89362382
Conc: 51.60 ng/ml



#16 AR-1242-1


R.T.: 5.590 min
Delta R.T.: 0.002 min
Response: 25590244
Conc: 520.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042623AR1242




#16 AR-1242-1

R.T.: 4.738 min
Delta R.T.: 0.000 min
Response: 23184889
Conc: 515.10 ng/ml




#17 AR-1242-2

R.T.: 5.612 min
Delta R.T.: 0.001 min
Response: 37095528
Conc: 520.37 ng/ml



#17 AR-1242-2


R.T.: 4.757 min
Delta R.T.: 0.000 min
Response: 32346087
Conc: 521.13 ng/ml



#18 AR-1242-3


R.T.: 5.675 min
Delta R.T.: 0.001 min
Response: 23576582
Conc: 526.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042623AR1242




#18 AR-1242-3

R.T.: 4.936 min
Delta R.T.: 0.000 min
Response: 18151810
Conc: 525.16 ng/ml




#19 AR-1242-4

R.T.: 5.774 min
Delta R.T.: 0.000 min
Response: 19013995
Conc: 528.31 ng/ml



#19 AR-1242-4


R.T.: 5.020 min
Delta R.T.: 0.000 min
Response: 20379227
Conc: 519.19 ng/ml



#20 AR-1242-5

R.T.: 6.516 min
Delta R.T.: 0.000 min
Response: 16685764
Conc: 527.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042623AR1242



#20 AR-1242-5

R.T.: 5.548 min
Delta R.T.: 0.000 min
Response: 22688260
Conc: 519.57 ng/ml