

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042721\
 Data File : PP035335.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Apr 2021 17:05
 Operator : DD\AJ
 Sample : M2158-06 20X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 17:34:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 27 07:44:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.836	4.238	70432	50047	0.869	0.958
2)	SA Decachlor...	10.734	0.000	77952	0	1.091	N.D. #
Target Compounds							
30)	L6 AR-1254-5	8.409f	0.000	6942	0	2.055	N.D. #
34)	L7 AR-1260-4	8.665f	0.000	2968	0	0.889	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042721\
Data File : PP035335.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Apr 2021 17:05
Operator : DD\AJ
Sample : M2158-06 20X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 27 17:34:13 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042621.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 27 07:44:36 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.836 min
Delta R.T.: -0.003 min
Response: 70432
Conc: 0.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.238 min
Delta R.T.: -0.004 min
Response: 50047
Conc: 0.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.734 min
Delta R.T.: -0.007 min
Response: 77952
Conc: 1.09 ng/ml



#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T. : 9.524 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 8.409 min
Delta R.T.: 0.029 min
Response: 6942
Conc: 2.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 7.592 min
Response: 0
Conc: N.D.



#34 AR-1260-4

R.T.: 8.665 min
Delta R.T.: -0.020 min
Response: 2968
Conc: 0.89 ng/ml



#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 7.892 min
Response: 0
Conc: N.D.