

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042822\
 Data File : PP046820.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2022 12:02
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 12:27:30 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 22 19:00:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.868	3.125	128.4E6	186.9E6	53.200	56.987
2) SA Decachlor...	9.958	8.444	97811844	142.2E6	66.470	59.609
Target Compounds						
26) L6 AR-1254-1	6.040	5.104	32245291	114.7E6	555.133	594.620
27) L6 AR-1254-2	6.276	5.264	66523333	89648768	539.770	538.356
28) L6 AR-1254-3	6.676	5.693	66552973	135.0E6	502.973	541.257
29) L6 AR-1254-4	6.989	5.942	54847604	99070886	551.836	577.423
30) L6 AR-1254-5	7.448	6.391	49733792	130.3E6	565.460	597.190



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042822\
Data File : PP046820.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2022 12:02
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 29 12:27:30 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 22 19:00:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.868 min
Delta R.T.: 0.001 min
Response: 128421038
Conc: 53.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.125 min
Delta R.T.: -0.005 min
Response: 186937172
Conc: 56.99 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.958 min
Delta R.T.: 0.009 min
Response: 97811844
Conc: 66.47 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.444 min
Delta R.T.: -0.005 min
Response: 142185820
Conc: 59.61 ng/ml



#26 AR-1254-1


R.T.: 6.040 min
Delta R.T.: 0.003 min
Response: 32245291
Conc: 555.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.104 min
Delta R.T.: -0.007 min
Response: 114740004
Conc: 594.62 ng/ml




#27 AR-1254-2

R.T.: 6.276 min
Delta R.T.: 0.004 min
Response: 66523333
Conc: 539.77 ng/ml



#27 AR-1254-2


R.T.: 5.264 min
Delta R.T.: -0.008 min
Response: 89648768
Conc: 538.36 ng/ml



#28 AR-1254-3


R.T.: 6.676 min
Delta R.T.: 0.003 min
Response: 66552973
Conc: 502.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.693 min
Delta R.T.: -0.008 min
Response: 134961634
Conc: 541.26 ng/ml




#29 AR-1254-4

R.T.: 6.989 min
Delta R.T.: 0.003 min
Response: 54847604
Conc: 551.84 ng/ml



#29 AR-1254-4


R.T.: 5.942 min
Delta R.T.: -0.008 min
Response: 99070886
Conc: 577.42 ng/ml



#30 AR-1254-5

R.T.: 7.448 min
Delta R.T.: 0.003 min
Response: 49733792
Conc: 565.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.391 min
Delta R.T.: -0.008 min
Response: 130308591
Conc: 597.19 ng/ml