

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
 Data File : PP071553.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Apr 2025 15:52
 Operator : YP\AJ
 Sample : Q1893-05DL 50X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 INTERIOR-3DL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 28 16:12:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.488f	3.808	1008532	1109303	0.509	0.787 #
2)	SA Decachlor...	10.248	8.841	864428	377606	0.597	0.431 #
Target Compounds							
3)	L1 AR-1016-1	5.666	4.895	17364851	16303332	258.984	304.544
4)	L1 AR-1016-2	5.688	4.913	29798596	21446022	290.355	280.840
5)	L1 AR-1016-3	5.750	5.090	18959959	13154835	307.556	311.649
6)	L1 AR-1016-4	5.847	5.132	19945552	15863558	388.816	460.001
7)	L1 AR-1016-5	6.140	5.347	33734762	27253932	699.334	613.174
8)	L2 AR-1221-1	4.731	4.010	2479316	514411	102.804	23.388 #
9)	L2 AR-1221-2	4.786	4.107	6683284	72020	370.944	4.349 #
10)	L2 AR-1221-3	4.875	4.174	7461197	4937190	131.797	101.846
11)	L3 AR-1232-1	4.875	4.174	7461197	4937190	166.129	128.848
12)	L3 AR-1232-2	5.401	4.913	5751844	21446022	249.938	561.600 #
13)	L3 AR-1232-3	5.688	5.090	29798596	13154835	631.091	621.235
14)	L3 AR-1232-4	5.847	5.174	19945552	20734236	860.488	1118.327 #
15)	L3 AR-1232-5	5.937	5.347	20472139	27253932	1267.961	1369.417
16)	L4 AR-1242-1	5.666	4.895	17364851	16303332	317.043	352.553
17)	L4 AR-1242-2	5.688	4.913	29798596	21446022	350.652	324.828
18)	L4 AR-1242-3	5.750	5.090	18959959	13154835	370.867	360.918
19)	L4 AR-1242-4	5.847	5.174	19945552	20734236	472.548	580.619
20)	L4 AR-1242-5	6.578	5.698	72404305	67148594	1564.129	1543.360
21)	L5 AR-1248-1	5.666	4.895	17364851	16303332	404.408	451.533
22)	L5 AR-1248-2	5.937	5.132	20472139	15863558	338.760	323.995
23)	L5 AR-1248-3	6.140	5.174	33734762	20734236	505.129	405.873
24)	L5 AR-1248-4	6.539	5.347	74573509	27253932	881.370	449.310 #
25)	L5 AR-1248-5	6.578	5.739	72404305	60620262	911.339	1039.961
26)	L6 AR-1254-1	6.513	5.698	46373423	67148594	565.866	734.664 #
27)	L6 AR-1254-2	6.733	5.846	88168554	24983883	693.824	317.408 #
28)	L6 AR-1254-3	7.095	6.250	34223199	29901899	267.840	252.319
29)	L6 AR-1254-4	7.377	6.478	33810914	25432703	287.720	329.586
30)	L6 AR-1254-5	7.793	6.896	13857783	9538519	128.634	94.005 #
31)	L7 AR-1260-1	7.259	6.394	6884846	11600424	71.699	160.450 #
32)	L7 AR-1260-2	7.512	6.568	11919267	6466266	83.306	73.933
33)	L7 AR-1260-3	7.870	6.721	3545634	5880963	31.628	75.024 #
34)	L7 AR-1260-4	8.090	7.192	23338695	3068353	206.784	48.619 #
35)	L7 AR-1260-5	8.414	7.434	21039294	10396517	92.859	68.953 #
36)	L8 AR-1262-1	8.090	6.932	23338695	3827662	161.523	33.691 #
37)	L8 AR-1262-2	8.414	7.192	21039294	3068353	76.734	33.229 #
38)	L8 AR-1262-3	8.738	7.716	12925839	2679766	70.680	35.402 #
39)	L8 AR-1262-4	8.817	7.779	8252242	6802500	61.487	54.496
40)	L8 AR-1262-5	9.471	8.280	3706916	2722409	42.047	49.058
41)	L9 AR-1268-1	8.738	7.716	12925839	2679766	41.605	13.420 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
Data File : PP071553.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Apr 2025 15:52
Operator : YP\AJ
Sample : Q1893-05DL 50X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 28 16:12:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.817	7.779	8252242	6802500	32.017	40.245 #
43)	L9 AR-1268-3	9.062	7.987	114533	144772	0.514	1.041 #
44)	L9 AR-1268-4	9.471	8.280	3706916	2722409	39.411	44.779
45)	L9 AR-1268-5	9.887	8.581	1932869	925236	3.231	2.479



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
Data File : PP071553.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Apr 2025 15:52
Operator : YP\AJ
Sample : Q1893-05DL 50X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 28 16:12:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.488 min
Delta R.T.: -0.029 min
Response: 1008532
Conc: 0.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#1 Tetrachloro-m-xylene

R.T.: 3.808 min
Delta R.T.: -0.004 min
Response: 1109303
Conc: 0.79 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.248 min
Delta R.T.: 0.010 min
Response: 864428
Conc: 0.60 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.841 min
Delta R.T.: -0.008 min
Response: 377606
Conc: 0.43 ng/ml



#3 AR-1016-1


R.T.: 5.666 min
Delta R.T.: -0.005 min
Response: 17364851
Conc: 258.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#3 AR-1016-1

R.T.: 4.895 min
Delta R.T.: -0.003 min
Response: 16303332
Conc: 304.54 ng/ml




#4 AR-1016-2

R.T.: 5.688 min
Delta R.T.: -0.005 min
Response: 29798596
Conc: 290.35 ng/ml



#4 AR-1016-2


R.T.: 4.913 min
Delta R.T.: -0.003 min
Response: 21446022
Conc: 280.84 ng/ml



#5 AR-1016-3


R.T.: 5.750 min
Delta R.T.: -0.005 min
Response: 18959959
Conc: 307.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#5 AR-1016-3

R.T.: 5.090 min
Delta R.T.: -0.003 min
Response: 13154835
Conc: 311.65 ng/ml




#6 AR-1016-4

R.T.: 5.847 min
Delta R.T.: -0.005 min
Response: 19945552
Conc: 388.82 ng/ml



#6 AR-1016-4


R.T.: 5.132 min
Delta R.T.: -0.004 min
Response: 15863558
Conc: 460.00 ng/ml



#7 AR-1016-5


R.T.: 6.140 min
Delta R.T.: -0.005 min
Response: 33734762
Conc: 699.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#7 AR-1016-5

R.T.: 5.347 min
Delta R.T.: -0.003 min
Response: 27253932
Conc: 613.17 ng/ml




#8 AR-1221-1

R.T.: 4.731 min
Delta R.T.: 0.016 min
Response: 2479316
Conc: 102.80 ng/ml



#8 AR-1221-1


R.T.: 4.010 min
Delta R.T.: -0.013 min
Response: 514411
Conc: 23.39 ng/ml



#9 AR-1221-2


R.T.: 4.786 min
Delta R.T.: -0.014 min
Response: 6683284
Conc: 370.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#9 AR-1221-2

R.T.: 4.107 min
Delta R.T.: -0.002 min
Response: 72020
Conc: 4.35 ng/ml




#10 AR-1221-3

R.T.: 4.875 min
Delta R.T.: 0.000 min
Response: 7461197
Conc: 131.80 ng/ml



#10 AR-1221-3


R.T.: 4.174 min
Delta R.T.: -0.011 min
Response: 4937190
Conc: 101.85 ng/ml



#11 AR-1232-1


R.T.: 4.875 min
Delta R.T.: -0.003 min
Response: 7461197
Conc: 166.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#11 AR-1232-1

R.T.: 4.174 min
Delta R.T.: -0.013 min
Response: 4937190
Conc: 128.85 ng/ml




#12 AR-1232-2

R.T.: 5.401 min
Delta R.T.: -0.003 min
Response: 5751844
Conc: 249.94 ng/ml



#12 AR-1232-2


R.T.: 4.913 min
Delta R.T.: -0.005 min
Response: 21446022
Conc: 561.60 ng/ml



#13 AR-1232-3


R.T.: 5.688 min
Delta R.T.: -0.002 min
Response: 29798596
Conc: 631.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#13 AR-1232-3

R.T.: 5.090 min
Delta R.T.: -0.004 min
Response: 13154835
Conc: 621.24 ng/ml




#14 AR-1232-4

R.T.: 5.847 min
Delta R.T.: -0.003 min
Response: 19945552
Conc: 860.49 ng/ml



#14 AR-1232-4


R.T.: 5.174 min
Delta R.T.: -0.004 min
Response: 20734236
Conc: 1118.33 ng/ml



#15 AR-1232-5


R.T.: 5.937 min
Delta R.T.: -0.003 min
Response: 20472139
Conc: 1267.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#15 AR-1232-5

R.T.: 5.347 min
Delta R.T.: -0.004 min
Response: 27253932
Conc: 1369.42 ng/ml




#16 AR-1242-1

R.T.: 5.666 min
Delta R.T.: -0.003 min
Response: 17364851
Conc: 317.04 ng/ml



#16 AR-1242-1


R.T.: 4.895 min
Delta R.T.: -0.003 min
Response: 16303332
Conc: 352.55 ng/ml



#17 AR-1242-2


R.T.: 5.688 min
Delta R.T.: -0.002 min
Response: 29798596
Conc: 350.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#17 AR-1242-2

R.T.: 4.913 min
Delta R.T.: -0.003 min
Response: 21446022
Conc: 324.83 ng/ml




#18 AR-1242-3

R.T.: 5.750 min
Delta R.T.: -0.002 min
Response: 18959959
Conc: 370.87 ng/ml



#18 AR-1242-3


R.T.: 5.090 min
Delta R.T.: -0.003 min
Response: 13154835
Conc: 360.92 ng/ml



#19 AR-1242-4


R.T.: 5.847 min
Delta R.T.: -0.002 min
Response: 19945552
Conc: 472.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#19 AR-1242-4

R.T.: 5.174 min
Delta R.T.: -0.003 min
Response: 20734236
Conc: 580.62 ng/ml




#20 AR-1242-5

R.T.: 6.578 min
Delta R.T.: -0.003 min
Response: 72404305
Conc: 1564.13 ng/ml



#20 AR-1242-5


R.T.: 5.698 min
Delta R.T.: -0.003 min
Response: 67148594
Conc: 1543.36 ng/ml



#21 AR-1248-1


R.T.: 5.666 min
Delta R.T.: -0.002 min
Response: 17364851
Conc: 404.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#21 AR-1248-1

R.T.: 4.895 min
Delta R.T.: -0.003 min
Response: 16303332
Conc: 451.53 ng/ml




#22 AR-1248-2

R.T.: 5.937 min
Delta R.T.: -0.003 min
Response: 20472139
Conc: 338.76 ng/ml



#22 AR-1248-2


R.T.: 5.132 min
Delta R.T.: -0.004 min
Response: 15863558
Conc: 323.99 ng/ml



#23 AR-1248-3


R.T.: 6.140 min
Delta R.T.: -0.002 min
Response: 33734762
Conc: 505.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#23 AR-1248-3

R.T.: 5.174 min
Delta R.T.: -0.003 min
Response: 20734236
Conc: 405.87 ng/ml




#24 AR-1248-4

R.T.: 6.539 min
Delta R.T.: -0.003 min
Response: 74573509
Conc: 881.37 ng/ml



#24 AR-1248-4


R.T.: 5.347 min
Delta R.T.: -0.003 min
Response: 27253932
Conc: 449.31 ng/ml



#25 AR-1248-5


R.T.: 6.578 min
Delta R.T.: -0.003 min
Response: 72404305
Conc: 911.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#25 AR-1248-5

R.T.: 5.739 min
Delta R.T.: -0.004 min
Response: 60620262
Conc: 1039.96 ng/ml




#26 AR-1254-1

R.T.: 6.513 min
Delta R.T.: -0.006 min
Response: 46373423
Conc: 565.87 ng/ml



#26 AR-1254-1


R.T.: 5.698 min
Delta R.T.: -0.004 min
Response: 67148594
Conc: 734.66 ng/ml



#27 AR-1254-2


R.T.: 6.733 min
Delta R.T.: -0.002 min
Response: 88168554
Conc: 693.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#27 AR-1254-2

R.T.: 5.846 min
Delta R.T.: -0.005 min
Response: 24983883
Conc: 317.41 ng/ml




#28 AR-1254-3

R.T.: 7.095 min
Delta R.T.: -0.004 min
Response: 34223199
Conc: 267.84 ng/ml



#28 AR-1254-3


R.T.: 6.250 min
Delta R.T.: -0.004 min
Response: 29901899
Conc: 252.32 ng/ml



#29 AR-1254-4


R.T.: 7.377 min
Delta R.T.: -0.004 min
Response: 33810914
Conc: 287.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#29 AR-1254-4

R.T.: 6.478 min
Delta R.T.: -0.005 min
Response: 25432703
Conc: 329.59 ng/ml




#30 AR-1254-5

R.T.: 7.793 min
Delta R.T.: -0.004 min
Response: 13857783
Conc: 128.63 ng/ml



#30 AR-1254-5


R.T.: 6.896 min
Delta R.T.: -0.005 min
Response: 9538519
Conc: 94.01 ng/ml



#31 AR-1260-1


R.T.: 7.259 min
Delta R.T.: -0.005 min
Response: 6884846
Conc: 71.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#31 AR-1260-1

R.T.: 6.394 min
Delta R.T.: 0.009 min
Response: 11600424
Conc: 160.45 ng/ml




#32 AR-1260-2

R.T.: 7.512 min
Delta R.T.: -0.007 min
Response: 11919267
Conc: 83.31 ng/ml



#32 AR-1260-2


R.T.: 6.568 min
Delta R.T.: -0.005 min
Response: 6466266
Conc: 73.93 ng/ml



#33 AR-1260-3


R.T.: 7.870 min
Delta R.T.: -0.007 min
Response: 3545634
Conc: 31.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#33 AR-1260-3

R.T.: 6.721 min
Delta R.T.: -0.006 min
Response: 5880963
Conc: 75.02 ng/ml




#34 AR-1260-4

R.T.: 8.090 min
Delta R.T.: -0.012 min
Response: 23338695
Conc: 206.78 ng/ml



#34 AR-1260-4


R.T.: 7.192 min
Delta R.T.: -0.006 min
Response: 3068353
Conc: 48.62 ng/ml



#35 AR-1260-5


R.T.: 8.414 min
Delta R.T.: -0.008 min
Response: 21039294
Conc: 92.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#35 AR-1260-5

R.T.: 7.434 min
Delta R.T.: -0.005 min
Response: 10396517
Conc: 68.95 ng/ml




#36 AR-1262-1

R.T.: 8.090 min
Delta R.T.: -0.011 min
Response: 23338695
Conc: 161.52 ng/ml



#36 AR-1262-1


R.T.: 6.932 min
Delta R.T.: -0.007 min
Response: 3827662
Conc: 33.69 ng/ml



#37 AR-1262-2


R.T.: 8.414 min
Delta R.T.: -0.006 min
Response: 21039294
Conc: 76.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#37 AR-1262-2

R.T.: 7.192 min
Delta R.T.: -0.006 min
Response: 3068353
Conc: 33.23 ng/ml




#38 AR-1262-3

R.T.: 8.738 min
Delta R.T.: 0.000 min
Response: 12925839
Conc: 70.68 ng/ml



#38 AR-1262-3


R.T.: 7.716 min
Delta R.T.: -0.005 min
Response: 2679766
Conc: 35.40 ng/ml



#39 AR-1262-4


R.T.: 8.817 min
Delta R.T.: -0.005 min
Response: 8252242
Conc: 61.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#39 AR-1262-4

R.T.: 7.779 min
Delta R.T.: -0.006 min
Response: 6802500
Conc: 54.50 ng/ml




#40 AR-1262-5

R.T.: 9.471 min
Delta R.T.: -0.007 min
Response: 3706916
Conc: 42.05 ng/ml



#40 AR-1262-5


R.T.: 8.280 min
Delta R.T.: -0.006 min
Response: 2722409
Conc: 49.06 ng/ml



#41 AR-1268-1


R.T.: 8.738 min
Delta R.T.: 0.005 min
Response: 12925839
Conc: 41.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#41 AR-1268-1

R.T.: 7.716 min
Delta R.T.: -0.006 min
Response: 2679766
Conc: 13.42 ng/ml




#42 AR-1268-2

R.T.: 8.817 min
Delta R.T.: -0.009 min
Response: 8252242
Conc: 32.02 ng/ml



#42 AR-1268-2


R.T.: 7.779 min
Delta R.T.: -0.009 min
Response: 6802500
Conc: 40.25 ng/ml



#43 AR-1268-3


R.T.: 9.062 min
Delta R.T.: 0.004 min
Response: 114533
Conc: 0.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL




#43 AR-1268-3

R.T.: 7.987 min
Delta R.T.: -0.006 min
Response: 144772
Conc: 1.04 ng/ml




#44 AR-1268-4

R.T.: 9.471 min
Delta R.T.: -0.007 min
Response: 3706916
Conc: 39.41 ng/ml



#44 AR-1268-4


R.T.: 8.280 min
Delta R.T.: -0.007 min
Response: 2722409
Conc: 44.78 ng/ml



#45 AR-1268-5

R.T.: 9.887 min
Delta R.T.: -0.006 min
Response: 1932869
Conc: 3.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
INTERIOR-3DL



#45 AR-1268-5

R.T.: 8.581 min
Delta R.T.: -0.008 min
Response: 925236
Conc: 2.48 ng/ml