

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
 Data File : PP071578.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2025 01:25
 Operator : YP\AJ
 Sample : Q1891-01
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 MH-C

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/29/2025
 Supervised By :mohammad ahmed 05/02/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 06:29:14 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.511	3.808	35442665	25637383	17.877m	18.200
2) SA Decachlor...	10.226	8.839	26782212	16657898	18.496	18.993

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
Data File : PP071578.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2025 01:25
Operator : YP\AJ
Sample : Q1891-01
Misc :
ALS Vial : 37 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
MH-C


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/29/2025
Supervised By :mohammad ahmed 05/02/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 29 06:29:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.511 min
Delta R.T.: -0.006 min
Response: 35442665
Conc: 17.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
MH-C


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/29/2025
Supervised By :mohammad ahmed 05/02/2025




#1 Tetrachloro-m-xylene

R.T.: 3.808 min
Delta R.T.: -0.004 min
Response: 25637383
Conc: 18.20 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.226 min
Delta R.T.: -0.012 min
Response: 26782212
Conc: 18.50 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.839 min
Delta R.T.: -0.010 min
Response: 16657898
Conc: 18.99 ng/ml