

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042919\
 Data File : PP030060.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Apr 2019 8:13
 Operator : EI/SJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 05:43:13 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042519.M
 Quant Title : 8080.M
 QLast Update : Fri Apr 26 10:04:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds

8) T	DICHLORPROP	13.679	0.000	1664906	0	3.506	N.D.	#
9) T	2,4-D	14.052	0.000	1815722	0	3.805	N.D.	#
15) T	Picloram	17.808	18.396	2922126	2131163	0.824	0.791	



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042919\
Data File : PP030060.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Apr 2019 8:13
Operator : EI/SJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 05:43:13 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042519.M
Quant Title : 8080.M
QLast Update : Fri Apr 26 10:04:13 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 0.000 min
Exp R.T. : 12.149 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
HEXANE




#4 2,4-DCAA

R.T.: 0.000 min
Exp R.T. : 12.219 min
Response: 0
Conc: N.D.




#8 DICHLORPROP

R.T.: 13.679 min
Delta R.T.: 0.002 min
Response: 1664906
Conc: 3.51 ng/ml



#8 DICHLORPROP


R.T.: 0.000 min
Exp R.T. : 13.649 min
Response: 0
Conc: N.D.



#9 2,4-D


R.T.: 14.052 min
Delta R.T.: 0.009 min
Response: 1815722
Conc: 3.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEXANE




#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 14.142 min
Response: 0
Conc: N.D.



#15 Picloram

R.T.: 17.808 min
Delta R.T.: 0.025 min
Response: 2922126
Conc: 0.82 ng/ml



#15 Picloram

R.T.: 18.396 min
Delta R.T.: 0.037 min
Response: 2131163
Conc: 0.79 ng/ml