

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP042921.M
 Title : GC EXTRACTABLES
 Last Update : Fri Apr 30 06:28:57 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP035414.D 1000 =PP035412.D 750 =PP035413.D
 250 =PP035415.D 50 =PP035416.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	4.847	4.682	4.757	4.874	5.084	4.849	E4	3.13
2) SA	Decachlorobiphenyl	4.322	4.052	4.167	4.404	4.460	4.281	E4	3.95
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1	2.806	2.620	2.697	2.888	3.250	2.852	E3	8.59
37) L8	AR-1262-2	4.912	4.727	4.815	4.971	4.814	4.848	E3	1.96
38) L8	AR-1262-3	3.400	3.227	3.308	3.435	3.407	3.355	E3	2.58
39) L8	AR-1262-4	2.627	2.492	2.560	2.662	2.688	2.606	E3	3.05
40) L8	AR-1262-5	1.732	1.659	1.696	1.736	1.693	1.703	E3	1.84
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =PP035414.D 1000 =PP035412.D 750 =PP035413.D
 250 =PP035415.D 50 =PP035416.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.403	3.229	3.293	3.496	3.747	3.434	E4	5.91
2) SA	Decachlorobiphenyl	2.100	1.920	1.981	2.229	2.353	2.116	E4	8.38
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP042921.M
 Title : GC EXTRACTABLES
 Last Update : Fri Apr 30 06:28:57 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP035414.D 1000 =PP035412.D 750 =PP035413.D
 250 =PP035415.D 50 =PP035416.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	0.954	0.871	0.907	0.984	1.053	0.954 E3	7.37
37) L8	AR-1262-2	2.889	2.755	2.826	3.007	3.303	2.956 E3	7.27
38) L8	AR-1262-3	1.221	1.119	1.155	1.346	1.401	1.248 E3	9.73
39) L8	AR-1262-4	2.206	2.048	2.104	2.308	2.448	2.223 E3	7.20
40) L8	AR-1262-5	0.999	0.942	0.970	1.049	1.057	1.003 E3	4.95
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range