

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP042921.M
 Title : GC EXTRACTABLES
 Last Update : Fri Apr 30 06:33:55 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP035419.D 1000 =PP035417.D 750 =PP035418.D
 250 =PP035420.D 50 =PP035421.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	4.772	4.639	4.715	4.895	5.105	4.825 E4	3.77
2) SA	Decachlorobiphenyl	7.322	7.001	7.127	7.598	8.048	7.419 E4	5.63
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	5.967	5.744	5.855	6.146	6.373	6.017 E3	4.13
42) L9	AR-1268-2	5.611	5.416	5.530	5.739	6.050	5.669 E3	4.29
43) L9	AR-1268-3	4.878	4.688	4.781	5.034	5.204	4.917 E3	4.17
44) L9	AR-1268-4	1.901	1.839	1.879	1.915	1.952	1.897 E3	2.21
45) L9	AR-1268-5	1.512	1.479	1.498	1.538	1.640	1.534 E4	4.14

Signal #2 Calibration Files

500 =PP035419.D 1000 =PP035417.D 750 =PP035418.D
 250 =PP035420.D 50 =PP035421.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	3.377	3.191	3.287	3.523	3.665	3.409 E4	5.52
2) SA	Decachlorobiphenyl	3.500	3.269	3.361	3.733	4.268	3.626 E4	11.00
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP042921.M
 Title : GC EXTRACTABLES
 Last Update : Fri Apr 30 06:33:55 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP035419.D 1000 =PP035417.D 750 =PP035418.D
 250 =PP035420.D 50 =PP035421.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	3.354	3.149	3.234	3.578	4.016	3.466 E3	10.02
42) L9	AR-1268-2	3.068	2.895	2.965	3.255	3.586	3.154 E3	8.79
43) L9	AR-1268-3	2.622	2.452	2.521	2.795	3.127	2.703 E3	9.98
44) L9	AR-1268-4	1.121	1.055	1.077	1.171	1.235	1.132 E3	6.44
45) L9	AR-1268-5	7.390	7.220	7.320	7.755	8.846	7.706 E3	8.67

(#) = Out of Range