

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042921\
 Data File : PP035405.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Apr 2021 20:39
 Operator : DD\AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 06:54:41 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 30 06:49:37 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.790	4.208	1220767	876403	25.866	26.690
2)	SA Decachlor...	10.665	9.476	1064493	530573	26.109	27.068
Target Compounds							
21)	L5 AR-1248-1	6.097	5.453	263791	159332	270.344	266.301
22)	L5 AR-1248-2	6.389	5.711	380055	227005	293.579	273.034
23)	L5 AR-1248-3	6.604	5.756	426156	234642	273.453	268.889
24)	L5 AR-1248-4	7.031	5.940	444780	273687	270.057	266.955
25)	L5 AR-1248-5	7.071	6.360	426463	215878	266.889	226.295



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042921\
Data File : PP035405.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Apr 2021 20:39
Operator : DD\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 06:54:41 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042921.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 30 06:49:37 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.790 min
Delta R.T.: -0.001 min
Response: 1220767
Conc: 25.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#1 Tetrachloro-m-xylene

R.T.: 4.208 min
Delta R.T.: 0.000 min
Response: 876403
Conc: 26.69 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.665 min
Delta R.T.: -0.002 min
Response: 1064493
Conc: 26.11 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.476 min
Delta R.T.: 0.000 min
Response: 530573
Conc: 27.07 ng/ml



#21 AR-1248-1


R.T.: 6.097 min
Delta R.T.: 0.000 min
Response: 263791
Conc: 270.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#21 AR-1248-1

R.T.: 5.453 min
Delta R.T.: 0.000 min
Response: 159332
Conc: 266.30 ng/ml




#22 AR-1248-2

R.T.: 6.389 min
Delta R.T.: 0.000 min
Response: 380055
Conc: 293.58 ng/ml



#22 AR-1248-2


R.T.: 5.711 min
Delta R.T.: 0.000 min
Response: 227005
Conc: 273.03 ng/ml



#23 AR-1248-3


R.T.: 6.604 min
Delta R.T.: 0.000 min
Response: 426156
Conc: 273.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#23 AR-1248-3

R.T.: 5.756 min
Delta R.T.: 0.000 min
Response: 234642
Conc: 268.89 ng/ml




#24 AR-1248-4

R.T.: 7.031 min
Delta R.T.: 0.000 min
Response: 444780
Conc: 270.06 ng/ml



#24 AR-1248-4


R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 273687
Conc: 266.95 ng/ml



#25 AR-1248-5

R.T.: 7.071 min
Delta R.T.: 0.000 min
Response: 426463
Conc: 266.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250



#25 AR-1248-5

R.T.: 6.360 min
Delta R.T.: 0.000 min
Response: 215878
Conc: 226.29 ng/ml