

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042922\
 Data File : PP046864.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Apr 2022 04:45
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 00:39:53 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 22 19:00:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.859	3.124	57526756	83750263	23.831	25.531
2)	SA Decachlor...	9.937	8.440	35377376	60778894	24.042	25.480
Target Compounds							
3)	L1 AR-1016-1	5.110	4.259	193101	24608	2.564	0.214 #
4)	L1 AR-1016-2	5.137	4.279	110375	81696	0.969	0.482 #
5)	L1 AR-1016-3	5.199	4.463	243829	21639	3.411	0.247 #
6)	L1 AR-1016-4	5.311	4.521	494363	81284	8.223	1.168 #
7)	L1 AR-1016-5	5.608f	4.751	347885	4401843	6.570	51.754 #
8)	L2 AR-1221-1	4.084	3.361	102648	14154	3.722	0.321 #
9)	L2 AR-1221-2	4.182	3.434	792189	1131439	39.018	33.499
10)	L2 AR-1221-3	4.263	3.523	264228	20523	4.053	0.207 #
11)	L3 AR-1232-1	4.263	3.523	264228	20523	5.397	0.271 #
12)	L3 AR-1232-2	4.826	4.279	149524	81696	6.160	1.133 #
13)	L3 AR-1232-3	5.137	4.463	110375	21639	2.337	0.594 #
14)	L3 AR-1232-4	5.311	4.556	494363	80461	20.317	2.526 #
15)	L3 AR-1232-5	5.410	4.751	230009	4401843	13.824	152.809 #
16)	L4 AR-1242-1	5.110	4.259	193101	24608	3.046	0.250 #
17)	L4 AR-1242-2	5.137	4.279	110375	81696	1.159	0.564 #
18)	L4 AR-1242-3	5.199	4.463	243829	21639	4.073	0.291 #
19)	L4 AR-1242-4	5.311	4.556	494363	80461	9.775	1.127 #
20)	L4 AR-1242-5	6.101	5.124	8143	3203099	0.162	33.196 #
21)	L5 AR-1248-1	5.110	4.266	193101	19702	4.620	0.287 #
22)	L5 AR-1248-2	5.410	4.509	230009	33092	3.892	0.354 #
23)	L5 AR-1248-3	5.608f	4.556	347885	80461	4.980	0.824 #
24)	L5 AR-1248-4	6.073	4.751	48080	4401843	0.627	37.393 #
25)	L5 AR-1248-5	6.101	5.157	8143	108273	0.105	0.888 #
26)	L6 AR-1254-1	6.036	5.101	202401	15297	3.485	0.079 #
27)	L6 AR-1254-2	6.264	5.269	356769	72716	2.895	0.437 #
28)	L6 AR-1254-3	6.675	5.697	317462	56518	2.399	0.227 #
29)	L6 AR-1254-4	6.991	5.949	162599	39771	1.636	0.232 #
30)	L6 AR-1254-5	7.448	6.399	140509	95391	1.598	0.437 #
31)	L7 AR-1260-1	6.855	5.837	93895	28815	1.113	0.181 #
32)	L7 AR-1260-2	7.137	6.052	56331	51262	0.559	0.252 #
33)	L7 AR-1260-3	7.523	6.205	41093	250321	0.507	1.465 #
34)	L7 AR-1260-4	7.778	6.718	110990	109848	1.393	0.756 #
35)	L7 AR-1260-5	8.117	6.979	17535	63965	0.110	0.182 #
36)	L8 AR-1262-1	7.523	6.437	41093	188586	0.340	0.763 #
37)	L8 AR-1262-2	8.117	6.718	17535	109848	0.091	0.568 #
38)	L8 AR-1262-3	8.442	7.282	68532	211545	0.510	1.303 #
39)	L8 AR-1262-4	8.537	7.370	77371	11822	1.228	0.041 #
40)	L8 AR-1262-5	9.200	7.902	43628	87134	0.632	0.664
41)	L9 AR-1268-1	8.442	7.282	68532	211545	0.285	0.441 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042922\
 Data File : PP046864.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Apr 2022 04:45
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 00:39:53 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 22 19:00:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.537	7.370	77371	11822	0.349	0.028 #
43) L9	AR-1268-3	8.760	7.585	125592	73165	0.654	0.207 #
44) L9	AR-1268-4	9.200	7.902	43628	87134	0.565	0.593
45) L9	AR-1268-5	9.628	8.205	107382	33004	0.192	0.032 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042922\
Data File : PP046864.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Apr 2022 04:45
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 00:39:53 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 22 19:00:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.859 min
Delta R.T.: -0.008 min
Response: 57526756
Conc: 23.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.124 min
Delta R.T.: -0.005 min
Response: 83750263
Conc: 25.53 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.937 min
Delta R.T.: -0.012 min
Response: 35377376
Conc: 24.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.440 min
Delta R.T.: -0.010 min
Response: 60778894
Conc: 25.48 ng/ml



#3 AR-1016-1


R.T.: 5.110 min
Delta R.T.: -0.002 min
Response: 193101
Conc: 2.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.259 min
Delta R.T.: 0.000 min
Response: 24608
Conc: 0.21 ng/ml




#4 AR-1016-2

R.T.: 5.137 min
Delta R.T.: 0.000 min
Response: 110375
Conc: 0.97 ng/ml



#4 AR-1016-2


R.T.: 4.279 min
Delta R.T.: -0.003 min
Response: 81696
Conc: 0.48 ng/ml



#5 AR-1016-3


R.T.: 5.199 min
Delta R.T.: -0.003 min
Response: 243829
Conc: 3.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.463 min
Delta R.T.: -0.003 min
Response: 21639
Conc: 0.25 ng/ml




#6 AR-1016-4

R.T.: 5.311 min
Delta R.T.: 0.005 min
Response: 494363
Conc: 8.22 ng/ml



#6 AR-1016-4


R.T.: 4.521 min
Delta R.T.: 0.006 min
Response: 81284
Conc: 1.17 ng/ml



#7 AR-1016-5


R.T.: 5.608 min
Delta R.T.: -0.016 min
Response: 347885
Conc: 6.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.751 min
Delta R.T.: 0.012 min
Response: 4401843
Conc: 51.75 ng/ml




#8 AR-1221-1

R.T.: 4.084 min
Delta R.T.: -0.007 min
Response: 102648
Conc: 3.72 ng/ml



#8 AR-1221-1


R.T.: 3.361 min
Delta R.T.: 0.006 min
Response: 14154
Conc: 0.32 ng/ml



#9 AR-1221-2


R.T.: 4.182 min
Delta R.T.: 0.000 min
Response: 792189
Conc: 39.02 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.434 min
Delta R.T.: -0.010 min
Response: 1131439
Conc: 33.50 ng/ml




#10 AR-1221-3

R.T.: 4.263 min
Delta R.T.: 0.000 min
Response: 264228
Conc: 4.05 ng/ml



#10 AR-1221-3


R.T.: 3.523 min
Delta R.T.: 0.002 min
Response: 20523
Conc: 0.21 ng/ml



#11 AR-1232-1


R.T.: 4.263 min
Delta R.T.: 0.000 min
Response: 264228
Conc: 5.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.523 min
Delta R.T.: 0.002 min
Response: 20523
Conc: 0.27 ng/ml




#12 AR-1232-2

R.T.: 4.826 min
Delta R.T.: 0.003 min
Response: 149524
Conc: 6.16 ng/ml



#12 AR-1232-2


R.T.: 4.279 min
Delta R.T.: -0.003 min
Response: 81696
Conc: 1.13 ng/ml



#13 AR-1232-3


R.T.: 5.137 min
Delta R.T.: 0.001 min
Response: 110375
Conc: 2.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.463 min
Delta R.T.: -0.003 min
Response: 21639
Conc: 0.59 ng/ml




#14 AR-1232-4

R.T.: 5.311 min
Delta R.T.: 0.004 min
Response: 494363
Conc: 20.32 ng/ml



#14 AR-1232-4


R.T.: 4.556 min
Delta R.T.: -0.002 min
Response: 80461
Conc: 2.53 ng/ml



#15 AR-1232-5


R.T.: 5.410 min
Delta R.T.: 0.002 min
Response: 230009
Conc: 13.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.751 min
Delta R.T.: 0.012 min
Response: 4401843
Conc: 152.81 ng/ml




#16 AR-1242-1

R.T.: 5.110 min
Delta R.T.: -0.003 min
Response: 193101
Conc: 3.05 ng/ml



#16 AR-1242-1


R.T.: 4.259 min
Delta R.T.: 0.000 min
Response: 24608
Conc: 0.25 ng/ml



#17 AR-1242-2


R.T.: 5.137 min
Delta R.T.: 0.000 min
Response: 110375
Conc: 1.16 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.279 min
Delta R.T.: -0.003 min
Response: 81696
Conc: 0.56 ng/ml




#18 AR-1242-3

R.T.: 5.199 min
Delta R.T.: -0.004 min
Response: 243829
Conc: 4.07 ng/ml



#18 AR-1242-3


R.T.: 4.463 min
Delta R.T.: -0.004 min
Response: 21639
Conc: 0.29 ng/ml



#19 AR-1242-4


R.T.: 5.311 min
Delta R.T.: 0.002 min
Response: 494363
Conc: 9.77 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.556 min
Delta R.T.: -0.003 min
Response: 80461
Conc: 1.13 ng/ml




#20 AR-1242-5

R.T.: 6.101 min
Delta R.T.: -0.005 min
Response: 8143
Conc: 0.16 ng/ml



#20 AR-1242-5


R.T.: 5.124 min
Delta R.T.: 0.011 min
Response: 3203099
Conc: 33.20 ng/ml



#21 AR-1248-1


R.T.: 5.110 min
Delta R.T.: -0.002 min
Response: 193101
Conc: 4.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.266 min
Delta R.T.: 0.003 min
Response: 19702
Conc: 0.29 ng/ml




#22 AR-1248-2

R.T.: 5.410 min
Delta R.T.: 0.002 min
Response: 230009
Conc: 3.89 ng/ml



#22 AR-1248-2


R.T.: 4.509 min
Delta R.T.: -0.005 min
Response: 33092
Conc: 0.35 ng/ml



#23 AR-1248-3


R.T.: 5.608 min
Delta R.T.: -0.016 min
Response: 347885
Conc: 4.98 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.556 min
Delta R.T.: -0.002 min
Response: 80461
Conc: 0.82 ng/ml




#24 AR-1248-4

R.T.: 6.073 min
Delta R.T.: 0.008 min
Response: 48080
Conc: 0.63 ng/ml



#24 AR-1248-4


R.T.: 4.751 min
Delta R.T.: 0.014 min
Response: 4401843
Conc: 37.39 ng/ml



#25 AR-1248-5


R.T.: 6.101 min
Delta R.T.: -0.005 min
Response: 8143
Conc: 0.10 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.157 min
Delta R.T.: 0.002 min
Response: 108273
Conc: 0.89 ng/ml




#26 AR-1254-1

R.T.: 6.036 min
Delta R.T.: 0.000 min
Response: 202401
Conc: 3.48 ng/ml



#26 AR-1254-1


R.T.: 5.101 min
Delta R.T.: -0.010 min
Response: 15297
Conc: 0.08 ng/ml



#27 AR-1254-2


R.T.: 6.264 min
Delta R.T.: -0.008 min
Response: 356769
Conc: 2.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.269 min
Delta R.T.: -0.002 min
Response: 72716
Conc: 0.44 ng/ml




#28 AR-1254-3

R.T.: 6.675 min
Delta R.T.: 0.002 min
Response: 317462
Conc: 2.40 ng/ml



#28 AR-1254-3


R.T.: 5.697 min
Delta R.T.: -0.003 min
Response: 56518
Conc: 0.23 ng/ml



#29 AR-1254-4


R.T.: 6.991 min
Delta R.T.: 0.005 min
Response: 162599
Conc: 1.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 5.949 min
Delta R.T.: 0.000 min
Response: 39771
Conc: 0.23 ng/ml




#30 AR-1254-5

R.T.: 7.448 min
Delta R.T.: 0.003 min
Response: 140509
Conc: 1.60 ng/ml



#30 AR-1254-5


R.T.: 6.399 min
Delta R.T.: 0.000 min
Response: 95391
Conc: 0.44 ng/ml



#31 AR-1260-1


R.T.: 6.855 min
Delta R.T.: 0.003 min
Response: 93895
Conc: 1.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.837 min
Delta R.T.: -0.003 min
Response: 28815
Conc: 0.18 ng/ml




#32 AR-1260-2

R.T.: 7.137 min
Delta R.T.: 0.002 min
Response: 56331
Conc: 0.56 ng/ml



#32 AR-1260-2


R.T.: 6.052 min
Delta R.T.: 0.005 min
Response: 51262
Conc: 0.25 ng/ml



#33 AR-1260-3


R.T.: 7.523 min
Delta R.T.: -0.005 min
Response: 41093
Conc: 0.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.205 min
Delta R.T.: -0.002 min
Response: 250321
Conc: 1.47 ng/ml




#34 AR-1260-4

R.T.: 7.778 min
Delta R.T.: 0.000 min
Response: 110990
Conc: 1.39 ng/ml



#34 AR-1260-4


R.T.: 6.718 min
Delta R.T.: 0.000 min
Response: 109848
Conc: 0.76 ng/ml



#35 AR-1260-5


R.T.: 8.117 min
Delta R.T.: 0.003 min
Response: 17535
Conc: 0.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 6.979 min
Delta R.T.: -0.007 min
Response: 63965
Conc: 0.18 ng/ml




#36 AR-1262-1

R.T.: 7.523 min
Delta R.T.: -0.005 min
Response: 41093
Conc: 0.34 ng/ml



#36 AR-1262-1


R.T.: 6.437 min
Delta R.T.: -0.002 min
Response: 188586
Conc: 0.76 ng/ml



#37 AR-1262-2


R.T.: 8.117 min
Delta R.T.: 0.001 min
Response: 17535
Conc: 0.09 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.718 min
Delta R.T.: 0.000 min
Response: 109848
Conc: 0.57 ng/ml




#38 AR-1262-3

R.T.: 8.442 min
Delta R.T.: 0.000 min
Response: 68532
Conc: 0.51 ng/ml



#38 AR-1262-3


R.T.: 7.282 min
Delta R.T.: -0.007 min
Response: 211545
Conc: 1.30 ng/ml



#39 AR-1262-4


R.T.: 8.537 min
Delta R.T.: 0.000 min
Response: 77371
Conc: 1.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.370 min
Delta R.T.: 0.010 min
Response: 11822
Conc: 0.04 ng/ml




#40 AR-1262-5

R.T.: 9.200 min
Delta R.T.: -0.006 min
Response: 43628
Conc: 0.63 ng/ml



#40 AR-1262-5


R.T.: 7.902 min
Delta R.T.: 0.000 min
Response: 87134
Conc: 0.66 ng/ml



#41 AR-1268-1


R.T.: 8.442 min
Delta R.T.: 0.003 min
Response: 68532
Conc: 0.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.282 min
Delta R.T.: -0.007 min
Response: 211545
Conc: 0.44 ng/ml




#42 AR-1268-2

R.T.: 8.537 min
Delta R.T.: 0.000 min
Response: 77371
Conc: 0.35 ng/ml



#42 AR-1268-2


R.T.: 7.370 min
Delta R.T.: 0.008 min
Response: 11822
Conc: 0.03 ng/ml



#43 AR-1268-3


R.T.: 8.760 min
Delta R.T.: -0.007 min
Response: 125592
Conc: 0.65 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.585 min
Delta R.T.: 0.007 min
Response: 73165
Conc: 0.21 ng/ml




#44 AR-1268-4

R.T.: 9.200 min
Delta R.T.: -0.006 min
Response: 43628
Conc: 0.57 ng/ml



#44 AR-1268-4


R.T.: 7.902 min
Delta R.T.: 0.000 min
Response: 87134
Conc: 0.59 ng/ml



#45 AR-1268-5

R.T.: 9.628 min
Delta R.T.: 0.008 min
Response: 107382
Conc: 0.19 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.205 min
Delta R.T.: 0.003 min
Response: 33004
Conc: 0.03 ng/ml