

Method Path : P:\HPCHEM1\ECD P\methods\
 Method File : PP043018.M
 Title : 8080.M
 Last Update : Tue May 01 05:53:49 2018
 Response Via : Initial Calibration

Calibration Files

750 =PP025359.D 1000 =PP025360.D 1500 =PP025361.D
 500 =PP025358.D 200 =PP025357.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	8.016	7.649	7.398	8.013	8.904	7.996	E5	7.14
2) T	3,5-DICHLOROBENZOI...	8.577	8.224	7.853	8.693	9.363	8.542	E5	6.61
3) T	4-Nitrophenol	3.785	3.624	3.462	3.809	4.093	3.754	E5	6.26
4) S	2,4-DCAA	5.721	5.482	5.260	5.803	6.341	5.721	E5	7.10
5) T	DICAMBA	2.529	2.413	2.282	2.556	2.660	2.488	E6	5.82
6) T	MCPP	1.690	1.675	1.633	1.650	1.315	1.593	E6	9.85
7) T	MCPA	2.434	2.354	2.286	2.399	2.418	2.378	E6	2.51
8) T	DICHLORPROP	6.213	5.966	5.730	6.289	6.807	6.201	E5	6.51
9) T	2,4-D	7.845	7.520	7.211	7.967	8.585	7.825	E5	6.60
10) T	Pentachlorophenol	0.986	0.918	0.813	1.054	1.171	0.988	E7	13.68
11) T	2,4,5-TP (SILVEX)	4.219	3.971	3.694	4.327	4.596	4.161	E6	8.27
12) T	2,4,5-T	3.911	3.676	3.433	4.020	4.261	3.860	E6	8.24
13) T	2,4-DB	4.853	4.665	4.542	4.832	5.046	4.788	E5	4.02
14) T	DINOSEB	2.759	2.631	2.517	2.773	2.840	2.704	E6	4.77
15) T	Picloram	6.009	5.721	5.356	6.116	6.264	5.893	E6	6.11
16) T	DCPA	5.003	4.756	4.366	5.258	5.587	4.994	E6	9.35

Signal #2 Calibration Files

750 =PP025359.D 1000 =PP025360.D 1500 =PP025361.D
 500 =PP025358.D 200 =PP025357.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	8.070	7.703	7.389	8.175	8.939	8.055	E5	7.25
2) T	3,5-DICHLOROBENZOI...	8.076	7.677	7.273	8.202	8.978	8.041	E5	7.93
3) T	4-Nitrophenol	3.573	3.379	3.214	3.601	3.939	3.541	E5	7.68
4) S	2,4-DCAA	6.090	5.798	5.527	6.175	6.792	6.076	E5	7.81
5) T	DICAMBA	2.323	2.205	2.074	2.342	2.447	2.278	E6	6.27
6) T	MCPP	1.482	1.428	1.382	1.441	1.495	1.446	E6	3.12
7) T	MCPA	2.164	2.084	2.010	2.134	2.184	2.116	E6	3.30
8) T	DICHLORPROP	5.660	5.388	5.118	5.732	6.161	5.612	E5	6.97
9) T	2,4-D	7.586	7.229	6.879	7.699	8.289	7.537	E5	7.03
10) T	Pentachlorophenol	0.873	0.801	0.714	0.923	1.024	0.867	E7	13.62
11) T	2,4,5-TP (SILVEX)	3.665	3.444	3.187	3.753	4.003	3.610	E6	8.58
12) T	2,4,5-T	3.033	2.866	2.667	3.084	3.248	2.979	E6	7.44
13) T	2,4-DB	4.276	4.088	3.919	4.286	4.573	4.229	E5	5.79
14) T	DINOSEB	2.307	2.170	2.069	2.292	2.360	2.240	E6	5.27
15) T	Picloram	4.934	4.624	4.310	4.976	5.159	4.801	E6	6.98
16) T	DCPA	4.291	3.991	3.670	4.423	4.785	4.232	E6	10.02

(#) = Out of Range