

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050123\
 Data File : PP057318.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2023 13:35
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 01 14:23:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 14:22:23 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.415	3.642	60512999	46235533	25.154	25.373
2) SA Decachlor...	10.264	8.716	41374379	51204682	25.940	26.661
Target Compounds						
16) L4 AR-1242-1	5.594	4.741	14419035	13307454	264.203	268.863
17) L4 AR-1242-2	5.617	4.761	20727550	18227448	263.096	259.700
18) L4 AR-1242-3	5.680	4.939	13277687	10307389	267.997	265.045
19) L4 AR-1242-4	5.779	5.024	10679430	11607194	268.660	270.026
20) L4 AR-1242-5	6.522	5.552	9290257	12747369	266.345	264.145



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050123\
Data File : PP057318.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2023 13:35
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 01 14:23:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050123.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 14:22:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.415 min
Delta R.T.: 0.000 min
Response: 60512999
Conc: 25.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.642 min
Delta R.T.: 0.000 min
Response: 46235533
Conc: 25.37 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.264 min
Delta R.T.: 0.002 min
Response: 41374379
Conc: 25.94 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.716 min
Delta R.T.: 0.000 min
Response: 51204682
Conc: 26.66 ng/ml



#16 AR-1242-1


R.T.: 5.594 min
Delta R.T.: 0.000 min
Response: 14419035
Conc: 264.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 4.741 min
Delta R.T.: 0.000 min
Response: 13307454
Conc: 268.86 ng/ml




#17 AR-1242-2

R.T.: 5.617 min
Delta R.T.: 0.000 min
Response: 20727550
Conc: 263.10 ng/ml



#17 AR-1242-2


R.T.: 4.761 min
Delta R.T.: 0.000 min
Response: 18227448
Conc: 259.70 ng/ml



#18 AR-1242-3


R.T.: 5.680 min
Delta R.T.: 0.000 min
Response: 13277687
Conc: 268.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 4.939 min
Delta R.T.: 0.000 min
Response: 10307389
Conc: 265.05 ng/ml




#19 AR-1242-4

R.T.: 5.779 min
Delta R.T.: 0.000 min
Response: 10679430
Conc: 268.66 ng/ml



#19 AR-1242-4


R.T.: 5.024 min
Delta R.T.: 0.000 min
Response: 11607194
Conc: 270.03 ng/ml



#20 AR-1242-5

R.T.: 6.522 min
Delta R.T.: 0.000 min
Response: 9290257
Conc: 266.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.552 min
Delta R.T.: 0.000 min
Response: 12747369
Conc: 264.14 ng/ml