

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050123\
 Data File : PP057337.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2023 18:51
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP050123AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 01 19:07:31 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 18:39:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.414	3.641	117.0E6	88048988	49.687	49.661
2) SA Decachlor...	10.260	8.713	74035640	92136360	44.877	46.950
Target Compounds						
16) L4 AR-1242-1	5.593	4.741	25656474	23847213	474.920	483.989
17) L4 AR-1242-2	5.615	4.760	37994333	33062901	482.865	483.129
18) L4 AR-1242-3	5.678	4.939	23918798	18654211	486.759	489.229
19) L4 AR-1242-4	5.777	5.023	19078101	20825310	486.147	489.137
20) L4 AR-1242-5	6.521	5.551	16737405	23148910	487.637	486.791



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050123\
Data File : PP057337.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2023 18:51
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP050123AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 01 19:07:31 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050123.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 18:39:24 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.414 min
Delta R.T.: -0.001 min
Response: 117015227
Conc: 49.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP050123AR1242




#1 Tetrachloro-m-xylene

R.T.: 3.641 min
Delta R.T.: 0.000 min
Response: 88048988
Conc: 49.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.260 min
Delta R.T.: -0.004 min
Response: 74035640
Conc: 44.88 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.713 min
Delta R.T.: 0.000 min
Response: 92136360
Conc: 46.95 ng/ml



#16 AR-1242-1


R.T.: 5.593 min
Delta R.T.: 0.000 min
Response: 25656474
Conc: 474.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP050123AR1242




#16 AR-1242-1

R.T.: 4.741 min
Delta R.T.: 0.000 min
Response: 23847213
Conc: 483.99 ng/ml




#17 AR-1242-2

R.T.: 5.615 min
Delta R.T.: -0.001 min
Response: 37994333
Conc: 482.87 ng/ml



#17 AR-1242-2


R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 33062901
Conc: 483.13 ng/ml



#18 AR-1242-3


R.T.: 5.678 min
Delta R.T.: -0.001 min
Response: 23918798
Conc: 486.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP050123AR1242




#18 AR-1242-3

R.T.: 4.939 min
Delta R.T.: 0.000 min
Response: 18654211
Conc: 489.23 ng/ml




#19 AR-1242-4

R.T.: 5.777 min
Delta R.T.: 0.000 min
Response: 19078101
Conc: 486.15 ng/ml



#19 AR-1242-4


R.T.: 5.023 min
Delta R.T.: 0.000 min
Response: 20825310
Conc: 489.14 ng/ml



#20 AR-1242-5

R.T.: 6.521 min
Delta R.T.: -0.001 min
Response: 16737405
Conc: 487.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP050123AR1242



#20 AR-1242-5

R.T.: 5.551 min
Delta R.T.: -0.001 min
Response: 23148910
Conc: 486.79 ng/ml