

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050123\
 Data File : PP057349.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2023 22:10
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 02:07:12 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 18:39:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.414	3.641	118.2E6	89835293	50.193	50.668
2) SA Decachlor...	10.258	8.713	77787266	94549023	47.151	48.180
Target Compounds						
3) L1 AR-1016-1	5.593	4.740	31538597	29038207	473.364	480.668
4) L1 AR-1016-2	5.615	4.759	46002335	40785808	473.415	484.636
5) L1 AR-1016-3	5.677	4.938	29074347	22863990	479.256	490.846
6) L1 AR-1016-4	5.777	4.980	23262359	20491757	480.393	491.322
7) L1 AR-1016-5	6.075	5.196	25717719	26288116	477.679	491.487
31) L7 AR-1260-1	7.211	6.240	47526257	54618253	485.291	478.581
32) L7 AR-1260-2	7.469	6.429	47003249	60965780	463.707	469.776
33) L7 AR-1260-3	7.833	6.584	38766021	59808671	478.295	479.766
34) L7 AR-1260-4	8.060	7.059	43959783	50220948	466.694	480.629
35) L7 AR-1260-5	8.386	7.302	70775390	98866663	450.028	466.197



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050123\
Data File : PP057349.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2023 22:10
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 02:07:12 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050123.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 18:39:24 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.414 min
Delta R.T.: -0.001 min
Response: 118207359
Conc: 50.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.641 min
Delta R.T.: 0.000 min
Response: 89835293
Conc: 50.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.258 min
Delta R.T.: -0.006 min
Response: 77787266
Conc: 47.15 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.713 min
Delta R.T.: -0.001 min
Response: 94549023
Conc: 48.18 ng/ml



#3 AR-1016-1


R.T.: 5.593 min
Delta R.T.: -0.002 min
Response: 31538597
Conc: 473.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.740 min
Delta R.T.: -0.001 min
Response: 29038207
Conc: 480.67 ng/ml




#4 AR-1016-2

R.T.: 5.615 min
Delta R.T.: -0.002 min
Response: 46002335
Conc: 473.41 ng/ml



#4 AR-1016-2


R.T.: 4.759 min
Delta R.T.: 0.000 min
Response: 40785808
Conc: 484.64 ng/ml



#5 AR-1016-3


R.T.: 5.677 min
Delta R.T.: -0.003 min
Response: 29074347
Conc: 479.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.938 min
Delta R.T.: 0.000 min
Response: 22863990
Conc: 490.85 ng/ml




#6 AR-1016-4

R.T.: 5.777 min
Delta R.T.: -0.002 min
Response: 23262359
Conc: 480.39 ng/ml



#6 AR-1016-4


R.T.: 4.980 min
Delta R.T.: 0.000 min
Response: 20491757
Conc: 491.32 ng/ml



#7 AR-1016-5


R.T.: 6.075 min
Delta R.T.: -0.002 min
Response: 25717719
Conc: 477.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.196 min
Delta R.T.: -0.001 min
Response: 26288116
Conc: 491.49 ng/ml




#31 AR-1260-1

R.T.: 7.211 min
Delta R.T.: -0.002 min
Response: 47526257
Conc: 485.29 ng/ml



#31 AR-1260-1


R.T.: 6.240 min
Delta R.T.: 0.000 min
Response: 54618253
Conc: 478.58 ng/ml



#32 AR-1260-2


R.T.: 7.469 min
Delta R.T.: -0.003 min
Response: 47003249
Conc: 463.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.429 min
Delta R.T.: 0.000 min
Response: 60965780
Conc: 469.78 ng/ml




#33 AR-1260-3

R.T.: 7.833 min
Delta R.T.: -0.003 min
Response: 38766021
Conc: 478.29 ng/ml



#33 AR-1260-3


R.T.: 6.584 min
Delta R.T.: -0.001 min
Response: 59808671
Conc: 479.77 ng/ml



#34 AR-1260-4


R.T.: 8.060 min
Delta R.T.: -0.003 min
Response: 43959783
Conc: 466.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.059 min
Delta R.T.: -0.002 min
Response: 50220948
Conc: 480.63 ng/ml



#35 AR-1260-5

R.T.: 8.386 min
Delta R.T.: -0.004 min
Response: 70775390
Conc: 450.03 ng/ml



#35 AR-1260-5

R.T.: 7.302 min
Delta R.T.: 0.000 min
Response: 98866663
Conc: 466.20 ng/ml