

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050223\
 Data File : PP057414.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2023 20:04
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 03 04:40:57 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 18:39:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.413	3.640	123.7E6	92481217	52.522	52.161
2)	SA Decachlor...	10.261	8.714	84013025	97808127	50.925	49.840
Target Compounds							
3)	L1 AR-1016-1	5.593	4.741	34445082	31419941	516.987	520.093
4)	L1 AR-1016-2	5.616	4.759	50825487	45260448	523.050	537.805
5)	L1 AR-1016-3	5.679	4.939	31498624	24690865	519.217	530.065
6)	L1 AR-1016-4	5.778	4.980	25357976	21682476	523.670	519.871
7)	L1 AR-1016-5	6.076	5.197	27123711	27903732	503.794	521.693
8)	L2 AR-1221-1	4.619	3.855	4381345	3272178	151.883	149.616
9)	L2 AR-1221-2	4.712	3.943	6183536	4879431	291.613	301.432
10)	L2 AR-1221-3	4.788	4.020	22001310	18410809	352.076	374.310
11)	L3 AR-1232-1	4.788	4.020	22001310	18410809	420.468	444.534
12)	L3 AR-1232-2	5.323	4.759	29342681	45260448	1100.559	1164.289
13)	L3 AR-1232-3	5.616	4.939	50825487	24690865	1145.945	1226.795
14)	L3 AR-1232-4	5.778	5.023	25357976	26538488	1147.241	1543.238 #
15)	L3 AR-1232-5	5.870	5.197	24344915	27903732	1161.310	1218.819
16)	L4 AR-1242-1	5.593	4.741	34445082	31419941	637.604	637.681
17)	L4 AR-1242-2	5.616	4.759	50825487	45260448	645.935	661.365
18)	L4 AR-1242-3	5.679	4.939	31498624	24690865	641.012	647.547
19)	L4 AR-1242-4	5.778	5.023	25357976	26538488	646.170	623.325
20)	L4 AR-1242-5	6.522	5.552	6332124	27456607	184.484	577.376 #
21)	L5 AR-1248-1	5.593	4.741	34445082	31419941	837.258	839.814
22)	L5 AR-1248-2	5.870	4.980	24344915	21682476	354.406	361.205
23)	L5 AR-1248-3	6.076	5.023	27123711	26538488	375.636	425.769
24)	L5 AR-1248-4	6.482	5.197	7429545	27903732	120.636	388.073 #
25)	L5 AR-1248-5	6.522	5.593	6332124	6336420	106.886	102.413
26)	L6 AR-1254-1	6.456	5.552	22793392	27456607	282.965	254.602
27)	L6 AR-1254-2	6.677	5.701	19561253	20301053	168.759	211.751 #
28)	L6 AR-1254-3	7.046	6.125f	8919953	38042999	81.454	252.992 #
29)	L6 AR-1254-4	7.335	6.339	5523942	2448421	87.293	30.986 #
30)	L6 AR-1254-5	7.755	6.760	54715451	67560624	729.282	531.914 #
31)	L7 AR-1260-1	7.213	6.240	49298552	57217643	503.388	501.358
32)	L7 AR-1260-2	7.472	6.430	51999989	65965868	513.002	508.305
33)	L7 AR-1260-3	7.834	6.585	40785323	63114277	503.209	506.283
34)	L7 AR-1260-4	8.061	7.061	46731441	52765068	496.119	504.977
35)	L7 AR-1260-5	8.388	7.303	80700811	108.4E6	513.139	511.380
36)	L8 AR-1262-1	7.834	6.853	40785323	29776905	389.496	491.772 #
37)	L8 AR-1262-2	8.388	7.061	80700811	52765068	540.263	424.652
38)	L8 AR-1262-3	8.719	7.589	47166126	27285793	452.078	286.389 #
39)	L8 AR-1262-4	8.800	7.653	13543707	80203620	181.716	500.286 #
40)	L8 AR-1262-5	9.477	8.153	20328144	27110809	431.594	391.096
41)	L9 AR-1268-1	8.719	7.589	47166126	27285793	224.928	96.247 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050223\
 Data File : PP057414.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2023 20:04
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 03 04:40:57 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 18:39:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.800	7.653	13543707	80203620	84.171	318.705 #
43) L9	AR-1268-3	9.039	7.863	1146645	1636908	6.250	7.120
44) L9	AR-1268-4	9.477	8.153	20328144	27110809	388.166	342.103
45) L9	AR-1268-5	9.909	8.451	6810896	8574843	13.639	14.139



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050223\
Data File : PP057414.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2023 20:04
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 03 04:40:57 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050123.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 18:39:24 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.413 min
Delta R.T.: -0.002 min
Response: 123690441
Conc: 52.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.640 min
Delta R.T.: 0.000 min
Response: 92481217
Conc: 52.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.261 min
Delta R.T.: -0.002 min
Response: 84013025
Conc: 50.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.714 min
Delta R.T.: 0.000 min
Response: 97808127
Conc: 49.84 ng/ml



#3 AR-1016-1


R.T.: 5.593 min
Delta R.T.: -0.001 min
Response: 34445082
Conc: 516.99 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.741 min
Delta R.T.: 0.000 min
Response: 31419941
Conc: 520.09 ng/ml




#4 AR-1016-2

R.T.: 5.616 min
Delta R.T.: 0.000 min
Response: 50825487
Conc: 523.05 ng/ml



#4 AR-1016-2


R.T.: 4.759 min
Delta R.T.: 0.000 min
Response: 45260448
Conc: 537.81 ng/ml



#5 AR-1016-3


R.T.: 5.679 min
Delta R.T.: -0.001 min
Response: 31498624
Conc: 519.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.939 min
Delta R.T.: 0.000 min
Response: 24690865
Conc: 530.07 ng/ml




#6 AR-1016-4

R.T.: 5.778 min
Delta R.T.: 0.000 min
Response: 25357976
Conc: 523.67 ng/ml



#6 AR-1016-4


R.T.: 4.980 min
Delta R.T.: 0.000 min
Response: 21682476
Conc: 519.87 ng/ml



#7 AR-1016-5


R.T.: 6.076 min
Delta R.T.: 0.000 min
Response: 27123711
Conc: 503.79 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.197 min
Delta R.T.: 0.000 min
Response: 27903732
Conc: 521.69 ng/ml




#8 AR-1221-1

R.T.: 4.619 min
Delta R.T.: -0.005 min
Response: 4381345
Conc: 151.88 ng/ml



#8 AR-1221-1


R.T.: 3.855 min
Delta R.T.: -0.003 min
Response: 3272178
Conc: 149.62 ng/ml



#9 AR-1221-2


R.T.: 4.712 min
Delta R.T.: 0.000 min
Response: 6183536
Conc: 291.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.943 min
Delta R.T.: -0.001 min
Response: 4879431
Conc: 301.43 ng/ml




#10 AR-1221-3

R.T.: 4.788 min
Delta R.T.: 0.000 min
Response: 22001310
Conc: 352.08 ng/ml



#10 AR-1221-3


R.T.: 4.020 min
Delta R.T.: -0.002 min
Response: 18410809
Conc: 374.31 ng/ml



#11 AR-1232-1


R.T.: 4.788 min
Delta R.T.: -0.001 min
Response: 22001310
Conc: 420.47 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.020 min
Delta R.T.: -0.001 min
Response: 18410809
Conc: 444.53 ng/ml




#12 AR-1232-2

R.T.: 5.323 min
Delta R.T.: 0.000 min
Response: 29342681
Conc: 1100.56 ng/ml



#12 AR-1232-2


R.T.: 4.759 min
Delta R.T.: 0.000 min
Response: 45260448
Conc: 1164.29 ng/ml



#13 AR-1232-3


R.T.: 5.616 min
Delta R.T.: 0.000 min
Response: 50825487
Conc: 1145.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.939 min
Delta R.T.: 0.000 min
Response: 24690865
Conc: 1226.79 ng/ml




#14 AR-1232-4

R.T.: 5.778 min
Delta R.T.: 0.000 min
Response: 25357976
Conc: 1147.24 ng/ml



#14 AR-1232-4


R.T.: 5.023 min
Delta R.T.: 0.000 min
Response: 26538488
Conc: 1543.24 ng/ml



#15 AR-1232-5


R.T.: 5.870 min
Delta R.T.: 0.000 min
Response: 24344915
Conc: 1161.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.197 min
Delta R.T.: 0.000 min
Response: 27903732
Conc: 1218.82 ng/ml




#16 AR-1242-1

R.T.: 5.593 min
Delta R.T.: 0.000 min
Response: 34445082
Conc: 637.60 ng/ml



#16 AR-1242-1


R.T.: 4.741 min
Delta R.T.: 0.000 min
Response: 31419941
Conc: 637.68 ng/ml



#17 AR-1242-2


R.T.: 5.616 min
Delta R.T.: 0.000 min
Response: 50825487
Conc: 645.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.759 min
Delta R.T.: -0.001 min
Response: 45260448
Conc: 661.36 ng/ml




#18 AR-1242-3

R.T.: 5.679 min
Delta R.T.: 0.000 min
Response: 31498624
Conc: 641.01 ng/ml



#18 AR-1242-3


R.T.: 4.939 min
Delta R.T.: 0.000 min
Response: 24690865
Conc: 647.55 ng/ml



#19 AR-1242-4


R.T.: 5.778 min
Delta R.T.: 0.000 min
Response: 25357976
Conc: 646.17 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.023 min
Delta R.T.: 0.000 min
Response: 26538488
Conc: 623.33 ng/ml




#20 AR-1242-5

R.T.: 6.522 min
Delta R.T.: 0.000 min
Response: 6332124
Conc: 184.48 ng/ml



#20 AR-1242-5


R.T.: 5.552 min
Delta R.T.: 0.000 min
Response: 27456607
Conc: 577.38 ng/ml



#21 AR-1248-1


R.T.: 5.593 min
Delta R.T.: -0.002 min
Response: 34445082
Conc: 837.26 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.741 min
Delta R.T.: -0.001 min
Response: 31419941
Conc: 839.81 ng/ml




#22 AR-1248-2

R.T.: 5.870 min
Delta R.T.: -0.001 min
Response: 24344915
Conc: 354.41 ng/ml



#22 AR-1248-2


R.T.: 4.980 min
Delta R.T.: -0.001 min
Response: 21682476
Conc: 361.20 ng/ml



#23 AR-1248-3


R.T.: 6.076 min
Delta R.T.: -0.001 min
Response: 27123711
Conc: 375.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.023 min
Delta R.T.: -0.001 min
Response: 26538488
Conc: 425.77 ng/ml




#24 AR-1248-4

R.T.: 6.482 min
Delta R.T.: -0.002 min
Response: 7429545
Conc: 120.64 ng/ml



#24 AR-1248-4


R.T.: 5.197 min
Delta R.T.: -0.001 min
Response: 27903732
Conc: 388.07 ng/ml



#25 AR-1248-5


R.T.: 6.522 min
Delta R.T.: 0.000 min
Response: 6332124
Conc: 106.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.593 min
Delta R.T.: -0.001 min
Response: 6336420
Conc: 102.41 ng/ml




#26 AR-1254-1

R.T.: 6.456 min
Delta R.T.: 0.000 min
Response: 22793392
Conc: 282.97 ng/ml



#26 AR-1254-1


R.T.: 5.552 min
Delta R.T.: 0.000 min
Response: 27456607
Conc: 254.60 ng/ml



#27 AR-1254-2


R.T.: 6.677 min
Delta R.T.: 0.000 min
Response: 19561253
Conc: 168.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.701 min
Delta R.T.: 0.000 min
Response: 20301053
Conc: 211.75 ng/ml




#28 AR-1254-3

R.T.: 7.046 min
Delta R.T.: 0.000 min
Response: 8919953
Conc: 81.45 ng/ml



#28 AR-1254-3


R.T.: 6.125 min
Delta R.T.: 0.016 min
Response: 38042999
Conc: 252.99 ng/ml



#29 AR-1254-4


R.T.: 7.335 min
Delta R.T.: 0.001 min
Response: 5523942
Conc: 87.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.339 min
Delta R.T.: 0.000 min
Response: 2448421
Conc: 30.99 ng/ml




#30 AR-1254-5

R.T.: 7.755 min
Delta R.T.: 0.000 min
Response: 54715451
Conc: 729.28 ng/ml



#30 AR-1254-5


R.T.: 6.760 min
Delta R.T.: 0.000 min
Response: 67560624
Conc: 531.91 ng/ml



#31 AR-1260-1


R.T.: 7.213 min
Delta R.T.: 0.000 min
Response: 49298552
Conc: 503.39 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.240 min
Delta R.T.: 0.000 min
Response: 57217643
Conc: 501.36 ng/ml




#32 AR-1260-2

R.T.: 7.472 min
Delta R.T.: 0.000 min
Response: 51999989
Conc: 513.00 ng/ml



#32 AR-1260-2


R.T.: 6.430 min
Delta R.T.: 0.000 min
Response: 65965868
Conc: 508.30 ng/ml



#33 AR-1260-3


R.T.: 7.834 min
Delta R.T.: -0.002 min
Response: 40785323
Conc: 503.21 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.585 min
Delta R.T.: 0.000 min
Response: 63114277
Conc: 506.28 ng/ml




#34 AR-1260-4

R.T.: 8.061 min
Delta R.T.: -0.002 min
Response: 46731441
Conc: 496.12 ng/ml



#34 AR-1260-4


R.T.: 7.061 min
Delta R.T.: 0.000 min
Response: 52765068
Conc: 504.98 ng/ml



#35 AR-1260-5


R.T.: 8.388 min
Delta R.T.: -0.001 min
Response: 80700811
Conc: 513.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.303 min
Delta R.T.: 0.000 min
Response: 108448780
Conc: 511.38 ng/ml




#36 AR-1262-1

R.T.: 7.834 min
Delta R.T.: 0.000 min
Response: 40785323
Conc: 389.50 ng/ml



#36 AR-1262-1


R.T.: 6.853 min
Delta R.T.: 0.000 min
Response: 29776905
Conc: 491.77 ng/ml



#37 AR-1262-2


R.T.: 8.388 min
Delta R.T.: 0.000 min
Response: 80700811
Conc: 540.26 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.061 min
Delta R.T.: 0.000 min
Response: 52765068
Conc: 424.65 ng/ml




#38 AR-1262-3

R.T.: 8.719 min
Delta R.T.: 0.011 min
Response: 47166126
Conc: 452.08 ng/ml



#38 AR-1262-3


R.T.: 7.589 min
Delta R.T.: 0.000 min
Response: 27285793
Conc: 286.39 ng/ml



#39 AR-1262-4


R.T.: 8.800 min
Delta R.T.: 0.000 min
Response: 13543707
Conc: 181.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.653 min
Delta R.T.: -0.002 min
Response: 80203620
Conc: 500.29 ng/ml




#40 AR-1262-5

R.T.: 9.477 min
Delta R.T.: 0.000 min
Response: 20328144
Conc: 431.59 ng/ml



#40 AR-1262-5


R.T.: 8.153 min
Delta R.T.: 0.000 min
Response: 27110809
Conc: 391.10 ng/ml



#41 AR-1268-1


R.T.: 8.719 min
Delta R.T.: 0.013 min
Response: 47166126
Conc: 224.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.589 min
Delta R.T.: 0.000 min
Response: 27285793
Conc: 96.25 ng/ml




#42 AR-1268-2

R.T.: 8.800 min
Delta R.T.: -0.003 min
Response: 13543707
Conc: 84.17 ng/ml



#42 AR-1268-2


R.T.: 7.653 min
Delta R.T.: -0.002 min
Response: 80203620
Conc: 318.71 ng/ml



#43 AR-1268-3


R.T.: 9.039 min
Delta R.T.: 0.001 min
Response: 1146645
Conc: 6.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.863 min
Delta R.T.: 0.000 min
Response: 1636908
Conc: 7.12 ng/ml




#44 AR-1268-4

R.T.: 9.477 min
Delta R.T.: 0.002 min
Response: 20328144
Conc: 388.17 ng/ml



#44 AR-1268-4


R.T.: 8.153 min
Delta R.T.: 0.000 min
Response: 27110809
Conc: 342.10 ng/ml



#45 AR-1268-5

R.T.: 9.909 min
Delta R.T.: 0.002 min
Response: 6810896
Conc: 13.64 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.451 min
Delta R.T.: 0.000 min
Response: 8574843
Conc: 14.14 ng/ml