

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050323\
 Data File : PP057466.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2023 02:12
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 04 03:44:42 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 04 03:44:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.393	3.626	206.0E6	201.6E6	94.704	96.096
2) SA Decachlor...	10.225	8.693	72308938	131.9E6	93.123	92.249
Target Compounds						
26) L6 AR-1254-1	6.437	5.536	55073212	89667335	902.886	919.314
27) L6 AR-1254-2	6.658	5.685	77140849	78705861	917.206	912.441
28) L6 AR-1254-3	7.026	6.093	74774162	126.9E6	933.702	927.823
29) L6 AR-1254-4	7.315	6.323	51246201	77082845	930.102	924.482
30) L6 AR-1254-5	7.738	6.744	55817506	116.3E6	922.835	918.331



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050323\
Data File : PP057466.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 02:12
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 03:44:42 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050323.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 04 03:44:16 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.393 min
Delta R.T.: 0.000 min
Response: 205982452
Conc: 94.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.626 min
Delta R.T.: 0.000 min
Response: 201647513
Conc: 96.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.225 min
Delta R.T.: 0.001 min
Response: 72308938
Conc: 93.12 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.693 min
Delta R.T.: 0.002 min
Response: 131942827
Conc: 92.25 ng/ml



#26 AR-1254-1


R.T.: 6.437 min
Delta R.T.: 0.000 min
Response: 55073212
Conc: 902.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#26 AR-1254-1

R.T.: 5.536 min
Delta R.T.: 0.000 min
Response: 89667335
Conc: 919.31 ng/ml




#27 AR-1254-2

R.T.: 6.658 min
Delta R.T.: 0.000 min
Response: 77140849
Conc: 917.21 ng/ml



#27 AR-1254-2


R.T.: 5.685 min
Delta R.T.: 0.000 min
Response: 78705861
Conc: 912.44 ng/ml



#28 AR-1254-3


R.T.: 7.026 min
Delta R.T.: 0.000 min
Response: 74774162
Conc: 933.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#28 AR-1254-3

R.T.: 6.093 min
Delta R.T.: 0.000 min
Response: 126941887
Conc: 927.82 ng/ml




#29 AR-1254-4

R.T.: 7.315 min
Delta R.T.: 0.001 min
Response: 51246201
Conc: 930.10 ng/ml



#29 AR-1254-4


R.T.: 6.323 min
Delta R.T.: 0.000 min
Response: 77082845
Conc: 924.48 ng/ml



#30 AR-1254-5

R.T.: 7.738 min
Delta R.T.: 0.000 min
Response: 55817506
Conc: 922.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000



#30 AR-1254-5

R.T.: 6.744 min
Delta R.T.: 0.000 min
Response: 116300346
Conc: 918.33 ng/ml