

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050323\
 Data File : PP057457.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 May 2023 23:43
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 04 01:37:04 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 04 01:36:57 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.393	3.626	156.5E6	152.6E6	75.000	75.078
2) SA Decachlor...	10.224	8.692	54634829	98109390	75.426	74.649
Target Compounds						
16) L4 AR-1242-1	5.573	4.724	30793854	35370575	753.393	743.682
17) L4 AR-1242-2	5.595	4.743	44181490	49210605	743.265	747.799
18) L4 AR-1242-3	5.658	4.922	28409722	28588755	750.678	748.214
19) L4 AR-1242-4	5.757	5.006	22838757	28240465	747.164	746.108
20) L4 AR-1242-5	6.501	5.534	19810187	34491535	747.891	746.530



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050323\
Data File : PP057457.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 May 2023 23:43
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 01:37:04 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050323.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 04 01:36:57 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.393 min
Delta R.T.: 0.000 min
Response: 156541748
Conc: 75.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.626 min
Delta R.T.: 0.000 min
Response: 152644627
Conc: 75.08 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.224 min
Delta R.T.: 0.000 min
Response: 54634829
Conc: 75.43 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.692 min
Delta R.T.: 0.000 min
Response: 98109390
Conc: 74.65 ng/ml



#16 AR-1242-1


R.T.: 5.573 min
Delta R.T.: 0.000 min
Response: 30793854
Conc: 753.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#16 AR-1242-1

R.T.: 4.724 min
Delta R.T.: 0.000 min
Response: 35370575
Conc: 743.68 ng/ml




#17 AR-1242-2

R.T.: 5.595 min
Delta R.T.: 0.000 min
Response: 44181490
Conc: 743.27 ng/ml



#17 AR-1242-2


R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 49210605
Conc: 747.80 ng/ml



#18 AR-1242-3


R.T.: 5.658 min
Delta R.T.: 0.000 min
Response: 28409722
Conc: 750.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#18 AR-1242-3

R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 28588755
Conc: 748.21 ng/ml




#19 AR-1242-4

R.T.: 5.757 min
Delta R.T.: 0.000 min
Response: 22838757
Conc: 747.16 ng/ml



#19 AR-1242-4


R.T.: 5.006 min
Delta R.T.: 0.000 min
Response: 28240465
Conc: 746.11 ng/ml



#20 AR-1242-5

R.T.: 6.501 min
Delta R.T.: 0.000 min
Response: 19810187
Conc: 747.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750



#20 AR-1242-5

R.T.: 5.534 min
Delta R.T.: 0.000 min
Response: 34491535
Conc: 746.53 ng/ml