

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050324\
 Data File : PP064748.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2024 02:26
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 04 06:30:14 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP043024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 01 05:01:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.379	3.454	74204038	113.5E6	42.160	48.160
2) SA Decachlor...	10.068	8.353	49873479	82029063	56.055	43.974
Target Compounds						
3) L1 AR-1016-1	5.542	4.514	20784356	34095561	434.801	496.141
4) L1 AR-1016-2	5.563	4.532	30630908	48928078	421.541	483.956
5) L1 AR-1016-3	5.625	4.704	20140895	26879221	426.075	491.728
6) L1 AR-1016-4	5.724	4.746	16205171	21535412	433.638	457.600
7) L1 AR-1016-5	6.018	4.954	16196029	28281701	420.411	470.114
31) L7 AR-1260-1	7.141	5.971	29961460	53926624	437.912	459.158
32) L7 AR-1260-2	7.397	6.158	33004548	63386800	464.916	484.677
33) L7 AR-1260-3	7.756	6.308	26110852	59597426	464.073	463.678
34) L7 AR-1260-4	7.980	6.775	29110267	47935991	494.049	461.217
35) L7 AR-1260-5	8.295	7.016	52050158	102.9E6	551.877	507.790



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050324\
Data File : PP064748.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2024 02:26
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 06:30:14 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP043024.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 01 05:01:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.379 min
Delta R.T.: 0.007 min
Response: 74204038
Conc: 42.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.454 min
Delta R.T.: 0.011 min
Response: 113548145
Conc: 48.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.068 min
Delta R.T.: -0.011 min
Response: 49873479
Conc: 56.06 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.353 min
Delta R.T.: -0.009 min
Response: 82029063
Conc: 43.97 ng/ml



#3 AR-1016-1


R.T.: 5.542 min
Delta R.T.: 0.000 min
Response: 20784356
Conc: 434.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.514 min
Delta R.T.: 0.002 min
Response: 34095561
Conc: 496.14 ng/ml




#4 AR-1016-2

R.T.: 5.563 min
Delta R.T.: 0.000 min
Response: 30630908
Conc: 421.54 ng/ml



#4 AR-1016-2


R.T.: 4.532 min
Delta R.T.: 0.002 min
Response: 48928078
Conc: 483.96 ng/ml



#5 AR-1016-3


R.T.: 5.625 min
Delta R.T.: 0.000 min
Response: 20140895
Conc: 426.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.704 min
Delta R.T.: 0.001 min
Response: 26879221
Conc: 491.73 ng/ml




#6 AR-1016-4

R.T.: 5.724 min
Delta R.T.: 0.000 min
Response: 16205171
Conc: 433.64 ng/ml



#6 AR-1016-4


R.T.: 4.746 min
Delta R.T.: 0.000 min
Response: 21535412
Conc: 457.60 ng/ml



#7 AR-1016-5


R.T.: 6.018 min
Delta R.T.: 0.000 min
Response: 16196029
Conc: 420.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 4.954 min
Delta R.T.: 0.000 min
Response: 28281701
Conc: 470.11 ng/ml




#31 AR-1260-1

R.T.: 7.141 min
Delta R.T.: -0.004 min
Response: 29961460
Conc: 437.91 ng/ml



#31 AR-1260-1


R.T.: 5.971 min
Delta R.T.: -0.004 min
Response: 53926624
Conc: 459.16 ng/ml



#32 AR-1260-2


R.T.: 7.397 min
Delta R.T.: -0.004 min
Response: 33004548
Conc: 464.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.158 min
Delta R.T.: -0.004 min
Response: 63386800
Conc: 484.68 ng/ml




#33 AR-1260-3

R.T.: 7.756 min
Delta R.T.: -0.004 min
Response: 26110852
Conc: 464.07 ng/ml



#33 AR-1260-3


R.T.: 6.308 min
Delta R.T.: -0.005 min
Response: 59597426
Conc: 463.68 ng/ml



#34 AR-1260-4


R.T.: 7.980 min
Delta R.T.: -0.006 min
Response: 29110267
Conc: 494.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.775 min
Delta R.T.: -0.006 min
Response: 47935991
Conc: 461.22 ng/ml



#35 AR-1260-5

R.T.: 8.295 min
Delta R.T.: -0.006 min
Response: 52050158
Conc: 551.88 ng/ml



#35 AR-1260-5

R.T.: 7.016 min
Delta R.T.: -0.006 min
Response: 102928921
Conc: 507.79 ng/ml