

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050421\
 Data File : PP035524.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 May 2021 00:07
 Operator : DD\AJ
 Sample : M2273-11 10X
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 05 04:49:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 03 09:49:05 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.788	4.206	108125	84600	2.130	2.485
2) SA Decachlor...	10.642f	9.460	109250	49211	2.690	2.540
Target Compounds						
28) L6 AR-1254-3	7.623	0.000	6897	0	2.347	N.D. #
33) L7 AR-1260-3	8.405	0.000	10272	0	6.061	N.D. #
34) L7 AR-1260-4	8.640	0.000	11274	0	5.325	N.D. #
36) L8 AR-1262-1	8.405	0.000	10272	0	3.602	N.D. #
38) L8 AR-1262-3	0.000	8.366	0	9115	N.D.	7.301 #
41) L9 AR-1268-1	0.000	8.366	0	9115	N.D.	2.630 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050421\
Data File : PP035524.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 May 2021 00:07
Operator : DD\AJ
Sample : M2273-11 10X
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 05 04:49:13 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042921.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 03 09:49:05 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.788 min
Delta R.T.: -0.003 min
Response: 108125
Conc: 2.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.206 min
Delta R.T.: -0.003 min
Response: 84600
Conc: 2.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.642 min
Delta R.T.: -0.018 min
Response: 109250
Conc: 2.69 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.460 min
Delta R.T.: -0.014 min
Response: 49211
Conc: 2.54 ng/ml



#28 AR-1254-3


R.T.: 7.623 min
Delta R.T.: 0.014 min
Response: 6897
Conc: 2.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.892 min
Response: 0
Conc: N.D.




#33 AR-1260-3

R.T.: 8.405 min
Delta R.T.: -0.001 min
Response: 10272
Conc: 6.06 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T. : 7.374 min
Response: 0
Conc: N.D.



#34 AR-1260-4


R.T.: 8.640 min
Delta R.T.: 0.003 min
Response: 11274
Conc: 5.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 7.853 min
Response: 0
Conc: N.D.




#36 AR-1262-1

R.T.: 8.405 min
Delta R.T.: 0.001 min
Response: 10272
Conc: 3.60 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T. : 7.551 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T.: 9.245 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#38 AR-1262-3

R.T.: 8.366 min
Delta R.T.: -0.013 min
Response: 9115
Conc: 7.30 ng/ml



#41 AR-1268-1

R.T.: 0.000 min
Exp R.T.: 9.244 min
Response: 0
Conc: N.D.



#41 AR-1268-1

R.T.: 8.366 min
Delta R.T.: -0.012 min
Response: 9115
Conc: 2.63 ng/ml