

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050423\
 Data File : PP057527.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2023 20:47
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 05 08:57:18 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 05 08:53:51 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.391	3.640	127.7E6	122.8E6	61.375	60.651
2) SA Decachlor...	10.217	8.702	48191452	86075378	65.933	66.164
Target Compounds						
21) L5 AR-1248-1	5.570	4.737	20995440	23850908	543.756	522.136
22) L5 AR-1248-2	5.847	4.977	32258385	33155054	548.553	518.745
23) L5 AR-1248-3	6.052	5.019	34135225	35147304	550.742	521.663
24) L5 AR-1248-4	6.459	5.193	31804891	41778205	567.166	528.609
25) L5 AR-1248-5	6.498	5.590	31143474	40211281	558.952	538.854



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050423\
Data File : PP057527.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 20:47
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 05 08:57:18 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050323.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 05 08:53:51 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.391 min
Delta R.T.: -0.002 min
Response: 127652610
Conc: 61.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.640 min
Delta R.T.: 0.015 min
Response: 122847168
Conc: 60.65 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.217 min
Delta R.T.: -0.006 min
Response: 48191452
Conc: 65.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.702 min
Delta R.T.: 0.009 min
Response: 86075378
Conc: 66.16 ng/ml



#21 AR-1248-1


R.T.: 5.570 min
Delta R.T.: -0.004 min
Response: 20995440
Conc: 543.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.737 min
Delta R.T.: 0.013 min
Response: 23850908
Conc: 522.14 ng/ml




#22 AR-1248-2

R.T.: 5.847 min
Delta R.T.: -0.004 min
Response: 32258385
Conc: 548.55 ng/ml



#22 AR-1248-2


R.T.: 4.977 min
Delta R.T.: 0.013 min
Response: 33155054
Conc: 518.74 ng/ml



#23 AR-1248-3


R.T.: 6.052 min
Delta R.T.: -0.004 min
Response: 34135225
Conc: 550.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.019 min
Delta R.T.: 0.012 min
Response: 35147304
Conc: 521.66 ng/ml




#24 AR-1248-4

R.T.: 6.459 min
Delta R.T.: -0.004 min
Response: 31804891
Conc: 567.17 ng/ml



#24 AR-1248-4


R.T.: 5.193 min
Delta R.T.: 0.013 min
Response: 41778205
Conc: 528.61 ng/ml



#25 AR-1248-5

R.T.: 6.498 min
Delta R.T.: -0.004 min
Response: 31143474
Conc: 558.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.590 min
Delta R.T.: 0.013 min
Response: 40211281
Conc: 538.85 ng/ml