

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050423\
 Data File : PP057530.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2023 21:36
 Operator : YP\AJ
 Sample : 02596-08
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 XYZ-1532

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/05/2023
 Supervised By :Ankita Jodhani 05/05/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 05 01:30:35 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 04 07:27:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.391	3.624	46093884	44306413	22.162	21.874
2) SA Decachlor...	10.217	8.685	19197210	38648599	26.265	29.708
Target Compounds						
26) L6 AR-1254-1	6.433	5.532	852153	1748978	13.905m	17.429 #
27) L6 AR-1254-2	6.656	5.681	1518470	1185127	17.830	13.328m#
28) L6 AR-1254-3	7.024	6.088	1580649	2347405	19.786	16.762
29) L6 AR-1254-4	7.311	6.319	1333568	1844016	23.887	21.622
30) L6 AR-1254-5	7.732	6.740	1483873	3446685	24.397m	26.354

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050423\
Data File : PP057530.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 21:36
Operator : YP\AJ
Sample : 02596-08
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
XYZ-1532


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/05/2023
Supervised By :Ankita Jodhani 05/05/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 05 01:30:35 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050323.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 04 07:27:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.391 min
Delta R.T.: -0.001 min
Response: 46093884
Conc: 22.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
XYZ-1532


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/05/2023
Supervised By :Ankita Jodhani 05/05/2023




#1 Tetrachloro-m-xylene

R.T.: 3.624 min
Delta R.T.: 0.000 min
Response: 44306413
Conc: 21.87 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.217 min
Delta R.T.: -0.007 min
Response: 19197210
Conc: 26.26 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.685 min
Delta R.T.: -0.007 min
Response: 38648599
Conc: 29.71 ng/ml




#26 AR-1254-1

R.T.: 6.433 min
Delta R.T.: -0.004 min
Response: 852153
Conc: 13.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
XYZ-1532


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/05/2023
Supervised By :Ankita Jodhani 05/05/2023




#26 AR-1254-1

R.T.: 5.532 min
Delta R.T.: -0.003 min
Response: 1748978
Conc: 17.43 ng/ml




#27 AR-1254-2

R.T.: 6.656 min
Delta R.T.: -0.002 min
Response: 1518470
Conc: 17.83 ng/ml



#27 AR-1254-2

R.T.: 5.681 min
Delta R.T.: -0.004 min
Response: 1185127
Conc: 13.33 ng/ml m




#28 AR-1254-3

R.T.: 7.024 min
Delta R.T.: -0.003 min
Response: 1580649
Conc: 19.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
XYZ-1532


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/05/2023
Supervised By :Ankita Jodhani 05/05/2023




#28 AR-1254-3

R.T.: 6.088 min
Delta R.T.: -0.004 min
Response: 2347405
Conc: 16.76 ng/ml




#29 AR-1254-4

R.T.: 7.311 min
Delta R.T.: -0.003 min
Response: 1333568
Conc: 23.89 ng/ml



#29 AR-1254-4

R.T.: 6.319 min
Delta R.T.: -0.004 min
Response: 1844016
Conc: 21.62 ng/ml




#30 AR-1254-5

R.T.: 7.732 min
Delta R.T.: -0.006 min
Response: 1483873
Conc: 24.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
XYZ-1532

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/05/2023
Supervised By :Ankita Jodhani 05/05/2023



#30 AR-1254-5

R.T.: 6.740 min
Delta R.T.: -0.004 min
Response: 3446685
Conc: 26.35 ng/ml