

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050521\
 Data File : PP035572.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 May 2021 22:32
 Operator : DD\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP050521AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 08:37:01 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 06 08:33:38 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	43.752	45.802 E3	-4.7	105	0.00
2 SA Decachlorobiphenyl	39.443	40.642 E3	-3.0	102	0.00
16 L4 AR-1242-1	1.142	1.123 E3	1.7	97	0.00
17 L4 AR-1242-2	1.717	1.720 E3	-0.2	99	0.00
18 L4 AR-1242-3	1.085	1.152 E3	-6.2	103	0.00
19 L4 AR-1242-4	876.323	920.984	-5.1	102	0.00
20 L4 AR-1242-5	882.957	895.451	-1.4	99	0.00

Signal #2

1 SA Tetrachloro-m-xylene	30.334	31.547 E3	-4.0	104	0.00
2 SA Decachlorobiphenyl	19.294	19.827 E3	-2.8	103	0.00
16 L4 AR-1242-1	709.684	683.083	3.7	95	0.00
17 L4 AR-1242-2	1.244	1.237 E3	0.6	98	0.00
18 L4 AR-1242-3	585.136	570.940	2.4	95	0.00
19 L4 AR-1242-4	586.662	580.116	1.1	97	0.00
20 L4 AR-1242-5	676.031	667.999	1.2	96	0.00

Evaluate Continuing Calibration Report - Not Found

3 L1 AR-1016-1	1.473	0.000 E3	100.0#	0#	-6.09#
4 L1 AR-1016-2	2.206	0.000 E3	100.0#	0#	-6.12#
5 L1 AR-1016-3	1.395	0.000 E3	100.0#	0#	-6.18#
6 L1 AR-1016-4	1.135	0.000 E3	100.0#	0#	-6.29#
7 L1 AR-1016-5	1.150	0.000 E3	100.0#	0#	-6.60#
8 L2 AR-1221-1	520.170	0.000	100.0#	0#	-5.03#
9 L2 AR-1221-2	393.834	0.000	100.0#	0#	-5.13#
10 L2 AR-1221-3	1.196	0.000 E3	100.0#	0#	-5.22#
11 L3 AR-1232-1	918.949	0.000	100.0#	0#	-5.22#
12 L3 AR-1232-2	498.166	0.000	100.0#	0#	-5.80#
13 L3 AR-1232-3	891.664	0.000	100.0#	0#	-6.12#
14 L3 AR-1232-4	451.944	0.000	100.0#	0#	-6.29#
15 L3 AR-1232-5	323.607	0.000	100.0#	0#	-6.38#
21 L5 AR-1248-1	832.649	0.000	100.0#	0#	-6.09#
22 L5 AR-1248-2	1.244	0.000 E3	100.0#	0#	-6.38#
23 L5 AR-1248-3	1.518	0.000 E3	100.0#	0#	-6.60#
24 L5 AR-1248-4	1.494	0.000 E3	100.0#	0#	-7.02#
25 L5 AR-1248-5	1.477	0.000 E3	100.0#	0#	-7.06#
26 L6 AR-1254-1	1.678	0.000 E3	100.0#	0#	-6.99#
27 L6 AR-1254-2	2.519	0.000 E3	100.0#	0#	-7.22#
28 L6 AR-1254-3	2.589	0.000 E3	100.0#	0#	-7.60#
29 L6 AR-1254-4	1.708	0.000 E3	100.0#	0#	-7.89#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050521\
 Data File : PP035572.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 May 2021 22:32
 Operator : DD\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP050521AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 08:37:01 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 06 08:33:38 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
30 L6 AR-1254-5	2.068	0.000 E3	100.0#	0#	-8.32#
31 L7 AR-1260-1	2.167	0.000 E3	100.0#	0#	-7.77#
32 L7 AR-1260-2	2.460	0.000 E3	100.0#	0#	-8.03#
33 L7 AR-1260-3	1.621	0.000 E3	100.0#	0#	-8.39#
34 L7 AR-1260-4	1.980	0.000 E3	100.0#	0#	-8.62#
35 L7 AR-1260-5	3.694	0.000 E3	100.0#	0#	-8.94#
36 L8 AR-1262-1	2.622	0.000 E3	100.0#	0#	-8.39#
37 L8 AR-1262-2	4.174	0.000 E3	100.0#	0#	-8.94#
38 L8 AR-1262-3	3.068	0.000 E3	100.0#	0#	-9.23#
39 L8 AR-1262-4	2.390	0.000 E3	100.0#	0#	-9.32#
40 L8 AR-1262-5	1.531	0.000 E3	100.0#	0#	-9.95#
41 L9 AR-1268-1	5.446	0.000 E3	100.0#	0#	-9.23#
42 L9 AR-1268-2	5.098	0.000 E3	100.0#	0#	-9.32#
43 L9 AR-1268-3	4.642	0.000 E3	100.0#	0#	-9.53#
44 L9 AR-1268-4	1.661	0.000 E3	100.0#	0#	-9.94#
45 L9 AR-1268-5	13.545	0.000 E3	100.0#	0#	-10.33#

Signal #2

3 L1 AR-1016-1	916.430	0.000	100.0#	0#	-5.45#
4 L1 AR-1016-2	1.520	0.000 E3	100.0#	0#	-5.47#
5 L1 AR-1016-3	752.628	0.000	100.0#	0#	-5.66#
6 L1 AR-1016-4	607.075	0.000	100.0#	0#	-5.71#
7 L1 AR-1016-5	764.269	0.000	100.0#	0#	-5.93#
8 L2 AR-1221-1	373.381	0.000	100.0#	0#	-4.46#
9 L2 AR-1221-2	280.193	0.000	100.0#	0#	-4.56#
10 L2 AR-1221-3	874.783	0.000	100.0#	0#	-4.65#
11 L3 AR-1232-1	673.071	0.000	100.0#	0#	-4.65#
12 L3 AR-1232-2	658.899	0.000	100.0#	0#	-5.47#
13 L3 AR-1232-3	302.914	0.000	100.0#	0#	-5.66#
14 L3 AR-1232-4	271.311	0.000	100.0#	0#	-5.75#
15 L3 AR-1232-5	288.876	0.000	100.0#	0#	-5.93#
21 L5 AR-1248-1	530.572	0.000	100.0#	0#	-5.45#
22 L5 AR-1248-2	809.551	0.000	100.0#	0#	-5.70#
23 L5 AR-1248-3	859.620	0.000	100.0#	0#	-5.75#
24 L5 AR-1248-4	982.691	0.000	100.0#	0#	-5.93#
25 L5 AR-1248-5	908.030	0.000	100.0#	0#	-6.35#
26 L6 AR-1254-1	1.420	0.000 E3	100.0#	0#	-6.31#
27 L6 AR-1254-2	1.290	0.000 E3	100.0#	0#	-6.46#
28 L6 AR-1254-3	1.898	0.000 E3	100.0#	0#	-6.88#
29 L6 AR-1254-4	1.100	0.000 E3	100.0#	0#	-7.12#
30 L6 AR-1254-5	1.677	0.000 E3	100.0#	0#	-7.54#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050521\
 Data File : PP035572.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 May 2021 22:32
 Operator : DD\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP050521AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 08:37:01 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 06 08:33:38 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
31 L7 AR-1260-1	1.355	0.000 E3	100.0#	0#	-7.02#
32 L7 AR-1260-2	1.577	0.000 E3	100.0#	0#	-7.21#
33 L7 AR-1260-3	1.492	0.000 E3	100.0#	0#	-7.36#
34 L7 AR-1260-4	1.077	0.000 E3	100.0#	0#	-7.84#
35 L7 AR-1260-5	2.378	0.000 E3	100.0#	0#	-8.08#
36 L8 AR-1262-1	864.151	0.000	100.0#	0#	-7.54#
37 L8 AR-1262-2	2.748	0.000 E3	100.0#	0#	-8.08#
38 L8 AR-1262-3	1.206	0.000 E3	100.0#	0#	-8.37#
39 L8 AR-1262-4	2.134	0.000 E3	100.0#	0#	-8.43#
40 L8 AR-1262-5	1023.856	0.000	100.0#	0#	-8.92#
41 L9 AR-1268-1	3.100	0.000 E3	100.0#	0#	-8.36#
42 L9 AR-1268-2	2.802	0.000 E3	100.0#	0#	-8.43#
43 L9 AR-1268-3	2.435	0.000 E3	100.0#	0#	-8.63#
44 L9 AR-1268-4	0.998	0.000 E3	100.0#	0#	-8.92#
45 L9 AR-1268-5	6.774	0.000 E3	100.0#	0#	-9.21#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0