

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050523\
 Data File : PP057587.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 May 2023 22:08
 Operator : YP\AJ
 Sample : 02610-04 10X
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 07:27:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 05 08:53:51 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.390	3.624	4749919	4736574	2.284	2.338
2)	SA Decachlor...	10.216	8.685	2981216	3044596	4.079	2.340 #
Target Compounds							
3)	L1 AR-1016-1	5.571	4.717	28888	47981	0.528	0.749 #
4)	L1 AR-1016-2	5.592	4.736	57092	96480	0.705	1.090 #
5)	L1 AR-1016-3	5.646	4.915	82030	74512	1.638	1.478
6)	L1 AR-1016-4	5.758	4.938f	27227	20116	0.675	0.499 #
7)	L1 AR-1016-5	6.048	0.000	97045	0	2.376	N.D. #
8)	L2 AR-1221-1	4.602	0.000	63420	0	2.539	N.D. #
9)	L2 AR-1221-2	4.701	3.928	171748	222829	9.292	11.922 #
10)	L2 AR-1221-3	4.760	0.000	64005	0	1.178	N.D. #
11)	L3 AR-1232-1	4.760	0.000	64005	0	1.576	N.D. #
12)	L3 AR-1232-2	5.317	4.736	152649	96480	6.522	2.327 #
13)	L3 AR-1232-3	5.592	4.915	57092	74512	1.547	3.129 #
14)	L3 AR-1232-4	5.758	4.999	27227	17709	1.501	0.833 #
15)	L3 AR-1232-5	5.853	0.000	66075	0	4.270	N.D. #
16)	L4 AR-1242-1	5.571	4.717	28888	47981	0.669	0.946 #
17)	L4 AR-1242-2	5.592	4.736	57092	96480	0.911	1.390 #
18)	L4 AR-1242-3	5.646	4.915	82030	74512	2.088	1.839
19)	L4 AR-1242-4	5.758	4.999	27227	17709	0.855	0.440 #
20)	L4 AR-1242-5	6.473f	5.543	89126	100682	3.205	2.051 #
21)	L5 AR-1248-1	5.571	4.717	28888	47981	0.748	1.050 #
22)	L5 AR-1248-2	5.853	4.938f	66075	20116	1.124	0.315 #
23)	L5 AR-1248-3	6.048	4.999	97045	17709	1.566	0.263 #
24)	L5 AR-1248-4	6.456	0.000	319401	0	5.696	N.D. #
25)	L5 AR-1248-5	6.473f	5.562	89126	12911	1.600	0.173 #
26)	L6 AR-1254-1	6.456f	5.543	319401	100682	5.212	1.003 #
27)	L6 AR-1254-2	6.652	5.682	482867	69191	5.670	0.778 #
28)	L6 AR-1254-3	7.026	6.090	457494	68132	5.727	0.486 #
29)	L6 AR-1254-4	7.323	6.317	143254	58759	2.566	0.689 #
30)	L6 AR-1254-5	7.729	6.741	385674	365227	6.341	2.793 #
31)	L7 AR-1260-1	7.171f	6.243f	71779	36635	1.006	0.394 #
32)	L7 AR-1260-2	7.435f	6.408	116688	215142	1.809	2.024
33)	L7 AR-1260-3	7.836f	6.567	85969	285687	2.072	2.745 #
34)	L7 AR-1260-4	8.044	7.043	608821	72848	11.998	0.990 #
35)	L7 AR-1260-5	8.363	7.283	128653	58793	1.550	0.372 #
36)	L8 AR-1262-1	7.836f	6.838	85969	282244	1.223	5.159 #
37)	L8 AR-1262-2	8.363	7.043	128653	72848	1.223	0.664 #
38)	L8 AR-1262-3	0.000	7.552f	0	704626	N.D.	8.414 #
39)	L8 AR-1262-4	8.793f	7.630	129697	89602	2.394	0.597 #
40)	L8 AR-1262-5	9.430f	0.000	37694	0	1.122	N.D. #
41)	L9 AR-1268-1	0.000	7.552f	0	704626	N.D.	2.871 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050523\
 Data File : PP057587.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 May 2023 22:08
 Operator : YP\AJ
 Sample : 02610-04 10X
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 07:27:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 05 08:53:51 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.793f	7.630	129697	89602	1.187	0.410 #
43) L9	AR-1268-3	9.010	0.000	46071	0	0.459	N.D. #
44) L9	AR-1268-4	9.430	8.104f	37694	75313	1.015	0.980
45) L9	AR-1268-5	9.881	8.429	34491	245827	0.135	0.486 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050523\
Data File : PP057587.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 May 2023 22:08
Operator : YP\AJ
Sample : 02610-04 10X
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 07:27:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050323.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 05 08:53:51 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.390 min
Delta R.T.: -0.002 min
Response: 4749919
Conc: 2.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.624 min
Delta R.T.: 0.000 min
Response: 4736574
Conc: 2.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.216 min
Delta R.T.: -0.008 min
Response: 2981216
Conc: 4.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.685 min
Delta R.T.: -0.008 min
Response: 3044596
Conc: 2.34 ng/ml



#3 AR-1016-1


R.T.: 5.571 min
Delta R.T.: 0.000 min
Response: 28888
Conc: 0.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.717 min
Delta R.T.: -0.007 min
Response: 47981
Conc: 0.75 ng/ml




#4 AR-1016-2

R.T.: 5.592 min
Delta R.T.: -0.002 min
Response: 57092
Conc: 0.71 ng/ml



#4 AR-1016-2


R.T.: 4.736 min
Delta R.T.: -0.007 min
Response: 96480
Conc: 1.09 ng/ml



#5 AR-1016-3


R.T.: 5.646 min
Delta R.T.: -0.012 min
Response: 82030
Conc: 1.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.915 min
Delta R.T.: -0.006 min
Response: 74512
Conc: 1.48 ng/ml




#6 AR-1016-4

R.T.: 5.758 min
Delta R.T.: 0.002 min
Response: 27227
Conc: 0.68 ng/ml



#6 AR-1016-4


R.T.: 4.938 min
Delta R.T.: -0.025 min
Response: 20116
Conc: 0.50 ng/ml



#7 AR-1016-5


R.T.: 6.048 min
Delta R.T.: -0.007 min
Response: 97045
Conc: 2.38 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T. : 5.179 min
Response: 0
Conc: N.D.




#8 AR-1221-1

R.T.: 4.602 min
Delta R.T.: 0.000 min
Response: 63420
Conc: 2.54 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.842 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.701 min
Delta R.T.: 0.011 min
Response: 171748
Conc: 9.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.928 min
Delta R.T.: 0.000 min
Response: 222829
Conc: 11.92 ng/ml




#10 AR-1221-3

R.T.: 4.760 min
Delta R.T.: -0.006 min
Response: 64005
Conc: 1.18 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T.: 4.006 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.760 min
Delta R.T.: -0.008 min
Response: 64005
Conc: 1.58 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 4.006 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.317 min
Delta R.T.: 0.014 min
Response: 152649
Conc: 6.52 ng/ml



#12 AR-1232-2


R.T.: 4.736 min
Delta R.T.: -0.008 min
Response: 96480
Conc: 2.33 ng/ml



#13 AR-1232-3


R.T.: 5.592 min
Delta R.T.: -0.003 min
Response: 57092
Conc: 1.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.915 min
Delta R.T.: -0.006 min
Response: 74512
Conc: 3.13 ng/ml




#14 AR-1232-4

R.T.: 5.758 min
Delta R.T.: 0.000 min
Response: 27227
Conc: 1.50 ng/ml



#14 AR-1232-4


R.T.: 4.999 min
Delta R.T.: -0.007 min
Response: 17709
Conc: 0.83 ng/ml



#15 AR-1232-5


R.T.: 5.853 min
Delta R.T.: 0.003 min
Response: 66075
Conc: 4.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 5.180 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.571 min
Delta R.T.: -0.001 min
Response: 28888
Conc: 0.67 ng/ml



#16 AR-1242-1


R.T.: 4.717 min
Delta R.T.: -0.007 min
Response: 47981
Conc: 0.95 ng/ml



#17 AR-1242-2


R.T.: 5.592 min
Delta R.T.: -0.003 min
Response: 57092
Conc: 0.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.736 min
Delta R.T.: -0.007 min
Response: 96480
Conc: 1.39 ng/ml




#18 AR-1242-3

R.T.: 5.646 min
Delta R.T.: -0.011 min
Response: 82030
Conc: 2.09 ng/ml



#18 AR-1242-3


R.T.: 4.915 min
Delta R.T.: -0.006 min
Response: 74512
Conc: 1.84 ng/ml



#19 AR-1242-4


R.T.: 5.758 min
Delta R.T.: 0.001 min
Response: 27227
Conc: 0.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.999 min
Delta R.T.: -0.006 min
Response: 17709
Conc: 0.44 ng/ml




#20 AR-1242-5

R.T.: 6.473 min
Delta R.T.: -0.028 min
Response: 89126
Conc: 3.21 ng/ml



#20 AR-1242-5


R.T.: 5.543 min
Delta R.T.: 0.009 min
Response: 100682
Conc: 2.05 ng/ml



#21 AR-1248-1


R.T.: 5.571 min
Delta R.T.: -0.002 min
Response: 28888
Conc: 0.75 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.717 min
Delta R.T.: -0.007 min
Response: 47981
Conc: 1.05 ng/ml




#22 AR-1248-2

R.T.: 5.853 min
Delta R.T.: 0.002 min
Response: 66075
Conc: 1.12 ng/ml



#22 AR-1248-2


R.T.: 4.938 min
Delta R.T.: -0.026 min
Response: 20116
Conc: 0.31 ng/ml



#23 AR-1248-3


R.T.: 6.048 min
Delta R.T.: -0.008 min
Response: 97045
Conc: 1.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.999 min
Delta R.T.: -0.007 min
Response: 17709
Conc: 0.26 ng/ml




#24 AR-1248-4

R.T.: 6.456 min
Delta R.T.: -0.007 min
Response: 319401
Conc: 5.70 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 5.180 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.473 min
Delta R.T.: -0.029 min
Response: 89126
Conc: 1.60 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.562 min
Delta R.T.: -0.015 min
Response: 12911
Conc: 0.17 ng/ml




#26 AR-1254-1

R.T.: 6.456 min
Delta R.T.: 0.019 min
Response: 319401
Conc: 5.21 ng/ml



#26 AR-1254-1


R.T.: 5.543 min
Delta R.T.: 0.008 min
Response: 100682
Conc: 1.00 ng/ml



#27 AR-1254-2


R.T.: 6.652 min
Delta R.T.: -0.005 min
Response: 482867
Conc: 5.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.682 min
Delta R.T.: -0.003 min
Response: 69191
Conc: 0.78 ng/ml




#28 AR-1254-3

R.T.: 7.026 min
Delta R.T.: 0.000 min
Response: 457494
Conc: 5.73 ng/ml



#28 AR-1254-3


R.T.: 6.090 min
Delta R.T.: -0.002 min
Response: 68132
Conc: 0.49 ng/ml



#29 AR-1254-4


R.T.: 7.323 min
Delta R.T.: 0.008 min
Response: 143254
Conc: 2.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.317 min
Delta R.T.: -0.005 min
Response: 58759
Conc: 0.69 ng/ml




#30 AR-1254-5

R.T.: 7.729 min
Delta R.T.: -0.009 min
Response: 385674
Conc: 6.34 ng/ml



#30 AR-1254-5


R.T.: 6.741 min
Delta R.T.: -0.003 min
Response: 365227
Conc: 2.79 ng/ml



#31 AR-1260-1


R.T.: 7.171 min
Delta R.T.: -0.022 min
Response: 71779
Conc: 1.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.243 min
Delta R.T.: 0.020 min
Response: 36635
Conc: 0.39 ng/ml




#32 AR-1260-2

R.T.: 7.435 min
Delta R.T.: -0.017 min
Response: 116688
Conc: 1.81 ng/ml



#32 AR-1260-2


R.T.: 6.408 min
Delta R.T.: -0.004 min
Response: 215142
Conc: 2.02 ng/ml



#33 AR-1260-3


R.T.: 7.836 min
Delta R.T.: 0.021 min
Response: 85969
Conc: 2.07 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.567 min
Delta R.T.: -0.001 min
Response: 285687
Conc: 2.74 ng/ml




#34 AR-1260-4

R.T.: 8.044 min
Delta R.T.: 0.003 min
Response: 608821
Conc: 12.00 ng/ml



#34 AR-1260-4


R.T.: 7.043 min
Delta R.T.: 0.000 min
Response: 72848
Conc: 0.99 ng/ml



#35 AR-1260-5


R.T.: 8.363 min
Delta R.T.: -0.001 min
Response: 128653
Conc: 1.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.283 min
Delta R.T.: -0.004 min
Response: 58793
Conc: 0.37 ng/ml




#36 AR-1262-1

R.T.: 7.836 min
Delta R.T.: 0.021 min
Response: 85969
Conc: 1.22 ng/ml



#36 AR-1262-1


R.T.: 6.838 min
Delta R.T.: 0.003 min
Response: 282244
Conc: 5.16 ng/ml



#37 AR-1262-2


R.T.: 8.363 min
Delta R.T.: -0.003 min
Response: 128653
Conc: 1.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.043 min
Delta R.T.: 0.000 min
Response: 72848
Conc: 0.66 ng/ml




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 8.684 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 7.552 min
Delta R.T.: -0.020 min
Response: 704626
Conc: 8.41 ng/ml



#39 AR-1262-4


R.T.: 8.793 min
Delta R.T.: 0.018 min
Response: 129697
Conc: 2.39 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.630 min
Delta R.T.: -0.006 min
Response: 89602
Conc: 0.60 ng/ml




#40 AR-1262-5

R.T.: 9.430 min
Delta R.T.: -0.017 min
Response: 37694
Conc: 1.12 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T.: 8.135 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T.: 8.680 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.552 min
Delta R.T.: -0.020 min
Response: 704626
Conc: 2.87 ng/ml




#42 AR-1268-2

R.T.: 8.793 min
Delta R.T.: 0.016 min
Response: 129697
Conc: 1.19 ng/ml



#42 AR-1268-2


R.T.: 7.630 min
Delta R.T.: -0.007 min
Response: 89602
Conc: 0.41 ng/ml



#43 AR-1268-3


R.T.: 9.010 min
Delta R.T.: 0.000 min
Response: 46071
Conc: 0.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 7.846 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 9.430 min
Delta R.T.: -0.015 min
Response: 37694
Conc: 1.01 ng/ml



#44 AR-1268-4


R.T.: 8.104 min
Delta R.T.: -0.030 min
Response: 75313
Conc: 0.98 ng/ml



#45 AR-1268-5

R.T.: 9.881 min
Delta R.T.: 0.010 min
Response: 34491
Conc: 0.13 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.429 min
Delta R.T.: -0.002 min
Response: 245827
Conc: 0.49 ng/ml