

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050523\
 Data File : PP057592.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 May 2023 23:31
 Operator : YP\AJ
 Sample : 02625-01 10X
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HEH657S-1-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/08/2023
 Supervised By :Ankita Jodhani 05/08/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 07:29:36 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 05 08:53:51 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.390	3.624	3860080	4197794	1.856	2.072
2) SA Decachlor...	10.217	8.683	2606758	2513581	3.566	1.932m#
Target Compounds						
26) L6 AR-1254-1	6.432	5.532	26122803	38250010	426.251	381.161
27) L6 AR-1254-2	6.653	5.680	38659059	36287025	453.932	408.101
28) L6 AR-1254-3	7.023	6.088	37504566	54990632	469.478	392.661
29) L6 AR-1254-4	7.311	6.318	26109329	33090818	467.675	388.013
30) L6 AR-1254-5	7.733	6.739	34532383	55130472	567.753	421.543 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050523\
Data File : PP057592.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 May 2023 23:31
Operator : YP\AJ
Sample : 02625-01 10X
Misc :
ALS Vial : 37 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
HEH657S-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/08/2023
Supervised By :Ankita Jodhani 05/08/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 07:29:36 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050323.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 05 08:53:51 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.390 min
Delta R.T.: -0.002 min
Response: 3860080
Conc: 1.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH657S-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/08/2023
Supervised By :Ankita Jodhani 05/08/2023




#1 Tetrachloro-m-xylene

R.T.: 3.624 min
Delta R.T.: 0.000 min
Response: 4197794
Conc: 2.07 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.217 min
Delta R.T.: -0.007 min
Response: 2606758
Conc: 3.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.683 min
Delta R.T.: -0.009 min
Response: 2513581
Conc: 1.93 ng/ml m




#26 AR-1254-1

R.T.: 6.432 min
Delta R.T.: -0.005 min
Response: 26122803
Conc: 426.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH657S-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/08/2023
Supervised By :Ankita Jodhani 05/08/2023




#26 AR-1254-1

R.T.: 5.532 min
Delta R.T.: -0.004 min
Response: 38250010
Conc: 381.16 ng/ml




#27 AR-1254-2

R.T.: 6.653 min
Delta R.T.: -0.005 min
Response: 38659059
Conc: 453.93 ng/ml



#27 AR-1254-2

R.T.: 5.680 min
Delta R.T.: -0.004 min
Response: 36287025
Conc: 408.10 ng/ml




#28 AR-1254-3

R.T.: 7.023 min
Delta R.T.: -0.004 min
Response: 37504566
Conc: 469.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH657S-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/08/2023
Supervised By :Ankita Jodhani 05/08/2023




#28 AR-1254-3

R.T.: 6.088 min
Delta R.T.: -0.005 min
Response: 54990632
Conc: 392.66 ng/ml




#29 AR-1254-4

R.T.: 7.311 min
Delta R.T.: -0.004 min
Response: 26109329
Conc: 467.67 ng/ml



#29 AR-1254-4

R.T.: 6.318 min
Delta R.T.: -0.005 min
Response: 33090818
Conc: 388.01 ng/ml




#30 AR-1254-5

R.T.: 7.733 min
Delta R.T.: -0.005 min
Response: 34532383
Conc: 567.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH657S-1-1

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/08/2023
Supervised By :Ankita Jodhani 05/08/2023



#30 AR-1254-5

R.T.: 6.739 min
Delta R.T.: -0.005 min
Response: 55130472
Conc: 421.54 ng/ml