

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050523\
 Data File : PP057604.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2023 04:03
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 07:32:48 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 05 08:53:51 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.390	3.624	86982287	92934433	41.821	45.883
2) SA Decachlor...	10.215	8.683	44678691	58944842	61.127	45.310 #
Target Compounds						
16) L4 AR-1242-1	5.568	4.721	19345439	22786452	448.008	449.320
17) L4 AR-1242-2	5.591	4.740	28001320	31406793	446.653	452.441
18) L4 AR-1242-3	5.654	4.918	18185723	18155649	462.913	448.149
19) L4 AR-1242-4	5.753	5.002	14357048	18060611	451.011	449.164
20) L4 AR-1242-5	6.497	5.530	12603896	22049148	453.270	449.245



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050523\
Data File : PP057604.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2023 04:03
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 07:32:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050323.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 05 08:53:51 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.390 min
Delta R.T.: -0.002 min
Response: 86982287
Conc: 41.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.624 min
Delta R.T.: -0.001 min
Response: 92934433
Conc: 45.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.215 min
Delta R.T.: -0.009 min
Response: 44678691
Conc: 61.13 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.683 min
Delta R.T.: -0.010 min
Response: 58944842
Conc: 45.31 ng/ml



#16 AR-1242-1


R.T.: 5.568 min
Delta R.T.: -0.004 min
Response: 19345439
Conc: 448.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.721 min
Delta R.T.: -0.003 min
Response: 22786452
Conc: 449.32 ng/ml




#17 AR-1242-2

R.T.: 5.591 min
Delta R.T.: -0.004 min
Response: 28001320
Conc: 446.65 ng/ml



#17 AR-1242-2


R.T.: 4.740 min
Delta R.T.: -0.003 min
Response: 31406793
Conc: 452.44 ng/ml



#18 AR-1242-3


R.T.: 5.654 min
Delta R.T.: -0.004 min
Response: 18185723
Conc: 462.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.918 min
Delta R.T.: -0.003 min
Response: 18155649
Conc: 448.15 ng/ml




#19 AR-1242-4

R.T.: 5.753 min
Delta R.T.: -0.004 min
Response: 14357048
Conc: 451.01 ng/ml



#19 AR-1242-4


R.T.: 5.002 min
Delta R.T.: -0.004 min
Response: 18060611
Conc: 449.16 ng/ml



#20 AR-1242-5

R.T.: 6.497 min
Delta R.T.: -0.004 min
Response: 12603896
Conc: 453.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.530 min
Delta R.T.: -0.004 min
Response: 22049148
Conc: 449.24 ng/ml